



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: KETANKUMAR

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB95222

Ambient Barometric Pressure (mmHg): 760

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	27.0	80.6	80.6	05/09/2018	09:15
2	J2787-02		1	>100.0	>212.0	>212.0	05/09/2018	09:40
3	J2787-02DUP		1	>100.0	>212.0	>212.0	05/09/2018	10:10
4	J2803-07		1	>100.0	>212.0	>212.0	05/09/2018	10:35
5	J2811-14		1	>100.0	>212.0	>212.0	05/09/2018	10:50

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.33

LB95222

WORKLIST(Hardcopy Internal Chain)

WorkList Name : FPW-050918

WorkList ID : 111826

Date : 5/9/2018 8:13:21 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
05/10/2018	Water	J2787-02	Flash Point	Cool 4 deg C	PSEG01	I31	V23608-VB102	05/07/2018	1010A
05/15/2018	Water	J2803-07	Flash Point	Cool 4 deg C	PSEG02	I22	OILY-WATER	05/08/2018	1010A
05/11/2018	Water	J2811-14	Flash Point	Cool 4 deg C	PSEG01	I33	TP-6	05/08/2018	1010A

Date/Time 05/09/18 8:30 AM
Received by: CP
Relinquished by: CP

Date/Time 05/09/18 10:55 AM
Received by: CP
Relinquished by: KS