



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: KETANKUMAR

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB95223

Ambient Barometric Pressure (mmHg): 760

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	28.0	82.4	82.4	05/09/2018	11:20
2	J1578-14		1	>100.0	>212.0	>212.0	05/09/2018	11:50
3	J1578-14DUP		1	>100.0	>212.0	>212.0	05/09/2018	12:25

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.33

LB95223

WORKLIST(Hardcopy Internal Chain)

WorkList Name : FP-050918 WorkList ID : 111825 Date : 5/9/2018 8:12:32 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
05/11/2018	Solid	J1578-14 A	Flash Point	Cool 4 deg C	PSEG02	I22	OILY-WATER	05/08/2018	1010

Date/Time 05/09/18 6:30 AM
Received by: KS
Relinquished by: CP

Date/Time 05/09/18 12:30 PM
Received by: CP
Relinquished by: KS