

=====

Test results

Aquakem 7.2AQ1

LB95224
Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

5/9/2018 14:59

Test: Total CN ug/l

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	88.470	0.0	0.089	
ICB1	0.447	0.0	0.004	
CCV1	242.377	0.0	0.237	
CCB1	0.774	0.0	0.004	
PB109095BL	-0.031	0.0	0.003	
PB109095BS	89.240	0.0	0.089	
J2803-06	1.240	0.0	0.005	
J2803-06DUP	1.475	0.0	0.005	
J2803-06MS	37.557	0.0	0.040	
J2803-06MSD	36.953	0.0	0.039	
J2811-13	1.918	0.0	0.005	
CCV2	260.385	0.0	0.254	
CCB2	0.556	0.0	0.004	

N	13
Mean	58.566
SD	91.5296
CV%	156.28

Aquakem v. 7.2AQ1

Results from time period:

Wed May 09 14:40:41 2018

Wed May 09 14:44:05 2018

AS

Sample Id	Sam/Ctr/c/	Test short	Test type	Result	Result unit	Result date and time	Stat
0.OPPBCN	A	Total CN	P	-0.012	µg/l	5/9/2018 11:21:46	
5.OPPBCN	A	Total CN	P	5.3383	µg/l	5/9/2018 11:21:47	
10PPBCN	A	Total CN	P	9.719	µg/l	5/9/2018 11:21:48	
50PPBCN	A	Total CN	P	52.3918	µg/l	5/9/2018 11:21:49	
100PPBCN	A	Total CN	P	93.7579	µg/l	5/9/2018 11:21:50	
250PPBCN	A	Total CN	P	255.587	µg/l	5/9/2018 11:21:51	
500PPBCN	A	Total CN	P	498.218	µg/l	5/9/2018 11:21:52	
ICV1	S	Total CN	P	88.4697	µg/l	5/9/2018 14:40:41	
ICB1	S	Total CN	P	0.4472	µg/l	5/9/2018 14:40:42	
CCV1	S	Total CN	P	242.3774	µg/l	5/9/2018 14:40:43	
CCB1	S	Total CN	P	0.7743	µg/l	5/9/2018 14:40:44	
PB109095BL	S	Total CN	P	-0.0306	µg/l	5/9/2018 14:40:45	
PB109095BS	S	Total CN	P	89.2398	µg/l	5/9/2018 14:40:46	
J2803-06	S	Total CN	P	1.2403	µg/l	5/9/2018 14:40:47	
J2803-06DUP	S	Total CN	P	1.475	µg/l	5/9/2018 14:40:48	
J2803-06MS	S	Total CN	P	37.557	µg/l	5/9/2018 14:40:49	
J2803-06MSD	S	Total CN	P	36.9528	µg/l	5/9/2018 14:40:50	
J2811-13	S	Total CN	P	1.9177	µg/l	5/9/2018 14:40:51	
CCV2	S	Total CN	P	260.3847	µg/l	5/9/2018 14:44:04	
CCB2	S	Total CN	P	0.5555	µg/l	5/9/2018 14:44:05	

CHEMTECH
 284 Sheffield Street,
 Mountainside, NJ 07092
 Reviewed by : AS

5/9/2018 11:38


Test Total CN

Accepted 5/9/2018 11:22

Factor 1038
 Bias 0.003

Coeff. of det. 0.999614

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.003	-0.0120	0.0000	
2	5.0PPBCN	0.009	5.3383	5.0000	
3	10PPBCN	0.013	9.7190	10.0000	
4	50PPBCN	0.054	52.3918	50.0000	
5	100PPBCN	0.094	93.7579	100.0000	
6	250PPBCN	0.250	255.5870	250.0000	
7	500PPBCN	0.484	498.2180	500.0000	