

245.1


LB 95232

INSTRUMENT ID: CV1

Linear

 μ Abs.:

35587



A= 0.0000e+000

B= 2.8146e-004

C= -3.5669e-002

Rho= 0.9999878

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.005	0.005	143	0.000	143				
0.05	0.050	0.051	0.001	307	0.0 %	307				
0.2	0.200	0.207	0.007	861	0.0 %	861				
2.5	2.500	2.470	-0.030	8902	0.0 %	8902				
5.0	5.000	5.006	0.006	17914	0.0 %	17914				
7.5	7.500	7.531	0.031	26884	0.0 %	26884				
10.0	10.000	9.981	-0.019	35587	0.0 %	35587				

LB95232 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 09 May 2018 16:15:42

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	SC	143	-	0.0000	245.1	PPB	09 May 2018 16:18:21	S	Std
0.05 - 1	50.05	307	-	0.0500	245.1	PPB	09 May 2018 16:20:29	S	Std
0.2 - 1	50.2	861	-	0.2000	245.1	PPB	09 May 2018 16:22:38	S	Std
2.5 - 1	52.5	8902	-	2.5000	245.1	PPB	09 May 2018 16:24:46	S	Std
5.0 - 1	55	17914	-	5.0000	245.1	PPB	09 May 2018 16:26:53	S	Std
7.5 - 1	57.5	26884	-	7.5000	245.1	PPB	09 May 2018 16:29:01	S	Std
10.0 - 1	510	35587	-	10.0000	245.1	PPB	09 May 2018 16:31:09	S	Std
ICV38 - 1	ICV38	13945	3.8893	-245.1		PPB	09 May 2018 16:36:07	U	SMPL
ICB38 - 1	ICB38	-5	-0.0371	-245.1		PPB	09 May 2018 16:38:13	U	SMPL
CCV75 - 1	CCV75	17954	5.0177	-245.1		PPB	09 May 2018 16:40:21	U	SMPL
CCB75 - 1	CCB75	-43	-0.0478	-245.1		PPB	09 May 2018 16:42:27	U	SMPL
CRA - 1	CRA	276	0.0420	-245.1		PPB	09 May 2018 16:44:34	U	SMPL
HighStd - 1	HighStd	34995	9.8140	-245.1		PPB	09 May 2018 16:46:41	U	SMPL
ChkStd - 1	ChkStd	23869	6.6825	-245.1		PPB	09 May 2018 16:48:47	U	SMPL
PB109082BL - 1	PBW	53	-0.0208	-245.1		PPB	09 May 2018 17:05:44	U	SMPL
PB109082BS - 1	LCSW	13832	3.8575	-245.1		PPB	09 May 2018 17:07:51	U	SMPL
J2797-01 - 1	050718-DBRI-E-D-B	-205	-0.0934	-245.1		PPB	09 May 2018 17:09:58	U	SMPL
J2797-02 - 1	050718-DBRI-E-D-M	-95	-0.0624	-245.1		PPB	09 May 2018 17:12:07	U	SMPL
J2797-03 - 1	050718-DBRI-E-D-S	-121	-0.0697	-245.1		PPB	09 May 2018 17:14:15	U	SMPL
J2797-04 - 1	050718-DBRI-E-U-B	-97	-0.0630	-245.1		PPB	09 May 2018 17:16:23	U	SMPL
J2797-05 - 1	050718-DBRI-E-U-M	-129	-0.0720	-245.1		PPB	09 May 2018 17:18:29	U	SMPL
CCV76 - 1	CCV76	17664	4.9360	-245.1		PPB	09 May 2018 17:20:36	U	SMPL
CCB76 - 1	CCB76	-30	-0.0441	-245.1		PPB	09 May 2018 17:22:43	U	SMPL
J2797-06 - 1	050718-DBRI-E-U-S	-150	-0.0779	-245.1		PPB	09 May 2018 17:24:51	U	SMPL
J2797-07 - 1	050718-SIDI-E-D-B	-83	-0.0590	-245.1		PPB	09 May 2018 17:26:58	U	SMPL
J2797-08 - 1	050718-SIDI-E-D-M	-81	-0.0585	-245.1		PPB	09 May 2018 17:29:05	U	SMPL
J2797-09 - 1	050718-SIDI-E-D-S	-93	-0.0618	-245.1		PPB	09 May 2018 17:31:12	U	SMPL
J2797-10 - 1	050718-SIDI-E-U-B	-89	-0.0607	-245.1		PPB	09 May 2018 17:33:19	U	SMPL
J2797-11 - 1	050718-SIDI-E-U-M	-32	-0.0447	-245.1		PPB	09 May 2018 17:35:27	U	SMPL
J2797-12 - 1	050718-SIDI-E-U-S	-113	-0.0675	-245.1		PPB	09 May 2018 17:37:34	U	SMPL
J2799-01 - 1	050818-SIDI-E-D-B	-78	-0.0576	-245.1		PPB	09 May 2018 17:39:43	U	SMPL
J2799-02 - 1	050818-SIDI-E-D-M	-105	-0.0652	-245.1		PPB	09 May 2018 17:41:51	U	SMPL
J2799-03 - 1	050818-SIDI-E-D-S	-109	-0.0663	-245.1		PPB	09 May 2018 17:43:58	U	SMPL
CCV77 - 1	CCV77	17916	5.0070	-245.1		PPB	09 May 2018 17:46:05	U	SMPL
CCB77 - 1	CCB77	-33	-0.0450	-245.1		PPB	09 May 2018 17:48:12	U	SMPL
J2799-04 - 1	050818-SIDI-E-U-B	-104	-0.0649	-245.1		PPB	09 May 2018 17:50:20	U	SMPL
J2799-05 - 1	050818-SIDI-E-U-M	-134	-0.0734	-245.1		PPB	09 May 2018 17:52:27	U	SMPL
J2799-06 - 1	050818-SIDI-E-U-S	-108	-0.0661	-245.1		PPB	09 May 2018 17:54:34	U	SMPL
J2799-06DUP - 1	050818-SIDI-E-U-SDUP	-84	-0.0593	-245.1		PPB	09 May 2018 17:56:41	U	SMPL
J2799-06MS - 1	050818-SIDI-E-U-SMS	12374	3.4471	-245.1		PPB	09 May 2018 17:58:49	U	SMPL
J2799-06MSD - 1	050818-SIDI-E-U-SMSD	12604	3.5118	-245.1		PPB	09 May 2018 18:00:56	U	SMPL
J2799-06LX5 - 1		-55	-0.0511	-245.1		PPB	09 May 2018 18:03:04	U	SMPL
J2799-06A - 1		13602	3.7927	-245.1		PPB	09 May 2018 18:05:12	U	SMPL
CCV78 - 1	CCV78	17791	4.9718	-245.1		PPB	09 May 2018 18:07:20	U	SMPL
CCB78 - 1	CCB78	-36	-0.0458	-245.1		PPB	09 May 2018 18:09:28	U	SMPL