



Analytical Summary Report

Analysis Method: SM5220 D

ANALYST: Indiana

Parameter: COD

SUPERVISOR REVIEW BY: apatel

Run Number: LB95233

Reagent/Standard	Lot/Log #
COD ICV-LCS std, 50ppm	WP63546
COD calibration std. 100 ppm	WP63549
COD calibration std. 10 ppm	WP63544
COD calibration std. 150 ppm	WP63550
COD calibration std. 50 ppm	WP63545
COD calibration std. 0 ppm	WP63543
COD Digestion Vials Low Level 0-150Mg/L	W2346
COD ICV-LCS std, 50ppm	WP64343
COD CCV std, 50ppm	WP64344

Temp In (C): <u>150</u>	Date In: <u>05/09/2018</u>	Time In: <u>16:20</u>
Temp Out (C): <u>150</u>	Date Out: <u>05/09/2018</u>	Time Out: <u>18:20</u>

Intercept: -0.0793

Slope: 1.011

Regression: 0.9999

Seq	Lab ID	TrueValue (mg/l)	DF	MATRIX	Reading	Anal Date	Anal Time
1	CAL1	0	1	Water	0.000	04/10/2018	13:10
2	CAL2	10	1	Water	11.000	04/10/2018	13:10
3	CAL3	50	1	Water	49.000	04/10/2018	13:11
4	CAL4	100	1	Water	101.000	04/10/2018	13:11
5	CAL5	150	1	Water	152.000	04/10/2018	13:12



Analytical Summary Report

Analysis Method: SM5220 D

ANALYST: Indiana

Parameter: COD

SUPERVISOR REVIEW BY: apatel

Run Number: LB95233

Seq	Lab ID	True Value (mg/l)	Initial Weight (g)	Final Vol (ml)	DF	MATRIX	Reading	Result	AnalDate	AnalTime
1	ICV	50			1	Water	50.000	49.534	04/10/2018	13:12
2	ICB				1	Water	0.000	0.078	05/10/2018	13:13
3	CCV1	50			1	Water	51.000	50.524	05/09/2018	18:50
4	CCB1				1	Water	0.000	0.078	05/09/2018	18:50
5	LB95233BL				1	Water	0.000	0.078	05/09/2018	18:51
6	LB95233BS	50			1	Water	51.000	50.524	05/09/2018	18:51
7	J1427-03				1	Water	5.000	5.024	05/09/2018	18:52
8	J2787-01				10	Water	38.000	37.665	05/09/2018	18:52
9	J2801-01				10	Water	95.000	94.045	05/09/2018	18:53
10	J2801-02				10	Water	57.000	56.458	05/09/2018	18:53
11	J2803-06				10	Water	25.000	24.806	05/09/2018	18:54
12	J2811-13				1	Water	39.000	38.654	05/09/2018	18:54
13	J2811-13DUP				1	Water	40.000	39.643	05/09/2018	18:55
14	J2811-13MS	50			1	Water	86.000	85.143	05/09/2018	18:55
15	CCV2	50			1	Water	52.000	51.513	05/09/2018	18:56
16	CCB2				1	Water	0.000	0.078	05/09/2018	18:56
17	J2811-13MSD	50			1	Water	87.000	86.132	05/09/2018	18:57
18	CCV3	50			1	Water	52.000	51.513	05/09/2018	18:57
19	CCB3				1	Water	0.000	0.078	05/09/2018	18:58