



Analytical Summary Report

Analysis Method: 9040C

Analyst By : Sweta

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB95240

Slope : 98.6

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.6	4.01	05/09/2018	16:40
2	CAL2	1	Water			21.4	7.00	05/09/2018	16:42
3	CAL3	1	Water			21.5	10.03	05/09/2018	16:44
4	ICV	1	Water			21.5	7.00	05/09/2018	16:46
5	CCV1	1	Water			21.6	2.00	05/09/2018	16:48
6	J1427-04	1	Water			21.6	8.12	05/09/2018	16:50
7	J1427-04DUP	1	Water			21.5	8.13	05/09/2018	16:52
8	CCV2	1	Water			21.5	12.00	05/09/2018	16:54

WORKLIST(Hardcopy Internal Chain)

LB 95240

WorkList Name : corrosivity-1

WorkList ID : 111861

Date : 5/9/2018 4:31:11 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
05/09/2018	Water	J1427-04	Corrosivity	Cool 4 deg C	PSEG01	I32	W-1	05/09/2018	9040C

Date/Time 05/09/18 4:35 PM
 Received by: SP
 Relinquished by: CP

Date/Time 05/09/18 5:00 PM
 Received by: CP
 Relinquished by: SP