


7470A

LB95243
INSTRUMENT ID: CV2

Linear

 μ Abs.:

35918



A= 0.0000e+000

B= 2.8018e-004

C= -6.7380e-002

Rho= 0.9999450

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.058	-0.058	35	0.000	35				
0.2	0.200	0.210	0.010	990	0.0 %	990				
2.5	2.500	2.555	0.055	9358	0.0 %	9358				
5.0	5.000	5.034	0.034	18208	0.0 %	18208				
7.5	7.500	7.463	-0.037	26875	0.0 %	26875				
10.0	10.000	9.996	-0.004	35918	0.0 %	35918				

LB95243 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 10 May 2018 10:07:37

Sample ID	Extended ID	µ Abs.	Conc.	Std CondMethod	Units	Date	Type	Type
0.0 - 1	50.5	35	-	0.00007470A	PPB	10 May 2018 10:12:39	S	Std
0.2 - 1	52.5	990	-	0.20007470A	PPB	10 May 2018 10:14:48	S	Std
2.5 - 1	55	9358	-	2.50007470A	PPB	10 May 2018 10:16:56	S	Std
5.0 - 1	57.5	18208	-	5.00007470A	PPB	10 May 2018 10:21:32	S	Std
7.5 - 1	510	26875	-	7.50007470A	PPB	10 May 2018 10:23:40	S	Std
10.0 - 1		35918	-	10.00007470A	PPB	10 May 2018 10:25:47	S	Std
ICV39 - 1	ICV39	14246	3.9241	-7470A	PPB	10 May 2018 10:28:20	U	SMPL
ICB39 - 1	ICB39	-14	-0.0713	-7470A	PPB	10 May 2018 10:33:23	U	SMPL
CCV79 - 1	CCV79	17085	4.7196	-7470A	PPB	10 May 2018 10:35:29	U	SMPL
CCB79 - 1	CCB79	-2	-0.0679	-7470A	PPB	10 May 2018 10:39:59	U	SMPL
CRA - 1	CRA	983	0.2080	-7470A	PPB	10 May 2018 10:42:05	U	SMPL
HighStd - 1	HighStd	35274	9.8158	-7470A	PPB	10 May 2018 10:44:11	U	SMPL
ChkStd - 1	ChkStd	25473	7.0697	-7470A	PPB	10 May 2018 10:46:18	U	SMPL
PB109096BL - 1	PBW	-9	-0.0699	-7470A	PPB	10 May 2018 10:48:27	U	SMPL
PB109096BS - 1	LCSW	11973	3.2873	-7470A	PPB	10 May 2018 10:50:36	U	SMPL
PB109096TB - 1	PB109058TB	8	-0.0651	-7470A	PPB	10 May 2018 10:52:44	U	SMPL
J1578-06 - 1	OILY-DEBRIS	109	-0.0368	-7470A	PPB	10 May 2018 10:54:51	U	SMPL
J1578-14 - 1	OILY-WATER	103	-0.0385	-7470A	PPB	10 May 2018 10:56:59	U	SMPL
J1585-11 - 1	FMS-SOIL-1	207	-0.0094	-7470A	PPB	10 May 2018 10:59:08	U	SMPL
J1585-21 - 1	FMS-SOIL-2	149	-0.0256	-7470A	PPB	10 May 2018 11:01:15	U	SMPL
CCV80 - 1	CCV80	17535	4.8456	-7470A	PPB	10 May 2018 11:03:22	U	SMPL
CCB80 - 1	CCB80	100	-0.0394	-7470A	PPB	10 May 2018 11:07:53	U	SMPL
J2669-06 - 1	SU-02-050818	140	-0.0282	-7470A	PPB	10 May 2018 11:09:59	U	SMPL
J2669-06DUP - 1	SU-02-050818DUP	155	-0.0240	-7470A	PPB	10 May 2018 11:12:06	U	SMPL
J2669-06MS - 1	SU-02-050818MS	11797	3.2379	-7470A	PPB	10 May 2018 11:14:13	U	SMPL
J2669-06MSD - 1	SU-02-050818MSD	11733	3.2200	-7470A	PPB	10 May 2018 11:16:20	U	SMPL
J2803-02 - 1	LS-OILY-STONE	56	-0.0517	-7470A	PPB	10 May 2018 11:18:28	U	SMPL
J2803-03 - 1	LS-OILY-WIPES	135	-0.0296	-7470A	PPB	10 May 2018 11:20:35	U	SMPL
J2805-05 - 1	WC-11(10-15)D3	156	-0.0237	-7470A	PPB	10 May 2018 11:22:43	U	SMPL
J2805-06 - 1	WC-12(5-10)J4	150	-0.0254	-7470A	PPB	10 May 2018 11:24:51	U	SMPL
J2806-02 - 1	CS-OILY-STONE	144	-0.0270	-7470A	PPB	10 May 2018 11:27:00	U	SMPL
J2806-03 - 1	CS-OILY-WIPES	112	-0.0360	-7470A	PPB	10 May 2018 11:29:07	U	SMPL
CCV81 - 1	CCV81	17538	4.8465	-7470A	PPB	10 May 2018 11:31:14	U	SMPL
CCB81 - 1	CCB81	-14	-0.0713	-7470A	PPB	10 May 2018 11:33:21	U	SMPL
J2807-04 - 1	BLACK-BROWN-SOIL	164	-0.0214	-7470A	PPB	10 May 2018 11:35:29	U	SMPL
J2807-05 - 1	RED-TILE	110	-0.0366	-7470A	PPB	10 May 2018 11:37:36	U	SMPL
J2807-08 - 1	RED-BLACK-CONCRETE-SLAB	147	-0.0262	-7470A	PPB	10 May 2018 11:39:43	U	SMPL
J2807-11 - 1	CONCRETE-FOOTING	19	-0.0621	-7470A	PPB	10 May 2018 11:41:50	U	SMPL
J2811-03 - 1	TP-4A	160	-0.0226	-7470A	PPB	10 May 2018 11:43:58	U	SMPL
J2811-06 - 1	TP-6A	138	-0.0287	-7470A	PPB	10 May 2018 11:46:05	U	SMPL
J2811-09 - 1	TP-6B	199	-0.0116	-7470A	PPB	10 May 2018 11:48:13	U	SMPL
PB109097BL - 1	PBW	117	-0.0346	-7470A	PPB	10 May 2018 11:50:22	U	SMPL
PB109097BS - 1	LCSW	14076	3.8765	-7470A	PPB	10 May 2018 11:52:30	U	SMPL
PB109097TB - 1	PB109060TB	-73	-0.0878	-7470A	PPB	10 May 2018 11:54:37	U	SMPL
CCV82 - 1	CCV82	17654	4.8790	-7470A	PPB	10 May 2018 11:56:44	U	SMPL
CCB82 - 1	CCB82	-69	-0.0867	-7470A	PPB	10 May 2018 11:58:52	U	SMPL
J2803-07 - 1	OILY-WATER	105	-0.0380	-7470A	PPB	10 May 2018 12:01:00	U	SMPL
J2811-14 - 1	TP-6	269	0.0080	-7470A	PPB	10 May 2018 12:03:07	U	SMPL
J2811-14DUP - 1	TP-6DUP	275	0.0097	-7470A	PPB	10 May 2018 12:05:13	U	SMPL
J2811-14MS - 1	TP-6MS	12796	3.5178	-7470A	PPB	10 May 2018 12:07:20	U	SMPL
J2811-14MSD - 1	TP-6MSD	12375	3.3999	-7470A	PPB	10 May 2018 12:09:28	U	SMPL
PB109098BL - 1	PBW	173	-0.0189	-7470A	PPB	10 May 2018 12:11:35	U	SMPL
PB109098BS - 1	LCSW	11750	3.2248	-7470A	PPB	10 May 2018 12:13:44	U	SMPL
J2803-06 - 1	OILY-WATER	43	-0.0553	-7470A	PPB	10 May 2018 12:15:52	U	SMPL
J2811-13 - 1	TP-6	568	0.0918	-7470A	PPB	10 May 2018 12:18:00	U	SMPL
J2811-13DUP - 1	TP-6DUP	533	0.0820	-7470A	PPB	10 May 2018 12:20:07	U	SMPL
CCV83 - 1	CCV83	17400	4.8078	-7470A	PPB	10 May 2018 12:22:14	U	SMPL
CCB83 - 1	CCB83	-21	-0.0733	-7470A	PPB	10 May 2018 12:24:21	U	SMPL
J2811-13MS - 1	TP-6MS	14355	3.9547	-7470A	PPB	10 May 2018 12:26:29	U	SMPL
J2811-13MSD - 1	TP-6MSD	13178	3.6249	-7470A	PPB	10 May 2018 12:28:36	U	SMPL
J2669-06LX5 - 1		33	-0.0581	-7470A	PPB	10 May 2018 12:30:44	U	SMPL
J2669-06A - 1		13844	3.8115	-7470A	PPB	10 May 2018 12:32:52	U	SMPL
J2811-14LX5 - 1		55	-0.0520	-7470A	PPB	10 May 2018 12:34:59	U	SMPL
J2811-14A - 1		14342	3.9510	-7470A	PPB	10 May 2018 12:37:08	U	SMPL
J2811-13LX5 - 1		137	-0.0290	-7470A	PPB	10 May 2018 12:39:16	U	SMPL
J2811-13A - 1		16987	4.6921	-7470A	PPB	10 May 2018 12:41:24	U	SMPL
CCV84 - 1	CCV84	17477	4.8294	-7470A	PPB	10 May 2018 12:43:33	U	SMPL
CCB84 - 1	CCB84	-78	-0.0892	-7470A	PPB	10 May 2018 12:45:41	U	SMPL