


7471B

LB95247
INSTRUMENT ID: CV1

Linear

 μ Abs.:

45250



A= 0.0000e+000

B= 2.2081e-004

C= -1.8229e-002

Rho= 0.9999486

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.021	0.021	177	0.000	177				
0.2	0.200	0.214	0.014	1053	0.0 %	1053				
2.5	2.500	2.439	-0.061	11129	0.0 %	11129				
5.0	5.000	4.994	-0.006	22700	0.0 %	22700				
7.5	7.500	7.558	0.058	34309	0.0 %	34309				
10.0	10.000	9.974	-0.026	45250	0.0 %	45250				

LB95247 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 10 May 2018 12:56:43

Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	177	-	0.0000	7471B	PPB	10 May 2018 12:58:03	S	Std
0.2 - 1	50.2	1053	-	0.2000	7471B	PPB	10 May 2018 13:00:20	S	Std
2.5 - 1	52.5	11129	-	2.5000	7471B	PPB	10 May 2018 13:02:38	S	Std
5.0 - 1	55	22700	-	5.0000	7471B	PPB	10 May 2018 13:04:55	S	Std
7.5 - 1	57.5	34309	-	7.5000	7471B	PPB	10 May 2018 13:07:12	S	Std
10.0 - 1	510	45250	-	10.0000	7471B	PPB	10 May 2018 13:12:20	S	Std
ICV40 - 1	ICV40	18357	4.0353	-	7471B	PPB	10 May 2018 13:14:59	U	SMPL
ICB40 - 1	ICB40	-16	-0.0218	-	7471B	PPB	10 May 2018 13:17:15	U	SMPL
CCV85 - 1	CCV85	21692	4.7717	-	7471B	PPB	10 May 2018 13:19:31	U	SMPL
CCB85 - 1	CCB85	8	-0.0165	-	7471B	PPB	10 May 2018 13:21:47	U	SMPL
CRA - 1	CRA	1021	0.2072	-	7471B	PPB	10 May 2018 13:24:03	U	SMPL
HighStd - 1	HighStd	45929	10.1236	-	7471B	PPB	10 May 2018 13:26:19	U	SMPL
ChkStd - 1	ChkStd	29090	6.4053	-	7471B	PPB	10 May 2018 13:28:35	U	SMPL
PB109103BL - 1	PBS	53	-0.0065	-	7471B	PPB	10 May 2018 13:30:54	U	SMPL
PB109103BS - 1	LCSS	15920	3.4971	-	7471B	PPB	10 May 2018 13:33:12	U	SMPL
J1427-01 - 1	S-1	10673	2.3385	-	7471B	PPB	10 May 2018 13:35:28	U	SMPL
J2671-05 - 1	AU-06-050918-A	1976	0.4181	-	7471B	PPB	10 May 2018 13:37:45	U	SMPL
J2671-07 - 1	AU-06-050918-B	1406	0.2922	-	7471B	PPB	10 May 2018 13:40:03	U	SMPL
J2671-09 - 1	AU-06-050918-C	1410	0.2931	-	7471B	PPB	10 May 2018 13:42:21	U	SMPL
J2672-03 - 1	BU-01-050918	1502	0.3134	-	7471B	PPB	10 May 2018 13:44:37	U	SMPL
CCV86 - 1	CCV86	21902	4.8181	-	7471B	PPB	10 May 2018 13:46:52	U	SMPL
CCB86 - 1	CCB86	37	-0.0101	-	7471B	PPB	10 May 2018 13:49:08	U	SMPL
J2676-11 - 1	CL-02-050918-A	2011	0.4258	-	7471B	PPB	10 May 2018 13:51:25	U	SMPL
J2676-13 - 1	CL-02-050918-B	4040	0.8739	-	7471B	PPB	10 May 2018 13:53:41	U	SMPL
J2676-15 - 1	CL-02-050918-C	2771	0.5936	-	7471B	PPB	10 May 2018 13:55:57	U	SMPL
J2676-17 - 1	CL-02-050918-D	3615	0.7800	-	7471B	PPB	10 May 2018 13:58:13	U	SMPL
J2676-19 - 1	CL-02-050918-E	2808	0.6018	-	7471B	PPB	10 May 2018 14:00:30	U	SMPL
J2676-19DUP - 1	CL-02-050918-EDUP	2511	0.5362	-	7471B	PPB	10 May 2018 14:02:47	U	SMPL
J2676-19MS - 1	CL-02-050918-EMS	19577	4.3047	-	7471B	PPB	10 May 2018 14:05:04	U	SMPL
J2676-19MSD - 1	CL-02-050918-EMSD	19582	4.3058	-	7471B	PPB	10 May 2018 14:07:21	U	SMPL
J2678-01 - 1	OR-02-050918	1485	0.3097	-	7471B	PPB	10 May 2018 14:09:39	U	SMPL
J2807-01 - 1	BLACK-BROWN-SOIL	2224	0.4729	-	7471B	PPB	10 May 2018 14:11:56	U	SMPL
CCV87 - 1	CCV87	22184	4.8803	-	7471B	PPB	10 May 2018 14:14:12	U	SMPL
CCB87 - 1	CCB87	-12	-0.0209	-	7471B	PPB	10 May 2018 14:16:28	U	SMPL
J2807-06 - 1	RED-BLACK-CONCRETE-SLAB	1193	0.2452	-	7471B	PPB	10 May 2018 14:18:45	U	SMPL
J2807-09 - 1	CONCRETE-FOOTING	993	0.2010	-	7471B	PPB	10 May 2018 14:21:01	U	SMPL
J2818-01 - 1	LS-RO-2711993	2054	0.4353	-	7471B	PPB	10 May 2018 14:23:18	U	SMPL
J2836-02 - 1	TP-9A	1557	0.3256	-	7471B	PPB	10 May 2018 14:25:34	U	SMPL
J2836-05 - 1	TP-9B	6195	1.3497	-	7471B	PPB	10 May 2018 14:27:51	U	SMPL
PB109106BL - 1	PBS	112	0.0065	-	7471B	PPB	10 May 2018 14:30:08	U	SMPL
PB109106BS - 1	LCSS	15895	3.4916	-	7471B	PPB	10 May 2018 14:32:26	U	SMPL
J2822-01 - 1	SB02-01	217	0.0297	-	7471B	PPB	10 May 2018 14:34:43	U	SMPL
J2822-01DUP - 1	SB02-01DUP	284	0.0445	-	7471B	PPB	10 May 2018 14:37:01	U	SMPL
J2822-01MS - 1	SB02-01MS	16767	3.6842	-	7471B	PPB	10 May 2018 14:39:17	U	SMPL
CCV88 - 1	CCV88	22416	4.9316	-	7471B	PPB	10 May 2018 14:41:34	U	SMPL
CCB88 - 1	CCB88	-44	-0.0279	-	7471B	PPB	10 May 2018 14:43:50	U	SMPL
J2822-01MSD - 1	SB02-01MSD	15994	3.5135	-	7471B	PPB	10 May 2018 14:46:07	U	SMPL
J2822-02 - 1	SB02-02	168	0.0189	-	7471B	PPB	10 May 2018 14:48:23	U	SMPL
J2822-03 - 1	SB01-01	761	0.1498	-	7471B	PPB	10 May 2018 14:50:40	U	SMPL
J2822-04 - 1	SB01-02	348	0.0586	-	7471B	PPB	10 May 2018 14:52:56	U	SMPL
J2822-08 - 1	SB04-01	2848	0.6107	-	7471B	PPB	10 May 2018 14:55:13	U	SMPL
J2822-09 - 1	SB04-02	1085	0.2214	-	7471B	PPB	10 May 2018 14:57:30	U	SMPL
J2823-01 - 1	GREENSAND-SUPPLY-A	3746	0.8089	-	7471B	PPB	10 May 2018 14:59:47	U	SMPL
J2823-02 - 1	GREENSAND-SUPPLY-B	8956	1.9594	-	7471B	PPB	10 May 2018 15:02:05	U	SMPL
J2825-01 - 1	FK-PS-01-025	628	0.1204	-	7471B	PPB	10 May 2018 15:04:23	U	SMPL
J2828-01 - 1	FK-PS-01.025	1464	0.3050	-	7471B	PPB	10 May 2018 15:06:39	U	SMPL
CCV89 - 1	CCV89	22636	4.9801	-	7471B	PPB	10 May 2018 15:08:55	U	SMPL
CCB89 - 1	CCB89	-19	-0.0224	-	7471B	PPB	10 May 2018 15:11:12	U	SMPL
J2837-01 - 1	EPE-F-7(0-5)	772	0.1522	-	7471B	PPB	10 May 2018 15:13:29	U	SMPL
J2837-02 - 1	EPE-F-7(5-10)	1144	0.2344	-	7471B	PPB	10 May 2018 15:15:45	U	SMPL
J2676-19LX5 - 1		570	0.1076	-	7471B	PPB	10 May 2018 15:18:01	U	SMPL
J2676-19A - 1		20406	4.4877	-	7471B	PPB	10 May 2018 15:20:18	U	SMPL
J2822-01LX5 - 1		28	-0.0120	-	7471B	PPB	10 May 2018 15:22:34	U	SMPL
J2822-01A - 1		17624	3.8734	-	7471B	PPB	10 May 2018 15:24:52	U	SMPL
CCV90 - 1	CCV90	22115	4.8651	-	7471B	PPB	10 May 2018 15:27:09	U	SMPL
CCB90 - 1	CCB90	4	-0.0173	-	7471B	PPB	10 May 2018 15:29:26	U	SMPL