



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Sweta

Parameter: pH

Supervisor Review By : apatel

Run Number: LB95248

Slope : 99.0

BalanceID: WC SC-4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.5	4.01	05/10/2018	12:00
2	CAL2	1	Water			21.4	7.00	05/10/2018	12:02
3	CAL3	1	Water			21.3	10.03	05/10/2018	12:04
4	ICV	1	Water			21.5	7.00	05/10/2018	12:06
5	CCV1	1	Water			21.6	2.00	05/10/2018	12:08
6	J2823-01	1	Solid	20.02	20	20.6	7.06	05/10/2018	12:10
7	J2823-01DUP	1	Solid	20.03	20	20.7	7.07	05/10/2018	12:12
8	J2823-02	1	Solid	20.05	20	20.2	7.41	05/10/2018	12:14
9	J2825-01	1	Solid	20.02	20	21.4	5.79	05/10/2018	12:16
10	J2828-01	1	Solid	20.04	20	20.6	5.38	05/10/2018	12:18
11	CCV2	1	Water			21.5	12.00	05/10/2018	12:20

WORKLIST(Hardcopy Internal Chain)

LB95248

WorkList Name : pH050918-1

WorkList ID : 111865

Date : 5/9/2018 5:06:08 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
05/08/2018	Solid	J2823-01	pH	Cool 4 deg C	YANN01	J63	GREENSAND-SUPPLY-A	05/08/2018	9045C
05/08/2018	Solid	J2823-02	pH	Cool 4 deg C	YANN01	J63	GREENSAND-SUPPLY-B	05/08/2018	9045C
05/09/2018	Solid	J2825-01	pH	Cool 4 deg C	TULL02	J23	FK-PS-01-025	05/09/2018	9045C
05/09/2018	Solid	J2828-01	pH	Cool 4 deg C	TULL02	J41	FK-PS-01.025	05/09/2018	9045C

Date/Time 05/10/18 11:50 AM

Received by: SP

Relinquished by: CP

Date/Time 05/10/18 12:30 PM

Received by: CP

Relinquished by: SP