



Analytical Summary Report

Analysis Method: SM 4500-PH B

Analyst By : Sweta

Parameter: pH

Supervisor Review By : apatel

Run Number: LB95274

Slope : 99.0

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.5	4.01	05/10/2018	12:20
2	CAL2	1	Water			21.4	7.00	05/10/2018	12:14
3	CAL3	1	Water			21.6	10.03	05/10/2018	12:16
4	ICV	1	Water			21.5	7.00	05/10/2018	12:18
5	CCV1	1	Water			21.5	2.00	05/10/2018	12:20
6	J2851-01	1	Water			19.6	4.53	05/10/2018	12:22
7	J2851-01DUP	1	Water			19.5	4.54	05/10/2018	12:24
8	CCV2	1	Water			21.6	12.00	05/10/2018	12:26