



Analytical Summary Report

Analysis Method: 9040C

Analyst By : Sweta

Parameter: pH

Supervisor Review By : apatel

Run Number: LB95275

Slope : 99.2

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.5	4.01	05/10/2018	17:30
2	CAL2	1	Water			21.4	7.00	05/10/2018	17:32
3	CAL3	1	Water			21.6	10.03	05/10/2018	17:34
4	ICV	1	Water			21.4	7.00	05/10/2018	17:36
5	CCV1	1	Water			21.5	2.00	05/10/2018	17:38
6	J2876-17	1	Water			20.6	6.55	05/10/2018	17:40
7	J2876-17DUP	1	Water			20.5	6.56	05/10/2018	17:42
8	CCV2	1	Water			21.6	12.00	05/10/2018	17:44

WORKLIST(Hardcopy Internal Chain)

LB95245

WorkList Name : pHJ2876

WorkList ID : 111935

Date : 5/10/2018 5:15:02 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
05/10/2018	Water	J2876-17	pH	Cool 4 deg C	PSEG01	I43	TP-1	05/10/2018	9040C

Date/Time

05/10/18 5:20 PM

Received by:

SP

Relinquished by:

SP

Date/Time

05/10/18 5:50 PM

Received by:

SP

Relinquished by:

SP