

7470A


LB 95287

INSTRUMENT ID: CV1

Linear

 μ Abs.:

38932



A= 0.0000e+000

B= 2.5782e-004

C= -3.8179e-002

Rho= 0.9999463

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.045	-0.045	-26	0.000	-26				
0.2	0.200	0.198	-0.002	917	0.0 %	917				
2.5	2.500	2.547	0.047	10029	0.0 %	10029				
5.0	5.000	5.047	0.047	19726	0.0 %	19726				
7.5	7.500	7.453	-0.047	29055	0.0 %	29055				
10.0	10.000	9.999	-0.001	38932	0.0 %	38932				

LB95287 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 11 May 2018 09:29:57

Sample ID	Extended ID	μ Abs.	Conc.	Std. Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-26	-	0.0000	7470A	PPB	11 May 2018 09:37:57	S	Std
0.2 - 1	50.2	917	-	0.2000	7470A	PPB	11 May 2018 09:40:05	S	Std
2.5 - 1	52.5	10029	-	2.5000	7470A	PPB	11 May 2018 09:42:13	S	Std
5.0 - 1	55	19726	-	5.0000	7470A	PPB	11 May 2018 09:52:39	S	Std
7.5 - 1	57.5	29055	-	7.5000	7470A	PPB	11 May 2018 09:54:47	S	Std
10.0 - 1	510	38932	-	10.0000	7470A	PPB	11 May 2018 09:56:55	S	Std
ICV42 - 1	ICV42	14704	3.7527	-	7470A	PPB	11 May 2018 10:00:00	U	SMPL
ICB42 - 1	ICB42	-185	-0.0859	-	7470A	PPB	11 May 2018 10:02:06	U	SMPL
CCV01 - 1	CCV01	19410	4.9660	-	7470A	PPB	11 May 2018 10:04:14	U	SMPL
CCB01 - 1	CCB01	-189	-0.0869	-	7470A	PPB	11 May 2018 10:06:20	U	SMPL
CRA - 1	CRA	911	0.1967	-	7470A	PPB	11 May 2018 10:08:28	U	SMPL
HighStd - 1	HighStd	38756	9.9537	-	7470A	PPB	11 May 2018 10:10:34	U	SMPL
ChkStd - 1	ChkStd	27364	7.0167	-	7470A	PPB	11 May 2018 10:12:41	U	SMPL
PB109148BL - 1	PBW	-12	-0.0413	-	7470A	PPB	11 May 2018 10:27:14	U	SMPL
PB109148BS - 1	LCSW	14722	3.7574	-	7470A	PPB	11 May 2018 10:29:21	U	SMPL
PB109148TB - 1	PB109101TB	-128	-0.0712	-	7470A	PPB	11 May 2018 10:31:28	U	SMPL
J1427-02 - 1	S-1	-62	-0.0542	-	7470A	PPB	11 May 2018 10:33:37	U	SMPL
J2671-06 - 1	AU-06-050918-A	-37	-0.0477	-	7470A	PPB	11 May 2018 10:35:45	U	SMPL
J2671-08 - 1	AU-06-050918-B	-18	-0.0428	-	7470A	PPB	11 May 2018 10:37:54	U	SMPL
J2671-10 - 1	AU-06-050918-C	9	-0.0359	-	7470A	PPB	11 May 2018 10:40:00	U	SMPL
CCV02 - 1	CCV02	19168	4.9036	-	7470A	PPB	11 May 2018 10:42:07	U	SMPL
CCB02 - 1	CCB02	-199	-0.0895	-	7470A	PPB	11 May 2018 10:44:13	U	SMPL
J2672-04 - 1	BU-01-050918	-6	-0.0397	-	7470A	PPB	11 May 2018 10:46:21	U	SMPL
J2676-12 - 1	CL-02-050918-A	-21	-0.0436	-	7470A	PPB	11 May 2018 10:48:28	U	SMPL
J2676-14 - 1	CL-02-050918-B	-32	-0.0464	-	7470A	PPB	11 May 2018 10:50:34	U	SMPL
J2676-16 - 1	CL-02-050918-C	-22	-0.0439	-	7470A	PPB	11 May 2018 10:52:41	U	SMPL
J2676-18 - 1	CL-02-050918-D	-58	-0.0531	-	7470A	PPB	11 May 2018 10:54:49	U	SMPL
J2676-20 - 1	CL-02-050918-E	-5	-0.0395	-	7470A	PPB	11 May 2018 10:56:56	U	SMPL
J2678-02 - 1	OR-02-050918	-21	-0.0436	-	7470A	PPB	11 May 2018 10:59:05	U	SMPL
J2818-02 - 1	LS-RO-2711993	-17	-0.0426	-	7470A	PPB	11 May 2018 11:01:13	U	SMPL
J2836-03 - 1	TP-9A	-24	-0.0444	-	7470A	PPB	11 May 2018 11:03:21	U	SMPL
J2836-06 - 1	TP-9B	38	-0.0284	-	7470A	PPB	11 May 2018 11:05:28	U	SMPL
CCV03 - 1	CCV03	18894	4.8330	-	7470A	PPB	11 May 2018 11:07:35	U	SMPL
CCB03 - 1	CCB03	-174	-0.0830	-	7470A	PPB	11 May 2018 11:09:41	U	SMPL
J2836-06DUP - 1	TP-9BDUP	-8	-0.0402	-	7470A	PPB	11 May 2018 11:11:49	U	SMPL
J2836-06MS - 1	TP-9BMS	12996	3.3124	-	7470A	PPB	11 May 2018 11:13:56	U	SMPL
J2836-06MSD - 1	TP-9BMSD	13122	3.3449	-	7470A	PPB	11 May 2018 11:16:02	U	SMPL
PB109149BL - 1	PBW	-182	-0.0851	-	7470A	PPB	11 May 2018 11:18:10	U	SMPL
PB109149BS - 1	LCSW	14486	3.6965	-	7470A	PPB	11 May 2018 11:20:19	U	SMPL
PB109149TB - 1	PB109104TB	-164	-0.0805	-	7470A	PPB	11 May 2018 11:22:26	U	SMPL
J1427-04 - 1	W-1	-48	-0.0506	-	7470A	PPB	11 May 2018 11:24:35	U	SMPL
J1427-04DUP - 1	W-1DUP	-28	-0.0454	-	7470A	PPB	11 May 2018 11:26:43	U	SMPL
J1427-04MS - 1	W-1MS	13087	3.3358	-	7470A	PPB	11 May 2018 11:28:52	U	SMPL
J1427-04MSD - 1	W-1MSD	13011	3.3163	-	7470A	PPB	11 May 2018 11:30:59	U	SMPL
CCV04 - 1	CCV04	18858	4.8237	-	7470A	PPB	11 May 2018 11:33:07	U	SMPL
CCB04 - 1	CCB04	-160	-0.0794	-	7470A	PPB	11 May 2018 11:35:15	U	SMPL
PB109150BL - 1	PBW	-32	-0.0464	-	7470A	PPB	11 May 2018 11:37:23	U	SMPL
PB109150BS - 1	LCSW	14853	3.7911	-	7470A	PPB	11 May 2018 11:39:30	U	SMPL
J1427-03 - 1	W-1	65	-0.0214	-	7470A	PPB	11 May 2018 11:46:12	U	SMPL
J1427-03DUP - 1	W-1DUP	-50	-0.0511	-	7470A	PPB	11 May 2018 11:48:19	U	SMPL
J1427-03MS - 1	W-1MS	16684	4.2632	-	7470A	PPB	11 May 2018 11:50:26	U	SMPL
J1427-03MSD - 1	W-1MSD	16301	4.1645	-	7470A	PPB	11 May 2018 11:52:34	U	SMPL
J2822-05 - 1	MW01	473	0.0838	-	7470A	PPB	11 May 2018 11:54:41	U	SMPL
CCV05 - 1	CCV05	18583	4.7528	-	7470A	PPB	11 May 2018 11:56:50	U	SMPL
CCB05 - 1	CCB05	-145	-0.0756	-	7470A	PPB	11 May 2018 11:58:58	U	SMPL
J2836-06LX5 - 1		-17	-0.0426	-	7470A	PPB	11 May 2018 12:01:07	U	SMPL
J2836-06A - 1		13589	3.4653	-	7470A	PPB	11 May 2018 12:03:14	U	SMPL
J1427-04LX5 - 1		-118	-0.0686	-	7470A	PPB	11 May 2018 12:05:21	U	SMPL
J1427-04A - 1		13505	3.4436	-	7470A	PPB	11 May 2018 12:07:29	U	SMPL
J1427-03LX5 - 1		8	-0.0361	-	7470A	PPB	11 May 2018 12:12:06	U	SMPL
J1427-03A - 1		16402	4.1905	-	7470A	PPB	11 May 2018 12:14:13	U	SMPL
CCV06 - 1	CCV06	18594	4.7556	-	7470A	PPB	11 May 2018 12:16:20	U	SMPL
CCB06 - 1	CCB06	-192	-0.0877	-	7470A	PPB	11 May 2018 12:18:27	U	SMPL