



Analytical Summary Report

Analysis Method: 9040C

Analyst By : Sweta

Parameter: pH

Supervisor Review By : apatel

Run Number: LB95312

Slope : 99.0

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.4	4.01	05/11/2018	17:15
2	CAL2	1	Water			21.5	7.00	05/11/2018	17:17
3	CAL3	1	Water			21.6	10.03	05/11/2018	17:19
4	ICV	1	Water			21.5	7.00	05/11/2018	17:21
5	CCV1	1	Water			21.4	2.00	05/11/2018	17:23
6	J2877-01	1	Water			19.1	6.80	05/11/2018	17:25
7	J2877-01DUP	1	Water			19.2	6.81	05/11/2018	17:27
8	J2883-03	1	Water			21.1	7.12	05/11/2018	17:29
9	CCV2	1	Water			21.4	12.00	05/11/2018	17:31

WORKLIST(Hardcopy Internal Chain)

LB95312

WorkList Name : ph051118

WorkList ID : 111996

Date : 5/11/2018 5:02:28 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
05/11/2018	Water	J2877-01	pH	Cool 4 deg C	PSEG01	I43	MS-WATER	05/11/2018	9040C
05/11/2018	Water	J2883-03	pH	Cool 4 deg C	PSEG01	I51	DRUM-1-5	05/11/2018	9040C

Date/Time

05/11/18 5:10 PM

Received by:

SP

Relinquished by:

CP

Date/Time

05/11/18 5:40 PM

Received by:

CP

Relinquished by:

SP