



Analytical Summary Report

Analysis Method: 9040C

Analyst By : Sweta

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB95313

Slope : 99.0

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.4	4.01	05/11/2018	17:15
2	CAL2	1	Water			21.5	7.00	05/11/2018	17:17
3	CAL3	1	Water			21.6	10.03	05/11/2018	17:19
4	ICV	1	Water			21.5	7.00	05/11/2018	17:21
5	CCV1	1	Water			21.4	2.00	05/11/2018	17:23
6	J2877-02	1	Water			19.1	6.80	05/11/2018	17:25
7	J2877-02DUP	1	Water			19.2	6.81	05/11/2018	17:27
8	J2883-04	1	Water			21.1	7.12	05/11/2018	17:29
9	CCV2	1	Water			21.4	12.00	05/11/2018	17:31