



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB95315

Slope : 97.8

BalanceID: WC SC-3

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.4	4.01	05/11/2018	17:35
2	CAL2	1	Water			21.4	7.00	05/11/2018	17:36
3	CAL3	1	Water			21.4	10.02	05/11/2018	17:37
4	ICV	1	Water			21.3	7.01	05/11/2018	17:38
5	CCV1	1	Water			21.4	2.01	05/11/2018	17:40
6	J2671-12	1	Solid	20.02	20	19.9	8.91	05/11/2018	17:42
7	J2671-12DUP	1	Solid	20.03	20	19.8	8.92	05/11/2018	17:43
8	J2671-14	1	Solid	20.04	20	19.6	9.35	05/11/2018	17:45
9	J2880-03	1	Solid	20.02	20	21.6	7.55	05/11/2018	17:46
10	J2880-06	1	Solid	20.03	20	21.6	8.45	05/11/2018	17:48
11	J2883-02	1	Solid	20.04	20	21.5	7.10	05/11/2018	17:50
12	J2888-04	1	Solid	20.02	20	21.5	9.32	05/11/2018	17:51
13	J2888-08	1	Solid	20.04	20	18.5	11.22	05/11/2018	18:00
14	J2888-12	1	Solid	20.01	20	18.4	10.56	05/11/2018	18:01
15	J2888-16	1	Solid	20.04	20	18.4	11.14	05/11/2018	18:02
16	CCV2	1	Water			21.2	12.01	05/11/2018	18:04
17	J2888-20	1	Solid	20.04	20	18.5	11.30	05/11/2018	18:05
18	J2888-24	1	Solid	20.02	20	18.4	11.17	05/11/2018	18:06
19	J2888-28	1	Solid	20.04	20	18.5	11.26	05/11/2018	18:09
20	J2888-32	1	Solid	20.02	20	18.3	11.34	05/11/2018	18:10

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
21	J2899-03	1	Solid	20.03	20	19.2	8.69	05/11/2018	18:11
22	J2899-07	1	Solid	20.03	20	18.9	7.71	05/11/2018	18:12
23	J2899-10	1	Solid	20.04	20	18.4	6.62	05/11/2018	18:14
24	J2899-10DUP	1	Solid	20.05	20	18.5	6.63	05/11/2018	18:15
25	CCV3	1	Water			21.4	2.02	05/11/2018	18:20

2BA5315

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosivity J2899

WorkList ID : 111997

Date : 5/11/2018 5:24:31 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
05/11/2018	Solid	J2671-12	Corrosivity	Cool 4 deg C	PSEG05		AU-02-051118-A	05/11/2018	9045C
05/11/2018	Solid	J2671-14	Corrosivity	Cool 4 deg C	PSEG05		AU-02-051118-B	05/11/2018	9045C
05/10/2018	Solid	J2880-03	Corrosivity	Cool 4 deg C	ENVI27	J42	PLANTER-AREA-2-PRECHAF	05/10/2018	9045C
05/10/2018	Solid	J2880-06	Corrosivity	Cool 4 deg C	ENVI27	J42	PLANTER-AREA-3-PRECHAF	05/10/2018	9045C
05/11/2018	Solid	J2883-02	Corrosivity	Cool 4 deg C	PSEG01	I51	DRUM-1-5	05/11/2018	9045C
05/11/2018	Solid	J2888-04	Corrosivity	Cool 4 deg C	PSEG01	I52	RRT-N1	05/11/2018	9045C
05/11/2018	Solid	J2888-08	Corrosivity	Cool 4 deg C	PSEG01	I52	RRT-N2	05/11/2018	9045C
05/11/2018	Solid	J2888-12	Corrosivity	Cool 4 deg C	PSEG01	I52	RRT-S1	05/11/2018	9045C
05/11/2018	Solid	J2888-16	Corrosivity	Cool 4 deg C	PSEG01	I52	RRT-S2	05/11/2018	9045C
05/11/2018	Solid	J2888-20	Corrosivity	Cool 4 deg C	PSEG01	I52	26KV-N1	05/11/2018	9045C
05/11/2018	Solid	J2888-24	Corrosivity	Cool 4 deg C	PSEG01	I52	26KV-N2	05/11/2018	9045C
05/11/2018	Solid	J2888-28	Corrosivity	Cool 4 deg C	PSEG01	I52	26KV-S1	05/11/2018	9045C
05/11/2018	Solid	J2888-32	Corrosivity	Cool 4 deg C	PSEG01	I52	26KV-S2	05/11/2018	9045C
05/11/2018	Solid	J2899-03	Corrosivity	Cool 4 deg C	PSEG01	I53	TP-7A	05/11/2018	9045C
05/11/2018	Solid	J2899-07	Corrosivity	Cool 4 deg C	PSEG01	I53	TP-8A	05/11/2018	9045C
05/11/2018	Solid	J2899-10	Corrosivity	Cool 4 deg C	PSEG01	I53	TP-8B	05/11/2018	9045C

Date/Time 05-11-18 5:25 PM

Received by: Sh

Relinquished by: CP

Date/Time 05-11-18 7:00 PM

Received by: g

Relinquished by: JP