



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB95316

Slope : 97.5

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.2	4.01	05/11/2018	18:20
2	CAL2	1	Water			21.2	7.01	05/11/2018	18:21
3	CAL3	1	Water			21.2	10.02	05/11/2018	18:22
4	ICV	1	Water			21.2	6.99	05/11/2018	18:24
5	CCV1	1	Water			21.3	2.02	05/11/2018	18:25
6	J2898-01	1	Solid	20.02	20	21.6	10.69	05/11/2018	18:30
7	J2898-01DUP	1	Solid	20.01	20	21.7	10.70	05/11/2018	18:31
8	J2898-02	1	Solid	20.03	20	21.4	10.90	05/11/2018	18:32
9	J2898-03	1	Solid	20.03	20	21.1	10.80	05/11/2018	18:34
10	J2898-04	1	Solid	20.02	20	20.9	9.80	05/11/2018	18:35
11	J2898-05	1	Solid	20.04	20	20.5	10.74	05/11/2018	18:36
12	J2898-06	1	Solid	20.02	20	20.7	10.87	05/11/2018	18:37
13	J2898-07	1	Solid	20.05	20	20.6	10.85	05/11/2018	18:39
14	CCV2	1	Water			21.3	12.01	05/11/2018	18:40

2505316

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH J2898

WorkList ID : 112000

Date : 5/11/2018 6:08:26 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
05/11/2018	Solid	J2898-01	pH	Cool 4 deg C	AREC01	J23	SP-01	05/11/2018	9045C
05/11/2018	Solid	J2898-02	pH	Cool 4 deg C	AREC01	J23	SP-02	05/11/2018	9045C
05/11/2018	Solid	J2898-03	pH	Cool 4 deg C	AREC01	J23	SP-03	05/11/2018	9045C
05/11/2018	Solid	J2898-04	pH	Cool 4 deg C	AREC01	J23	SP-04	05/11/2018	9045C
05/11/2018	Solid	J2898-05	pH	Cool 4 deg C	AREC01	J23	SP-05	05/11/2018	9045C
05/11/2018	Solid	J2898-06	pH	Cool 4 deg C	AREC01	J23	SP-06	05/11/2018	9045C
05/11/2018	Solid	J2898-07	pH	Cool 4 deg C	AREC01	J23	SP-07	05/11/2018	9045C

Date/Time 05-11-18 6:08 PM

Received by: Ju

Relinquished by: SW

Date/Time 05-11-18 7:20 PM

Received by: SW

Relinquished by: Ju