



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: KETANKUMAR

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB95317

Ambient Barometric Pressure (mmHg): 755

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	27.0	80.6	82.3	05/11/2018	09:10
2	J2832-01		1	>100.0	>212.0	>212.0	05/11/2018	09:35
3	J2832-01DUP		1	>100.0	>212.0	>212.0	05/11/2018	09:55
4	J2832-03		1	>100.0	>212.0	>212.0	05/11/2018	10:20

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.33

LB95317

WORKLIST(Hardcopy Internal Chain)

WorkList Name : FP-051118

WorkList ID : 111954

Date : 5/11/2018 8:02:11 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
05/13/2018	Solid	J2832-01 B	Flash Point	Cool 4 deg C	PSEG01	I41	WASTE-OIL	05/09/2018	1010A
05/13/2018	Solid	J2832-03 B	Flash Point	Cool 4 deg C	PSEG01	I41	PRODUCT-OIL	05/09/2018	1010A

Date/Time 05/11/18 8:30 AM

Received by: KS

Relinquished by: CP

Date/Time 05/11/18 10:25 AM

Received by: CP

Relinquished by: KS