



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: KETANKUMAR

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB95318

Ambient Barometric Pressure (mmHg): 755

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	27.0	80.6	82.3	05/11/2018	10:55
2	J2876-18		1	>100.0	>212.0	>212.0	05/11/2018	11:30
3	J2876-18DUP		1	>100.0	>212.0	>212.0	05/11/2018	12:05

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.33

LB 95318

WORKLIST(Hardcopy Internal Chain)

WorkList Name : FPW-051118

WorkList ID : 111955

Date : 5/11/2018 8:03:28 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
05/15/2018	Water	J2876-18	C Flash Point	Cool 4 deg C	PSEG01	I43	TP-1	05/10/2018	1010A

Date/Time 05/11/18 8:30 AM
Received by: KS
Relinquished by: CP

Date/Time 05/11/18 12:10 PM
Received by: CP
Relinquished by: KS