



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB95320

Slope : 98.2

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.2	4.01	05/14/2018	15:00
2	CAL2	1	Water			21.2	7.01	05/12/2018	15:01
3	CAL3	1	Water			21.2	10.03	05/12/2018	15:02
4	ICV	1	Water			21.3	6.99	05/12/2018	15:05
5	CCV1	1	Water			21.3	2.01	05/12/2018	15:06
6	J2670-04	1	Solid	20.02	20	21.7	9.00	05/12/2018	15:08
7	J2674-04	1	Solid	20.03	20	20.3	6.45	05/12/2018	15:10
8	J2679-08	1	Solid	20.04	20	19.3	9.53	05/12/2018	15:11
9	J2679-10	1	Solid	20.05	20	19.1	5.60	05/12/2018	15:12
10	J2679-12	1	Solid	20.01	20	19.4	9.60	05/12/2018	15:15
11	J2901-09	1	Solid	20.03	20	20.7	8.02	05/12/2018	15:16
12	J2901-10	1	Solid	20.04	20	20.5	7.53	05/12/2018	15:18
13	J2901-11	1	Solid	20.02	20	20.4	7.83	05/12/2018	15:20
14	J2901-12	1	Solid	20.05	20	20.5	8.71	05/12/2018	15:21
15	J2902-05	1	Solid	20.03	20	20.9	10.69	05/12/2018	15:22
16	CCV2	1	Water			21.3	12.02	05/12/2018	15:25
17	J2902-06	1	Solid	20.04	20	20.8	9.06	05/12/2018	15:26
18	J2902-06DUP	1	Solid	20.03	20	20.8	9.07	05/12/2018	15:28
19	CCV3	1	Water			21.2	2.02	05/12/2018	15:30

2B 05320

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosivity J2901

WorkList ID : 112013

Date : 5/12/2018 12:37:11 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
05/11/2018	Solid	J2670-04	Corrosivity	Cool 4 deg C	PSEG05		TR-05-051118	05/11/2018	9045C
05/11/2018	Solid	J2674-04	Corrosivity	Cool 4 deg C	PSEG05		HD-01-051118	05/11/2018	9045C
05/11/2018	Solid	J2679-08	Corrosivity	Cool 4 deg C	PSEG05		EO-05-051118-A	05/11/2018	9045C
05/11/2018	Solid	J2679-10	Corrosivity	Cool 4 deg C	PSEG05		EO-05-051118-B	05/11/2018	9045C
05/11/2018	Solid	J2679-12	Corrosivity	Cool 4 deg C	PSEG05		EO-05-051118-C	05/11/2018	9045C
05/10/2018	Solid	J2901-09	Corrosivity	Cool 4 deg C	LIRO01	J51	251-COMP1-SL	05/10/2018	9045C
05/10/2018	Solid	J2901-10	Corrosivity	Cool 4 deg C	LIRO01	J51	251-COMP2-SL	05/10/2018	9045C
05/10/2018	Solid	J2901-11	Corrosivity	Cool 4 deg C	LIRO01	J51	260-COMP1-SL	05/10/2018	9045C
05/10/2018	Solid	J2901-12	Corrosivity	Cool 4 deg C	LIRO01	J51	260-COMP2-SL	05/10/2018	9045C
05/11/2018	Solid	J2902-05	Corrosivity	Cool 4 deg C	LIRO01	J41	295-COMP1-SL	05/11/2018	9045C
05/11/2018	Solid	J2902-06	Corrosivity	Cool 4 deg C	LIRO01	J41	228-COMP1-SL	05/11/2018	9045C

Date/Time 05/12/18 2:26 PM
 Received by: JD
 Relinquished by: CD

Date/Time 05/12/18 4:10 PM
 Received by: CD
 Relinquished by: JD