

7470A


LB95325

INSTRUMENT ID: CV1

Linear

 μ Abs.:

36128



A= 0.0000e+000

B= 2.7836e-004

C= -3.6007e-002

Rho= 0.9998946

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.029	-0.029	25	0.000	25				
0.2	0.200	0.193	-0.007	822	0.0 %	822				
2.5	2.500	2.603	0.103	9481	0.0 %	9481				
5.0	5.000	4.929	-0.071	17838	0.0 %	17838				
7.5	7.500	7.483	-0.017	27012	0.0 %	27012				
10.0	10.000	10.021	0.021	36128	0.0 %	36128				

LB95325 INSTRUMENT ID :CV1

Method: 7470A

Operator: Admin

Date of Analysis: 14 May 2018 09:10:21

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	25	-	0.0000	7470A	PPB	14 May 2018 09:14:17	S	Std
0.2 - 1	50.2	822	-	0.2000	7470A	PPB	14 May 2018 09:16:25	S	Std
2.5 - 1	52.5	9481	-	2.5000	7470A	PPB	14 May 2018 09:18:34	S	Std
5.0 - 1	55	17838	-	5.0000	7470A	PPB	14 May 2018 09:20:42	S	Std
7.5 - 1	57.5	27012	-	7.5000	7470A	PPB	14 May 2018 09:22:50	S	Std
10.0 - 1	510	36128	-	10.0000	7470A	PPB	14 May 2018 09:27:22	S	Std
ICV44 - 1	ICV44	14275	3.9376	-	7470A	PPB	14 May 2018 09:30:00	U	SMPL
ICB44 - 1	ICB44	-159	-0.0803	-	7470A	PPB	14 May 2018 09:32:06	U	SMPL
CCV14 - 1	CCV14	18446	5.0986	-	7470A	PPB	14 May 2018 09:34:12	U	SMPL
CCB14 - 1	CCB14	-158	-0.0800	-	7470A	PPB	14 May 2018 09:36:18	U	SMPL
CRA - 1	CRA	808	0.1889	-	7470A	PPB	14 May 2018 09:38:26	U	SMPL
HighStd - 1	HighStd	36318	10.0735	-	7470A	PPB	14 May 2018 09:40:32	U	SMPL
ChkStd - 1	ChkStd	26365	7.3030	-	7470A	PPB	14 May 2018 09:42:39	U	SMPL
PB109188BL - 1	PBW	-205	-0.0931	-	7470A	PPB	14 May 2018 09:44:48	U	SMPL
PB109188BS - 1	LCSW	0	-0.0360	-	7470A	PPB	14 May 2018 09:46:57	U	SMPL
PB109188BL - 1	PBW	-31	-0.0446	-	7470A	PPB	14 May 2018 09:47:12	U	SMPL
PB109188BS - 1	LCSW	14349	3.9582	-	7470A	PPB	14 May 2018 09:49:19	U	SMPL
PB109188TB - 1	PB109155TB	-52	-0.0505	-	7470A	PPB	14 May 2018 09:51:26	U	SMPL
J2675-04 - 1	HR-06-051018-A	5	-0.0346	-	7470A	PPB	14 May 2018 09:53:34	U	SMPL
J2675-04DUP - 1	HR-06-051018-ADUP	14	-0.0321	-	7470A	PPB	14 May 2018 09:55:43	U	SMPL
J2675-04MS - 1	HR-06-051018-AMS	14553	4.0150	-	7470A	PPB	14 May 2018 09:57:51	U	SMPL
J2675-04MSD - 1	HR-06-051018-AMSD	13961	3.8502	-	7470A	PPB	14 May 2018 09:59:58	U	SMPL
CCV15 - 1	CCV15	18244	5.0424	-	7470A	PPB	14 May 2018 10:02:05	U	SMPL
CCB15 - 1	CCB15	-164	-0.0817	-	7470A	PPB	14 May 2018 10:04:11	U	SMPL
J2675-06 - 1	HR-06-051018-B	-6	-0.0377	-	7470A	PPB	14 May 2018 10:06:18	U	SMPL
J2675-08 - 1	HR-06-051018-C	16	-0.0316	-	7470A	PPB	14 May 2018 10:08:25	U	SMPL
J2859-04 - 1	RB-44791RT-RB-47188RT	26	-0.0288	-	7470A	PPB	14 May 2018 10:10:31	U	SMPL
J2862-03 - 1	RB-27681RT	33	-0.0268	-	7470A	PPB	14 May 2018 10:12:38	U	SMPL
J2868-04 - 1	PS-RO-VNJ-227	37	-0.0257	-	7470A	PPB	14 May 2018 10:14:45	U	SMPL
J2868-08 - 1	PS-RO-327	24	-0.0293	-	7470A	PPB	14 May 2018 10:16:53	U	SMPL
J2868-12 - 1	PS-VNJ-238	101	-0.0079	-	7470A	PPB	14 May 2018 10:19:01	U	SMPL
J2870-04 - 1	CONCRETE-RUBBLE-PILE-A	19	-0.0307	-	7470A	PPB	14 May 2018 10:21:09	U	SMPL
J2870-08 - 1	CONCRETE-RUBBLE-PILE-B	21	-0.0302	-	7470A	PPB	14 May 2018 10:23:18	U	SMPL
J2870-13 - 1	SP-A	7	-0.0341	-	7470A	PPB	14 May 2018 10:25:25	U	SMPL
CCV16 - 1	CCV16	17929	4.9547	-	7470A	PPB	14 May 2018 10:27:32	U	SMPL
CCB16 - 1	CCB16	-168	-0.0828	-	7470A	PPB	14 May 2018 10:29:39	U	SMPL
J2870-17 - 1	SP-A2	-16	-0.0405	-	7470A	PPB	14 May 2018 10:31:46	U	SMPL
J2870-22 - 1	SP-B	16	-0.0316	-	7470A	PPB	14 May 2018 10:33:53	U	SMPL
J2872-03 - 1	237-COMP-SL	13	-0.0324	-	7470A	PPB	14 May 2018 10:36:01	U	SMPL
PB109189BL - 1	PBW	10	-0.0332	-	7470A	PPB	14 May 2018 10:38:08	U	SMPL
PB109189BS - 1	LCSW	14238	3.9273	-	7470A	PPB	14 May 2018 10:40:15	U	SMPL
PB109189TB - 1	PB109158TB	-73	-0.0563	-	7470A	PPB	14 May 2018 10:42:23	U	SMPL
J2861-01 - 1	S-1-COMP	-1	-0.0363	-	7470A	PPB	14 May 2018 10:44:31	U	SMPL
J2861-02 - 1	S-2-COMP	18	-0.0310	-	7470A	PPB	14 May 2018 10:46:39	U	SMPL
J2861-03 - 1	S-3-COMP	27	-0.0285	-	7470A	PPB	14 May 2018 10:48:48	U	SMPL
J2861-04 - 1	S-4-COMP	28	-0.0282	-	7470A	PPB	14 May 2018 10:50:55	U	SMPL
CCV17 - 1	CCV17	18253	5.0449	-	7470A	PPB	14 May 2018 10:53:02	U	SMPL
CCB17 - 1	CCB17	-187	-0.0881	-	7470A	PPB	14 May 2018 10:55:09	U	SMPL
J2861-04DUP - 1	S-4-COMP-DUP	27	-0.0285	-	7470A	PPB	14 May 2018 10:57:17	U	SMPL
J2861-04MS - 1	S-4-COMPMS	14444	3.9846	-	7470A	PPB	14 May 2018 10:59:24	U	SMPL
J2861-04MSD - 1	S-4-COMPMSD	14300	3.9446	-	7470A	PPB	14 May 2018 11:01:31	U	SMPL
PB109190BL - 1	PBW	-95	-0.0625	-	7470A	PPB	14 May 2018 11:03:38	U	SMPL
PB109190BS - 1	LCSW	15004	4.1405	-	7470A	PPB	14 May 2018 11:05:46	U	SMPL
J2856-01 - 1	MW-01	-110	-0.0666	-	7470A	PPB	14 May 2018 11:10:33	U	SMPL
J2856-02 - 1	MW-02	-134	-0.0733	-	7470A	PPB	14 May 2018 11:12:40	U	SMPL
J2856-03 - 1	MW-03	-108	-0.0661	-	7470A	PPB	14 May 2018 11:14:48	U	SMPL
J2856-03DUP - 1	MW-03DUP	-116	-0.0683	-	7470A	PPB	14 May 2018 11:16:57	U	SMPL
J2856-04 - 1	J2856-03MS	13865	3.8235	-	7470A	PPB	14 May 2018 11:19:04	U	SMPL
CCV18 - 1	CCV18	18346	5.0708	-	7470A	PPB	14 May 2018 11:21:11	U	SMPL
CCB18 - 1	CCB18	-180	-0.0861	-	7470A	PPB	14 May 2018 11:23:18	U	SMPL
J2856-05 - 1	J2856-03MSD	14013	3.8647	-	7470A	PPB	14 May 2018 11:25:26	U	SMPL
J2856-06 - 1	MW-DUPLICATE	-246	-0.1045	-	7470A	PPB	14 May 2018 11:27:33	U	SMPL
J2858-01 - 1	MW-8R	24	-0.0293	-	7470A	PPB	14 May 2018 11:29:40	U	SMPL
J2858-02 - 1	SW-3	-148	-0.0772	-	7470A	PPB	14 May 2018 11:31:47	U	SMPL
J2858-03 - 1	MW-9R	722	0.1650	-	7470A	PPB	14 May 2018 11:33:54	U	SMPL
J2858-04 - 1	MW-6R	120	-0.0026	-	7470A	PPB	14 May 2018 11:36:02	U	SMPL
J2860-01 - 1	MW-14	-17	-0.0407	-	7470A	PPB	14 May 2018 11:38:10	U	SMPL
J2860-02 - 1	SW-02	-168	-0.0828	-	7470A	PPB	14 May 2018 11:40:19	U	SMPL
J2860-03 - 1	MW-2R	-116	-0.0683	-	7470A	PPB	14 May 2018 11:42:27	U	SMPL
J2860-04 - 1	MW-12	-25	-0.0430	-	7470A	PPB	14 May 2018 11:44:35	U	SMPL
CCV19 - 1	CCV19	17992	4.9723	-	7470A	PPB	14 May 2018 11:46:42	U	SMPL
CCB19 - 1	CCB19	-166	-0.0822	-	7470A	PPB	14 May 2018 11:48:49	U	SMPL
J2860-05 - 1	SW-01	-133	-0.0730	-	7470A	PPB	14 May 2018 11:50:57	U	SMPL
J2876-17 - 1	TP-1	3168	0.8458	-	7470A	PPB	14 May 2018 11:53:04	U	SMPL
J2877-01 - 1	MS-WATER	-207	-0.0936	-	7470A	PPB	14 May 2018 11:55:12	U	SMPL
J2883-03 - 1	DRUM-1-5	782	0.1817	-	7470A	PPB	14 May 2018 11:57:19	U	SMPL
J2675-04LX5 - 1		-32	-0.0449	-	7470A	PPB	14 May 2018 11:59:27	U	SMPL
J2675-04A - 1		13691	3.7750	-	7470A	PPB	14 May 2018 12:01:35	U	SMPL
J2861-04LX5 - 1		-82	-0.0588	-	7470A	PPB	14 May 2018 12:03:42	U	SMPL
J2861-04A - 1		14342	3.9562	-	7470A	PPB	14 May 2018 12:05:50	U	SMPL
J2856-03LX5 - 1		-141	-0.0753	-	7470A	PPB	14 May 2018 12:07:59	U	SMPL

LB95325 INSTRUMENT ID :CV1

Method: 7470A

Operator: Admin

Date of Analysis: 14 May 2018 09:10:21

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
J2856-03A - 1		13410	3.6968		-7470A	PPB	14 May 2018 12:10:07	U	SMPL
CCV20 - 1	CCV20	17843	4.9308		-7470A	PPB	14 May 2018 12:12:15	U	SMPL
CCB20 - 1	CCB20	-186	-0.0878		-7470A	PPB	14 May 2018 12:14:22	U	SMPL