

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

5/14/2018 9:04

Test: Phenolics-mg/l

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	1.030	0.0	0.111	
ICB1	0.000	0.0	0.004	
CCV1	1.041	0.0	0.112	
CCB1	0.000	0.0	0.004	
PB109205BL	0.001	0.0	0.004	
PB109205BS	1.041	0.0	0.112	
J2858-01	0.000	0.0	0.004	
J2858-01DUP	0.001	0.0	0.004	
J2858-01MS	0.997	0.0	0.108	
J2858-01MSD	0.993	0.0	0.107	
J2858-02	0.000	0.0	0.004	
J2858-03	-0.000	0.0	0.004	
J2858-04	-0.002	0.0	0.003	
J2860-01	0.000	0.0	0.004	
CCV2	1.033	0.0	0.111	
CCB2	-0.001	0.0	0.003	
J2860-02	-0.004	0.0	0.003	
J2860-03	-0.003	0.0	0.003	
J2860-04	0.003	0.0	0.004	
J2860-05	0.002	0.0	0.004	
CCV3	1.033	0.0	0.111	
CCB3	0.001	0.0	0.004	

N 22
Mean 0.326
SD 0.4883
CV% 149.86

AS

Aquakem v. 7.2AQ1

Results from time period:

Mon May 14 08:18:14 2018

Mon May 14 08:53:39 2018

Sample Id	Sam/Ctr/c/	Test short nar	Test type	Result	Result unit	Result date and time	Stat
0.05PPM	A	Phenolics-	P	0.0459	mg/l	5/14/2018 8:18:14	
0.1PPM	A	Phenolics-	P	0.1016	mg/l	5/14/2018 8:18:15	
0.25PPM	A	Phenolics-	P	0.2503	mg/l	5/14/2018 8:18:16	
0.50PPM	A	Phenolics-	P	0.5056	mg/l	5/14/2018 8:18:17	
1.0PPM	A	Phenolics-	P	0.9961	mg/l	5/14/2018 8:18:18	
2.0PPM	A	Phenolics-	P	2.0005	mg/l	5/14/2018 8:18:19	
ICV1	S	Phenolics-	P	1.0297	mg/l	5/14/2018 8:43:34	
ICB1	S	Phenolics-	P	0.0005	mg/l	5/14/2018 8:43:35	
CCV1	S	Phenolics-	P	1.0406	mg/l	5/14/2018 8:43:36	
CCB1	S	Phenolics-	P	0.0005	mg/l	5/14/2018 8:43:37	
PB109205BL	S	Phenolics-	P	0.0013	mg/l	5/14/2018 8:43:38	
PB109205BS	S	Phenolics-	P	1.0415	mg/l	5/14/2018 8:43:39	
J2858-01	S	Phenolics-	P	0.0005	mg/l	5/14/2018 8:43:40	
J2858-01DUP	S	Phenolics-	P	0.0012	mg/l	5/14/2018 8:43:41	
J2858-01MS	S	Phenolics-	P	0.9974	mg/l	5/14/2018 8:43:42	
J2858-01MSD	S	Phenolics-	P	0.9931	mg/l	5/14/2018 8:43:43	
J2858-02	S	Phenolics-	P	0.0004	mg/l	5/14/2018 8:43:44	
J2858-03	S	Phenolics-	P	-0.0001	mg/l	5/14/2018 8:43:45	
J2858-04	S	Phenolics-	P	-0.0018	mg/l	5/14/2018 8:53:29	
J2860-01	S	Phenolics-	P	0	mg/l	5/14/2018 8:53:30	
CCV2	S	Phenolics-	P	1.0334	mg/l	5/14/2018 8:53:31	
CCB2	S	Phenolics-	P	-0.0011	mg/l	5/14/2018 8:53:32	
J2860-02	S	Phenolics-	P	-0.0042	mg/l	5/14/2018 8:53:33	
J2860-03	S	Phenolics-	P	-0.0026	mg/l	5/14/2018 8:53:34	
J2860-04	S	Phenolics-	P	0.0026	mg/l	5/14/2018 8:53:35	
J2860-05	S	Phenolics-	P	0.0018	mg/l	5/14/2018 8:53:36	
CCV3	S	Phenolics-	P	1.0329	mg/l	5/14/2018 8:53:37	
CCB3	S	Phenolics-	P	0.0012	mg/l	5/14/2018 8:53:38	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

5/14/2018 8:18


Test Phenolics-

Accepted 5/14/2018 8:18

Factor 9.594
Bias 0.004

Coeff. of det. 0.999976

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	Phenol-2	0.008	0.0459	0.0500	
2	Phenol-2	0.014	0.1016	0.1000	
3	Phenol-2	0.030	0.2503	0.2500	
4	Phenol-2	0.056	0.5056	0.5000	
5	Phenol-2	0.107	0.9961	1.0000	
6	Phenol-2	0.212	2.0005	2.0000	