

21395329

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : As

5/14/2018 9:46

Test: Phenolics-mg/d

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	1.035	0.0	0.111	
ICB1	-0.003	0.0	0.003	
CCV1	1.027	0.0	0.111	
CCB1	-0.002	0.0	0.003	
PB109204BL	0.003	0.0	0.004	
PB109204BS	1.029	0.0	0.111	
J2856-01	-0.001	0.0	0.003	
J2856-02	-0.001	0.0	0.003	
J2856-03	-0.002	0.0	0.003	
J2856-03DUP	-0.005	0.0	0.003	
J2856-04	1.017	0.0	0.110	
J2856-05	1.023	0.0	0.110	
J2856-06	-0.002	0.0	0.003	
CCV2	1.029	0.0	0.111	
CCB2	-0.005	0.0	0.003	

N 15
Mean 0.409
SD 0.5216
CV% 127.39

AS

Aquakem v. 7.2AQ1

Results from time period:

Mon May 14 09:35:08 2018

Mon May 14 09:39:36 2018

Sample Id	Sam/Ctr/c/	Test short nan	Test type	Result	Result unit	Result date and time	Stat
0.05PPM	A	Phenolics-	P	0.0459	mg/l	5/14/2018 8:18:14	
0.1PPM	A	Phenolics-	P	0.1016	mg/l	5/14/2018 8:18:15	
0.25PPM	A	Phenolics-	P	0.2503	mg/l	5/14/2018 8:18:16	
0.50PPM	A	Phenolics-	P	0.5056	mg/l	5/14/2018 8:18:17	
1.0PPM	A	Phenolics-	P	0.9961	mg/l	5/14/2018 8:18:18	
2.0PPM	A	Phenolics-	P	2.0005	mg/l	5/14/2018 8:18:19	
ICV1	S	Phenolics-	P	1.0352	mg/l	5/14/2018 9:35:08	
ICB1	S	Phenolics-	P	-0.0027	mg/l	5/14/2018 9:35:09	
CCV1	S	Phenolics-	P	1.0271	mg/l	5/14/2018 9:35:10	
CCB1	S	Phenolics-	P	-0.0021	mg/l	5/14/2018 9:35:11	
PB109204BL	S	Phenolics-	P	0.0031	mg/l	5/14/2018 9:35:12	
PB109204BS	S	Phenolics-	P	1.0289	mg/l	5/14/2018 9:35:13	
J2856-01	S	Phenolics-	P	-0.0014	mg/l	5/14/2018 9:35:14	
J2856-02	S	Phenolics-	P	-0.0009	mg/l	5/14/2018 9:35:15	
J2856-03	S	Phenolics-	P	-0.0023	mg/l	5/14/2018 9:35:16	
J2856-03DUP	S	Phenolics-	P	-0.0047	mg/l	5/14/2018 9:35:17	
J2856-04	S	Phenolics-	P	1.0167	mg/l	5/14/2018 9:35:18	
J2856-05	S	Phenolics-	P	1.0226	mg/l	5/14/2018 9:35:19	
J2856-06	S	Phenolics-	P	-0.0017	mg/l	5/14/2018 9:39:34	
CCV2	S	Phenolics-	P	1.029	mg/l	5/14/2018 9:39:35	
CCB2	S	Phenolics-	P	-0.0051	mg/l	5/14/2018 9:39:36	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

5/14/2018 8:19


Test Phenolics-

Accepted 5/14/2018 8:18

Factor 9.594
Bias 0.004

Coeff. of det. 0.999976

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	Phenol-2	0.008	0.0459	0.0500	
2	Phenol-2	0.014	0.1016	0.1000	
3	Phenol-2	0.030	0.2503	0.2500	
4	Phenol-2	0.056	0.5056	0.5000	
5	Phenol-2	0.107	0.9961	1.0000	
6	Phenol-2	0.212	2.0005	2.0000	