



Extraction and Analytical Summary Report

Analysis Method: 1664A
Test: TPH
Run Number: LB95333
Analysis Date: 05/14/2018
BalanceID: WC SC-3
OvenID: EX Oven-2

ANALYST: JIGNESH
REVIEWED BY: apatel
Extraction Date: 05/14/2018
Extraction IN Time: 08:00
Extraction OUT Time: 08:40

Dish #	Lab ID	Client ID	Matrix	pH	Sample Vol (ml)	Final Volume (ml)	Empty Dish Weight	Final Empty Dish Weight (g)	Silica Gel Weight (g)	Weight After Drying (g)	Final Weight After Drying (g)	Change Weight (g)	Result in ppm
1	LB95333BL	LB95333BL	WATER	2	1000	100	3.0625	3.0625	3.02	3.0625	3.0625	0.0000	0
2	LB95333BS	LB95333BS	WATER	2	1000	100	3.1857	3.1857	3.01	3.2045	3.2045	0.0188	18.8
3	LB95333BSD	LB95333BSD	WATER	2	1000	100	3.0144	3.0144	3.03	3.0335	3.0335	0.0191	19.1
4	J2877-01	MS-WATER	WATER	2	1000	100	3.1632	3.1632	3.04	3.2538	3.2538	0.0906	90.6
5	J2883-03	DRUM-1-5	WATER	2	1000	100	3.0953	3.0953	3.05	3.0969	3.0969	0.0016	1.6



QC Batch# LB95333 Test: TPH Analysis Date: 05/14/2018

Chemicals Used:

Chemical Name	Chemical Lot #
HEXANE	W2362
pH Paper 0-14	W2329
Sodium Sulfate	EP1843
1:1 HCL	WP64228
Silica Gel	W2199
Sand	NA

Standards Used:

Standard Name	Amount Used	Standard Lot #
LCSW	10.00 ML	WP64229
LCSWD	10.00 ML	WP64230
MS/MSD	NA	NA

BALANCE CALIBRATION / OVEN Dessicator Data

Analytical Balance ID # : WC SC-3

Before Analysis

0.0020 gram Balance:	<u>0.0018</u>	In OVEN TEMP :	<u>71</u>	Dessicator Time In :	<u>10:20</u>
1.0000 gram Balance:	<u>1.0005</u>	In Time:	<u>09:15</u>		
Bal Check Time:	<u>08:22</u>	Out OVEN TEMP:	<u>71</u>	Dessicator Time Out:	<u>10:50</u>
		Out Time:	<u>10:19</u>		

After Analysis

0.0020 gram Balance:	<u>0.0021</u>	In OVEN TEMP :	<u>70</u>	Dessicator Time In :	<u>11:51</u>
1.0000 gram Balance:	<u>1.0003</u>	In Time:	<u>11:20</u>		
Bal Check Time:	<u>12:50</u>	Out OVEN TEMP:	<u>70</u>	Dessicator Time Out:	<u>12:25</u>
		Out Time:	<u>11:50</u>		

WorkList Name : TPH J2883

WorkList ID : 112022

Date : 5/14/2018 7:42:36 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
05/16/2018	Water	J2877-01 - 12	TPH	Conc H2SO4 1c	PSEG01	I43	MS-WATER	05/11/2018	1664A
05/16/2018	Water	J2883-03 - 12	TPH	Conc H2SO4 1c	PSEG01	I51	DRUM-1-5	05/11/2018	1664A

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TPH J2883

WorkList ID : 112022

Date : 5/14/2018 7:42:36 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
05/16/2018	Water	J2877-01 - 12	TPH	Conc H2SO4 1c	PSEG01	I43	MS-WATER	05/11/2018	1664A
05/16/2018	Water	J2883-03 - 12	TPH	Conc H2SO4 1c	PSEG01	I51	DRUM-1-5	05/11/2018	1664A

Date/Time 05-14-18 7:50 AM
Received by: JD
Relinquished by: CB

Date/Time 05-14-18 9:45 AM
Received by: CB
Relinquished by: JD