

7471B


LB95335

INSTRUMENT ID: CV2

Linear

μ Abs.:

45332



A= 0.0000e+000
B= 2.2004e-004
C= -9.2472e-003
Rho= 0.9997366
Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.000	0.000	43	0.000	43				
0.2	0.200	0.197	-0.003	936	0.0 %	936				
2.5	2.500	2.408	-0.092	10984	0.0 %	10984				
5.0	5.000	5.176	0.176	23567	0.0 %	23567				
7.5	7.500	7.453	-0.047	33914	0.0 %	33914				
10.0	10.000	9.966	-0.034	45332	0.0 %	45332				

LB95335 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 14 May 2018 13:57:33

Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	43	-	0.0000	7471B	PPB	14 May 2018 14:00:50	S	Std
0.2 - 1	50.2	936	-	0.2000	7471B	PPB	14 May 2018 14:03:08	S	Std
2.5 - 1	52.5	10984	-	2.5000	7471B	PPB	14 May 2018 14:05:25	S	Std
5.0 - 1	55	23567	-	5.0000	7471B	PPB	14 May 2018 14:07:43	S	Std
7.5 - 1	57.5	33914	-	7.5000	7471B	PPB	14 May 2018 14:10:00	S	Std
10.0 - 1	510	45332	-	10.0000	7471B	PPB	14 May 2018 14:12:17	S	Std
ICV45 - 1	ICV45	17494	3.8402	-	7471B	PPB	14 May 2018 14:15:01	U	SMPL
ICB45 - 1	ICB45	-142	-0.0405	-	7471B	PPB	14 May 2018 14:17:18	U	SMPL
CCV21 - 1	CCV21	21585	4.7403	-	7471B	PPB	14 May 2018 14:24:56	U	SMPL
CCB21 - 1	CCB21	-165	-0.0456	-	7471B	PPB	14 May 2018 14:27:11	U	SMPL
CRA - 1	CRA	871	0.1824	-	7471B	PPB	14 May 2018 14:29:28	U	SMPL
HighStd - 1	HighStd	45765	10.0610	-	7471B	PPB	14 May 2018 14:31:43	U	SMPL
ChkStd - 1	ChkStd	32357	7.1106	-	7471B	PPB	14 May 2018 14:34:00	U	SMPL
PB109210BL - 1	PBS	-210	-0.0555	-	7471B	PPB	14 May 2018 14:36:18	U	SMPL
PB109210BS - 1	LCSS	15961	3.5028	-	7471B	PPB	14 May 2018 14:38:36	U	SMPL
J2671-11 - 1	AU-02-051118-A	3928	0.8551	-	7471B	PPB	14 May 2018 14:40:53	U	SMPL
J2671-13 - 1	AU-02-051118-B	3341	0.7259	-	7471B	PPB	14 May 2018 14:43:11	U	SMPL
J2878-01 - 1	P003-S006-0001-01	3182	0.6909	-	7471B	PPB	14 May 2018 14:45:28	U	SMPL
J2878-02 - 1	P003-S007-0001-01	5090	1.1108	-	7471B	PPB	14 May 2018 14:47:46	U	SMPL
J2878-03 - 1	P003-S008-0001-01	6222	1.3599	-	7471B	PPB	14 May 2018 14:50:02	U	SMPL
CCV22 - 1	CCV22	21751	4.7769	-	7471B	PPB	14 May 2018 14:52:18	U	SMPL
CCB22 - 1	CCB22	-117	-0.0350	-	7471B	PPB	14 May 2018 14:54:34	U	SMPL
J2880-02 - 1	PLANTER-AREA-2-PRECHAR	6705	1.4661	-	7471B	PPB	14 May 2018 14:56:51	U	SMPL
J2880-05 - 1	PLANTER-AREA-3-PRECHAR	5651	1.2342	-	7471B	PPB	14 May 2018 14:59:06	U	SMPL
J2882-01 - 1	P003-S005-0001-01	2591	0.5609	-	7471B	PPB	14 May 2018 15:01:22	U	SMPL
J2882-01DUP - 1	P003-S005-0001-01DUP	2526	0.5466	-	7471B	PPB	14 May 2018 15:03:38	U	SMPL
J2882-02 - 1	J2882-01MS	19246	4.2257	-	7471B	PPB	14 May 2018 15:05:55	U	SMPL
J2882-03 - 1	J2882-01MSD	19204	4.2164	-	7471B	PPB	14 May 2018 15:08:11	U	SMPL
J2882-04 - 1	P003-S005-0001-02	4648	1.0135	-	7471B	PPB	14 May 2018 15:10:29	U	SMPL
J2888-01 - 1	RRT-N1	857	0.1793	-	7471B	PPB	14 May 2018 15:12:46	U	SMPL
J2888-05 - 1	RRT-N2	2296	0.4960	-	7471B	PPB	14 May 2018 15:15:04	U	SMPL
J2888-09 - 1	RRT-S1	1117	0.2365	-	7471B	PPB	14 May 2018 15:17:20	U	SMPL
CCV23 - 1	CCV23	22255	4.8878	-	7471B	PPB	14 May 2018 15:19:36	U	SMPL
CCB23 - 1	CCB23	-131	-0.0381	-	7471B	PPB	14 May 2018 15:21:53	U	SMPL
J2888-13 - 1	RRT-S2	285	0.0535	-	7471B	PPB	14 May 2018 15:24:10	U	SMPL
J2888-17 - 1	26KV-N1	895	0.1877	-	7471B	PPB	14 May 2018 15:26:26	U	SMPL
J2888-21 - 1	26KV-N2	668	0.1377	-	7471B	PPB	14 May 2018 15:28:42	U	SMPL
J2888-25 - 1	26KV-S1	235	0.0425	-	7471B	PPB	14 May 2018 15:30:58	U	SMPL
J2888-29 - 1	26KV-S2	228	0.0409	-	7471B	PPB	14 May 2018 15:33:15	U	SMPL
J2899-02 - 1	TP-7A	2675	0.5794	-	7471B	PPB	14 May 2018 15:35:32	U	SMPL
J2899-06 - 1	TP-8A	2501	0.5411	-	7471B	PPB	14 May 2018 15:37:49	U	SMPL
J2899-09 - 1	TP-8B	1714	0.3679	-	7471B	PPB	14 May 2018 15:40:07	U	SMPL
PB109211BL - 1	PBS	22	-0.0044	-	7471B	PPB	14 May 2018 15:42:24	U	SMPL
PB109211BS - 1	LCSS	16169	3.5486	-	7471B	PPB	14 May 2018 15:44:40	U	SMPL
CCV24 - 1	CCV24	22199	4.8755	-	7471B	PPB	14 May 2018 15:46:56	U	SMPL
CCB24 - 1	CCB24	-126	-0.0370	-	7471B	PPB	14 May 2018 15:49:13	U	SMPL
J2883-01 - 1	DRUM-1-5	221	0.0394	-	7471B	PPB	14 May 2018 15:51:29	U	SMPL
J2898-01 - 1	SP-01	15243	3.3448	-	7471B	PPB	14 May 2018 15:53:46	U	SMPL
J2898-01DUP - 1	SP-01DUP	14791	3.2454	-	7471B	PPB	14 May 2018 15:58:38	U	SMPL
J2898-01MS - 1	SP-01MS	33828	7.4343	-	7471B	PPB	14 May 2018 16:06:19	U	SMPL
J2898-01MSD - 1	SP-01MSD	33728	7.4123	-	7471B	PPB	14 May 2018 16:08:35	U	SMPL
J2898-02 - 1	SP-02	32968	7.2451	-	7471B	PPB	14 May 2018 16:10:53	U	SMPL
J2898-03 - 1	SP-03	20362	4.4712	-	7471B	PPB	14 May 2018 16:13:12	U	SMPL
J2898-04 - 1	SP-04	41305	9.0796	-	7471B	PPB	14 May 2018 16:15:31	U	SMPL
J2898-05 - 1	SP-05	95942	21.1020	-	7471B	PPB	14 May 2018 16:17:49	U	SMPL
J2898-06 - 1	SP-06	29376	6.4547	-	7471B	PPB	14 May 2018 16:20:07	U	SMPL
CCV25 - 1	CCV25	21940	4.8185	-	7471B	PPB	14 May 2018 16:22:26	U	SMPL
CCB25 - 1	CCB25	-183	-0.0495	-	7471B	PPB	14 May 2018 16:24:44	U	SMPL
J2898-07 - 1	SP-07	54707	12.0286	-	7471B	PPB	14 May 2018 16:27:02	U	SMPL
J2900-01 - 1	EPE-MX-5A(115-117)	162	0.0264	-	7471B	PPB	14 May 2018 16:29:18	U	SMPL
J2900-02 - 1	EPE-MX-5A(125-127)	812	0.1694	-	7471B	PPB	14 May 2018 16:31:37	U	SMPL
J2900-03 - 1	EPE-F-7A(65-67)	115	0.0161	-	7471B	PPB	14 May 2018 16:33:53	U	SMPL
J2900-04 - 1	EPE-F-7A(70-72)	657	0.1353	-	7471B	PPB	14 May 2018 16:36:09	U	SMPL
J2900-05 - 1	EPE-F-7A(80-82)	499	0.1006	-	7471B	PPB	14 May 2018 16:38:26	U	SMPL
J2900-06 - 1	EPE-F-5A(0-5)	1056	0.2231	-	7471B	PPB	14 May 2018 16:40:43	U	SMPL
J2903-02 - 1	S-1-COMP	27113	5.9567	-	7471B	PPB	14 May 2018 16:43:01	U	SMPL
J2903-04 - 1	S-2-COMP	28388	6.2373	-	7471B	PPB	14 May 2018 16:45:18	U	SMPL
J2903-06 - 1	S-3-COMP	58984	12.9697	-	7471B	PPB	14 May 2018 16:47:35	U	SMPL
CCV26 - 1	CCV26	21984	4.8281	-	7471B	PPB	14 May 2018 16:49:53	U	SMPL
CCB26 - 1	CCB26	-175	-0.0478	-	7471B	PPB	14 May 2018 16:52:12	U	SMPL
J2903-08 - 1	S-4-COMP	20031	4.3984	-	7471B	PPB	14 May 2018 16:54:30	U	SMPL
J2903-10 - 1	S-5-COMP	21155	4.6457	-	7471B	PPB	14 May 2018 16:56:46	U	SMPL
J2903-12 - 1	S-6-COMP	31886	7.0070	-	7471B	PPB	14 May 2018 16:59:03	U	SMPL
J2882-01LX5 - 1		329	0.0631	-	7471B	PPB	14 May 2018 17:01:19	U	SMPL
J2882-01A - 1		19829	4.3540	-	7471B	PPB	14 May 2018 17:03:37	U	SMPL
J2898-01LX5 - 1		2806	0.6082	-	7471B	PPB	14 May 2018 17:05:54	U	SMPL
J2898-01A - 1		33331	7.3250	-	7471B	PPB	14 May 2018 17:08:12	U	SMPL
J2898-05DLX5 - 1		18522	4.0664	-	7471B	PPB	14 May 2018 17:10:29	U	SMPL
J2898-07DLX2 - 1		26390	5.7976	-	7471B	PPB	14 May 2018 17:12:48	U	SMPL
J2903-06DLX2 - 1		28513	6.2648	-	7471B	PPB	14 May 2018 17:15:05	U	SMPL
CCV27 - 1	CCV27	21807	4.7892	-	7471B	PPB	14 May 2018 17:17:23	U	SMPL

LB95335 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 14 May 2018 13:57:33

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
CCB27 - 1	CCB27	-128	-0.0374		7471B	PPB	14 May 2018 17:19:41	U	SMPL