

Analytical Summary ReportAnalysis Method: 365.3ANALYST: AMANDEEPParameter: Phosphorus, TotalSUPERVISOR REVIEW BY: apatelRun Number: LB95336BALANCE ID: WC SC-4

Reagent/Standard	Lot/Log #
5N sulfuric acid	WP61786
calibration std. phosphate 1 ppm	WP64462
calibration std. phosphate 0.5 ppm	WP64461
calibration std. phosphate 0.3 ppm	WP64460
calibration std. 0 ppm	WP64457
phosphate ICV std.	WP64466
phosphate CCV std.	WP64465
calibration std. phosphate 0.1 ppm	WP64459
calibration std. phosphate 0.05 ppm	WP64458
Combined reagent	WP64464
11 N sulfuric acid	WP61601
Phenolphthalein indicator	WP61777
Sodium hydroxide, 1N	WP61776

Intercept: -0.0012Slope: 0.6509Regression: 0.999861

Seq	Lab ID	True Value (mg/L)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	AnalDate	AnalTime
1	CAL1	0.00	1	50	50	0.000	05/14/2018	11:31
2	CAL2	0.05	1	50	50	0.033	05/14/2018	11:32
3	CAL3	0.10	1	50	50	0.065	05/14/2018	11:32
4	CAL4	0.30	1	50	50	0.186	05/14/2018	11:33
5	CAL5	0.50	1	50	50	0.328	05/14/2018	11:33
6	CAL6	1.00	1	50	50	0.650	05/14/2018	11:34



Analytical Summary Report

Analysis Method: 365.3

ANALYST: AMANDEEP

Parameter: Phosphorus, Total

SUPERVISOR REVIEW BY: apatel

Run Number: LB95336

BALANCE ID: WC SC-4

Seq	Lab ID	True Value (mg/l)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	Result (mg/L)	AnalDate	AnalTime
1	ICV	0.50	1	50	50	0.320	0.493	05/14/2018	11:34
2	ICB		1	50	50	0.000	0.002	05/14/2018	11:34
3	CCV1	0.50	1	50	50	0.325	0.501	05/14/2018	11:35
4	CCB1		1	50	50	0.000	0.002	05/14/2018	11:35
5	PB109268BL		1	1.00	50	0.000	0.002	05/14/2018	11:36
6	PB109268BS	0.50	1	1.01	50	0.313	0.483	05/14/2018	11:36
7	J2823-01		1	1.04	50	1.368	2.104	05/14/2018	11:37
8	J2823-02		1	1.00	50	1.333	2.050	05/14/2018	11:37
9	J2823-02DUP		1	1.03	50	1.320	2.030	05/14/2018	11:38
10	J2823-02MS	0.50	1	1.01	50	1.440	2.214	05/14/2018	11:39
11	J2823-02MSD	0.50	1	1.01	50	1.435	2.206	05/14/2018	11:39
12	J2823-01		50	1.04	50	0.358	0.552	05/14/2018	12:01
13	J2823-02		50	1.00	50	0.293	0.452	05/14/2018	12:01
14	J2823-02DUP		50	1.03	50	0.272	0.420	05/14/2018	12:02
15	CCV2	0.50	1	50	50	0.328	0.506	05/14/2018	12:02
16	CCB2		1	50	50	0.000	0.002	05/14/2018	12:03
17	J2823-02MS	0.50	50	1.01	50	0.322	0.497	05/14/2018	12:04
18	J2823-02MSD	0.50	50	1.01	50	0.324	0.500	05/14/2018	12:04
19	CCV3	0.50	1	50	50	0.327	0.504	05/14/2018	12:05
20	CCB3		1	50	50	0.000	0.002	05/14/2018	12:05