

=====

Test results

Aquakem 7.2AQ1

21395341
Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : JS

5/14/2018 16:16

Test: Total CN Mg/L

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	95.163	0.0	0.095	
ICB1	0.825	0.0	0.004	
CCV1	244.158	0.0	0.239	
CCB1	1.124	0.0	0.004	
PB109241BL	0.932	0.0	0.004	
PB109241BS	97.354	0.0	0.097	
MIDPB109241	244.829	0.0	0.239	
J2904-01	1.705	0.0	0.004	
J2904-01DUP	1.419	0.0	0.004	
J2904-01MS	26.036	0.0	0.028	
J2904-01MSD	25.493	0.0	0.027	
CCV2	256.696	0.0	0.251	
CCB2	-0.916	0.0	0.002	

N	13
Mean	76.525
SD	103.7584
CV%	135.59

Aquakem v. 7.2AQ1

Results from time period:

Mon May 14 16:03:35 2018

Mon May 14 16:06:59 2018

AS

Sample Id	Sam/Ctr/c/	Test short	Test type	Result	Result unit	Result date and time	Stat
0.0PPBCN	A	Total CN	P	0.8381	µg/l	5/14/2018 11:08:17	
5.0PPBCN	A	Total CN	P	4.9513	µg/l	5/14/2018 11:08:18	
10PPBCN	A	Total CN	P	10.3396	µg/l	5/14/2018 11:08:19	
50PPBCN	A	Total CN	P	47.6196	µg/l	5/14/2018 11:08:20	
100PPBCN	A	Total CN	P	99.3178	µg/l	5/14/2018 11:08:21	
250PPBCN	A	Total CN	P	253.1309	µg/l	5/14/2018 11:08:22	
500PPBCN	A	Total CN	P	498.8027	µg/l	5/14/2018 11:08:23	
ICV1	S	Total CN	P	95.1632	µg/l	5/14/2018 16:03:35	
ICB1	S	Total CN	P	0.8252	µg/l	5/14/2018 16:03:36	
CCV1	S	Total CN	P	244.158	µg/l	5/14/2018 16:03:37	
CCB1	S	Total CN	P	1.1241	µg/l	5/14/2018 16:03:38	
PB109241BL	S	Total CN	P	0.9318	µg/l	5/14/2018 16:03:39	
PB109241BS	S	Total CN	P	97.354	µg/l	5/14/2018 16:03:40	
MIDPB109241	S	Total CN	P	244.8293	µg/l	5/14/2018 16:03:41	
J2904-01	S	Total CN	P	1.7047	µg/l	5/14/2018 16:03:42	
J2904-01DUP	S	Total CN	P	1.4195	µg/l	5/14/2018 16:03:43	
J2904-01MS	S	Total CN	P	26.0356	µg/l	5/14/2018 16:03:44	
J2904-01MSD	S	Total CN	P	25.4934	µg/l	5/14/2018 16:03:45	
CCV2	S	Total CN	P	256.6961	µg/l	5/14/2018 16:06:58	
CCB2	S	Total CN	P	-0.9163	µg/l	5/14/2018 16:06:59	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

5/14/2018 11:19


Test Total CN

Accepted 5/14/2018 11:18

Factor 1035
Bias 0.003

Coeff. of det. 0.999911

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.004	0.8381	0.0000	
2	5.0PPBCN	0.008	4.9513	5.0000	
3	10PPBCN	0.013	10.3396	10.0000	
4	50PPBCN	0.049	47.6196	50.0000	
5	100PPBCN	0.099	99.3178	100.0000	
6	250PPBCN	0.247	253.1309	250.0000	
7	500PPBCN	0.485	498.8027	500.0000	