



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB95342

Slope : 98.2

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.1	4.01	05/14/2018	16:15
2	CAL2	1	Water			21.1	7.00	05/14/2018	16:16
3	CAL3	1	Water			21.1	10.02	05/14/2018	16:18
4	ICV	1	Water			21.2	7.01	05/14/2018	16:19
5	CCV1	1	Water			21.2	2.01	05/14/2018	16:20
6	J1759-02	1	Solid	20.02	20	20.8	8.34	05/14/2018	16:21
7	J2267-02	1	Solid	20.03	20	21.3	5.60	05/14/2018	16:22
8	J2267-02DUP	1	Solid	20.05	20	21.4	5.61	05/14/2018	16:24
9	CCV2	1	Water			21.1	12.01	05/14/2018	16:25

1805342

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosivity j2267

WorkList ID : 112072

Date : 5/14/2018 3:51:02 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
05/14/2018	Solid	J1759-02	Corrosivity	Cool 4 deg C	PSEG01	I62	15TH-ST-SP-A	05/14/2018	9045C
05/14/2018	Solid	J2267-02	Corrosivity	Cool 4 deg C	PSEG01	I63	KG-GRAVEL	05/14/2018	9045C

Date/Time 05-14-18 4:10 PM
 Received by: JP
 Relinquished by: CD

Date/Time 05-14-18 5:00 PM
 Received by: LP
 Relinquished by: JH