


7470A

LB 95352
INSTRUMENT ID: CV1

Linear

 μ Abs.:

36404



A= 0.0000e+000

B= 2.7502e-004

C= 4.5620e-003

Rho= 0.9997792

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.028	0.028	86	0.000	86				
0.2	0.200	0.216	0.016	768	0.0 %	768				
2.5	2.500	2.386	-0.114	8660	0.0 %	8660				
5.0	5.000	5.128	0.128	18628	0.0 %	18628				
7.5	7.500	7.426	-0.074	26984	0.0 %	26984				
10.0	10.000	10.016	0.016	36404	0.0 %	36404				

LB95352 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 15 May 2018 08:47:52

Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	86	-	0.0000	7470A	PPB	15 May 2018 08:53:01	S	Std
0.2 - 1	50.2	768	-	0.2000	7470A	PPB	15 May 2018 08:55:09	S	Std
2.5 - 1	52.5	8660	-	2.5000	7470A	PPB	15 May 2018 08:57:18	S	Std
5.0 - 1	55	18628	-	5.0000	7470A	PPB	15 May 2018 08:59:26	S	Std
7.5 - 1	57.5	26984	-	7.5000	7470A	PPB	15 May 2018 09:01:34	S	Std
10.0 - 1	510	36404	-	10.0000	7470A	PPB	15 May 2018 09:06:10	S	Std
ICV46 - 1	ICV46	13745	3.7847	-	7470A	PPB	15 May 2018 09:08:42	U	SMPL
ICB46 - 1	ICB46	-138	-0.0334	-	7470A	PPB	15 May 2018 09:10:49	U	SMPL
CCV28 - 1	CCV28	18054	4.9698	-	7470A	PPB	15 May 2018 09:12:57	U	SMPL
CCB28 - 1	CCB28	-90	-0.0202	-	7470A	PPB	15 May 2018 09:15:03	U	SMPL
CRA - 1	CRA	789	0.2216	-	7470A	PPB	15 May 2018 09:17:11	U	SMPL
HighStd - 1	HighStd	36566	10.0610	-	7470A	PPB	15 May 2018 09:19:17	U	SMPL
ChkStd - 1	ChkStd	24545	6.7549	-	7470A	PPB	15 May 2018 09:21:24	U	SMPL
PB109245BL - 1	PBW	-267	-0.0689	-	7470A	PPB	15 May 2018 09:23:33	U	SMPL
PB109245BS - 1	LCSW	12998	3.5793	-	7470A	PPB	15 May 2018 09:25:43	U	SMPL
PB109245TB - 1	PB109199TB	-173	-0.0430	-	7470A	PPB	15 May 2018 09:27:50	U	SMPL
J2671-12 - 1	AU-02-051118-A	-92	-0.0207	-	7470A	PPB	15 May 2018 09:29:59	U	SMPL
J2671-12DUP - 1	AU-02-051118-ADUP	-74	-0.0158	-	7470A	PPB	15 May 2018 09:32:07	U	SMPL
J2671-12MS - 1	AU-02-051118-AMS	12762	3.5144	-	7470A	PPB	15 May 2018 09:34:16	U	SMPL
J2671-12MSD - 1	AU-02-051118-AMSD	13102	3.6079	-	7470A	PPB	15 May 2018 09:36:22	U	SMPL
CCV29 - 1	CCV29	18030	4.9632	-	7470A	PPB	15 May 2018 09:38:29	U	SMPL
CCB29 - 1	CCB29	-153	-0.0375	-	7470A	PPB	15 May 2018 09:40:36	U	SMPL
J2671-14 - 1	AU-02-051118-B	-78	-0.0169	-	7470A	PPB	15 May 2018 09:42:44	U	SMPL
J2876-03 - 1	TP-1B	-37	-0.0056	-	7470A	PPB	15 May 2018 09:44:50	U	SMPL
J2876-06 - 1	TP-10	-25	-0.0023	-	7470A	PPB	15 May 2018 09:46:56	U	SMPL
J2876-09 - 1	TP-11A	3	0.0054	-	7470A	PPB	15 May 2018 09:49:04	U	SMPL
J2876-12 - 1	TP-11B	-87	-0.0194	-	7470A	PPB	15 May 2018 09:51:11	U	SMPL
J2880-03 - 1	PLANTER-AREA-2-PRECHAR	-39	-0.0062	-	7470A	PPB	15 May 2018 09:53:19	U	SMPL
J2880-06 - 1	PLANTER-AREA-3-PRECHAR	-56	-0.0108	-	7470A	PPB	15 May 2018 09:55:27	U	SMPL
J2883-02 - 1	DRUM-1-5	-15	0.0004	-	7470A	PPB	15 May 2018 09:57:35	U	SMPL
J2888-04 - 1	RRT-N1	-34	-0.0048	-	7470A	PPB	15 May 2018 09:59:43	U	SMPL
J2888-08 - 1	RRT-N2	-12	0.0013	-	7470A	PPB	15 May 2018 10:01:50	U	SMPL
CCV30 - 1	CCV30	18113	4.9860	-	7470A	PPB	15 May 2018 10:03:57	U	SMPL
CCB30 - 1	CCB30	-144	-0.0350	-	7470A	PPB	15 May 2018 10:06:04	U	SMPL
J2888-12 - 1	RRT-S1	-19	-0.0007	-	7470A	PPB	15 May 2018 10:08:13	U	SMPL
J2888-16 - 1	RRT-S2	-45	-0.0078	-	7470A	PPB	15 May 2018 10:10:20	U	SMPL
J2888-20 - 1	26KV-N1	-10	0.0018	-	7470A	PPB	15 May 2018 10:12:26	U	SMPL
J2888-24 - 1	26KV-N2	5	0.0059	-	7470A	PPB	15 May 2018 10:14:33	U	SMPL
J2888-28 - 1	26KV-S1	-14	0.0007	-	7470A	PPB	15 May 2018 10:16:40	U	SMPL
J2888-32 - 1	26KV-S2	-63	-0.0128	-	7470A	PPB	15 May 2018 10:18:48	U	SMPL
J2899-03 - 1	TP-7A	-27	-0.0029	-	7470A	PPB	15 May 2018 10:20:56	U	SMPL
J2899-07 - 1	TP-8A	-62	-0.0125	-	7470A	PPB	15 May 2018 10:23:04	U	SMPL
J2899-10 - 1	TP-8B	-49	-0.0089	-	7470A	PPB	15 May 2018 10:25:13	U	SMPL
PB109247BL - 1	PBW	-28	-0.0031	-	7470A	PPB	15 May 2018 10:27:20	U	SMPL
CCV31 - 1	CCV31	17913	4.9310	-	7470A	PPB	15 May 2018 10:29:27	U	SMPL
CCB31 - 1	CCB31	-40	-0.0064	-	7470A	PPB	15 May 2018 10:31:34	U	SMPL
PB109247BS - 1	LCSW	13910	3.8301	-	7470A	PPB	15 May 2018 10:33:42	U	SMPL
PB109247TB - 1	PB109202TB	-115	-0.0271	-	7470A	PPB	15 May 2018 10:35:49	U	SMPL
J2876-18 - 1	TP-1	1	0.0048	-	7470A	PPB	15 May 2018 10:37:57	U	SMPL
J2877-02 - 1	MS-WATER	-15	0.0004	-	7470A	PPB	15 May 2018 10:40:04	U	SMPL
J2883-04 - 1	DRUM-1-5	-13	0.0010	-	7470A	PPB	15 May 2018 10:42:11	U	SMPL
J2883-04DUP - 1	DRUM-1-5DUP	-16	0.0002	-	7470A	PPB	15 May 2018 10:44:19	U	SMPL
J2883-04MS - 1	DRUM-1-5MS	13103	3.6082	-	7470A	PPB	15 May 2018 10:46:27	U	SMPL
J2883-04MSD - 1	DRUM-1-5MSD	13085	3.6032	-	7470A	PPB	15 May 2018 10:48:36	U	SMPL
PB109248BL - 1	PBW	-116	-0.0273	-	7470A	PPB	15 May 2018 10:50:44	U	SMPL
PB109248BS - 1	LCSW	14089	3.8793	-	7470A	PPB	15 May 2018 10:52:52	U	SMPL
CCV32 - 1	CCV32	18079	4.9767	-	7470A	PPB	15 May 2018 10:54:59	U	SMPL
CCB32 - 1	CCB32	-70	-0.0147	-	7470A	PPB	15 May 2018 10:57:07	U	SMPL
PB109248TB - 1	PB109208TB	-3	0.0037	-	7470A	PPB	15 May 2018 10:59:15	U	SMPL
J2674-04 - 1	HD-01-051118	9	0.0070	-	7470A	PPB	15 May 2018 11:01:22	U	SMPL
J2674-04DUP - 1	HD-01-051118DUP	31	0.0131	-	7470A	PPB	15 May 2018 11:03:29	U	SMPL
J2674-04MS - 1	HD-01-051118MS	12522	3.4484	-	7470A	PPB	15 May 2018 11:05:36	U	SMPL
J2674-04MSD - 1	HD-01-051118MSD	13085	3.6032	-	7470A	PPB	15 May 2018 11:07:44	U	SMPL
J2679-08 - 1	EO-05-051118-A	-112	-0.0262	-	7470A	PPB	15 May 2018 11:09:52	U	SMPL
J2679-10 - 1	EO-05-051118-B	-37	-0.0056	-	7470A	PPB	15 May 2018 11:12:00	U	SMPL
J2679-12 - 1	EO-05-051118-C	31	0.0131	-	7470A	PPB	15 May 2018 11:14:08	U	SMPL
PB109249BL - 1	PBW	-36	-0.0053	-	7470A	PPB	15 May 2018 11:16:17	U	SMPL
PB109249BS - 1	LCSW	13975	3.8480	-	7470A	PPB	15 May 2018 11:18:25	U	SMPL
CCV33 - 1	CCV33	18011	4.9580	-	7470A	PPB	15 May 2018 11:20:32	U	SMPL
CCB33 - 1	CCB33	-74	-0.0158	-	7470A	PPB	15 May 2018 11:22:40	U	SMPL
PB109249TB - 1	PB109224TB	-38	-0.0059	-	7470A	PPB	15 May 2018 11:24:48	U	SMPL
J2670-04 - 1	TR-05-051118	-3	0.0037	-	7470A	PPB	15 May 2018 11:26:55	U	SMPL
J2670-04DUP - 1	TR-05-051118DUP	23	0.0109	-	7470A	PPB	15 May 2018 11:29:02	U	SMPL
J2670-04MS - 1	TR-05-051118MS	13486	3.7135	-	7470A	PPB	15 May 2018 11:31:10	U	SMPL
J2670-04MSD - 1	TR-05-051118MSD	13732	3.7811	-	7470A	PPB	15 May 2018 11:33:16	U	SMPL
J2901-09 - 1	251-COMP1-SL	-113	-0.0265	-	7470A	PPB	15 May 2018 11:35:24	U	SMPL
J2901-10 - 1	251-COMP2-SL	-19	-0.0007	-	7470A	PPB	15 May 2018 11:37:33	U	SMPL
J2901-11 - 1	260-COMP1-SL	16	0.0090	-	7470A	PPB	15 May 2018 11:39:41	U	SMPL
J2901-12 - 1	260-COMP2-SL	10	0.0073	-	7470A	PPB	15 May 2018 11:41:50	U	SMPL
J2902-05 - 1	295-COMP1-SL	43	0.0164	-	7470A	PPB	15 May 2018 11:43:57	U	SMPL
CCV34 - 1	CCV34	18088	4.9791	-	7470A	PPB	15 May 2018 11:46:05	U	SMPL

LB95352 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 15 May 2018 08:47:52

Sample ID	Extended ID	μ Abs.	Conc.	Std Cond	Method	Units	Date	Type	Type
CCB34 - 1	CCB34	-75	-0.0161	-7470A	PPB		15 May 2018 11:48:13	U	SMPL
J2902-06 - 1	228-COMP1-SL	-24	-0.0020	-7470A	PPB		15 May 2018 11:50:21	U	SMPL
J2671-12LX5 - 1		55	0.0197	-7470A	PPB		15 May 2018 11:52:29	U	SMPL
J2671-12A - 1		13742	3.7839	-7470A	PPB		15 May 2018 11:54:37	U	SMPL
J2883-04LX5 - 1		-36	-0.0053	-7470A	PPB		15 May 2018 11:56:44	U	SMPL
J2883-04A - 1		13748	3.7855	-7470A	PPB		15 May 2018 11:58:52	U	SMPL
J2674-04LX5 - 1		-50	-0.0092	-7470A	PPB		15 May 2018 12:00:59	U	SMPL
J2674-04A - 1		13844	3.8119	-7470A	PPB		15 May 2018 12:03:08	U	SMPL
J2670-04LX5 - 1		-59	-0.0117	-7470A	PPB		15 May 2018 12:05:16	U	SMPL
J2670-04A - 1		14128	3.8901	-7470A	PPB		15 May 2018 12:07:25	U	SMPL
CCV35 - 1	CCV35	18356	5.0528	-7470A	PPB		15 May 2018 12:09:32	U	SMPL
CCB35 - 1	CCB35	-69	-0.0144	-7470A	PPB		15 May 2018 12:11:40	U	SMPL