



Analytical Summary Report

Analysis Method: 9040C

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB95373

Slope : 97.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.1	3.99	05/15/2018	15:25
2	CAL2	1	Water			21.1	7.01	05/15/2018	15:26
3	CAL3	1	Water			21.1	10.03	05/15/2018	15:27
4	ICV	1	Water			21.2	6.99	05/15/2018	15:29
5	CCV1	1	Water			21.2	2.01	05/15/2018	15:30
6	J2519-03	1	Water			22.4	5.64	05/15/2018	15:31
7	J2519-03DUP	1	Water			22.5	5.65	05/15/2018	15:32
8	CCV2	1	Water			21.1	12.01	05/15/2018	15:34

1895373

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrsivity j2519

WorkList ID : 112116

Date : 5/15/2018 3:15:59 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
05/15/2018	Water	J2519-03-18	Corrosivity	Cool 4 deg C	PSEG02	I62	OILY-WATER-TOTE	05/15/2018	9040C

Date/Time 05-15-18 3:12:01 PM
Received by: JD
Relinquished by: [Signature]

Date/Time 05-15-18 4:00 PM
Received by: [Signature]
Relinquished by: [Signature]