



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB95377

Slope : 98.3

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.2	4.01	05/15/2018	17:00
2	CAL2	1	Water			21.2	7.01	05/15/2018	17:01
3	CAL3	1	Water			21.2	10.02	05/15/2018	17:02
4	ICV	1	Water			21.3	7.00	05/15/2018	17:05
5	CCV1	1	Water			21.3	2.01	05/15/2018	17:06
6	J2361-03	1	Solid	20.02	20	21.2	6.50	05/15/2018	17:08
7	J2672-06	1	Solid	20.03	20	23.6	7.49	05/15/2018	17:10
8	J2679-14	1	Solid	20.05	20	22.2	6.01	05/15/2018	17:11
9	J2679-16	1	Solid	20.02	20	21.5	9.14	05/15/2018	17:12
10	J2679-18	1	Solid	20.03	20	22.1	6.26	05/15/2018	17:14
11	J2680-12	1	Solid	20.05	20	21.4	8.46	05/15/2018	17:15
12	J2680-14	1	Solid	20.02	20	21.7	7.88	05/15/2018	17:18
13	J2680-16	1	Solid	20.04	20	21.5	8.36	05/15/2018	17:20
14	J2680-18	1	Solid	20.05	20	22.2	10.00	05/15/2018	17:21
15	J2680-20	1	Solid	20.03	20	22.0	10.45	05/15/2018	17:22
16	CCV2	1	Water			21.2	12.02	05/15/2018	17:24
17	J2924-04	1	Solid	20.01	20	22.2	7.76	05/15/2018	17:25
18	J2924-06	1	Solid	20.02	20	21.9	8.22	05/15/2018	17:26
19	J2925-03	1	Solid	20.03	20	22.7	8.76	05/15/2018	17:29
20	J2925-03DUP	1	Solid	20.04	20	22.8	8.77	05/15/2018	17:30

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
21	CCV3	1	Water			21.2	2.01	05/15/2018	17:31

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosivity J2924

WorkList ID : 112115

Date : 5/15/2018 3:15:13 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
05/15/2018	Solid	J2361-03	Corrosivity	Cool 4 deg C	PSEG01	I63	JC-RT-4763-OILY-WIPES	05/15/2018	9045C
05/15/2018	Solid	J2672-06	Corrosivity	Cool 4 deg C	PSEG05		BU-03-051518	05/15/2018	9045C
05/15/2018	Solid	J2679-14	Corrosivity	Cool 4 deg C	PSEG05		EO-01-051518-A	05/15/2018	9045C
05/15/2018	Solid	J2679-16	Corrosivity	Cool 4 deg C	PSEG05		EO-01-051518-B	05/15/2018	9045C
05/15/2018	Solid	J2679-18	Corrosivity	Cool 4 deg C	PSEG05		EO-01-051518-C	05/15/2018	9045C
05/15/2018	Solid	J2680-12	Corrosivity	Cool 4 deg C	PSEG05		PL-01-051518-A	05/15/2018	9045C
05/15/2018	Solid	J2680-14	Corrosivity	Cool 4 deg C	PSEG05		PL-01-051518-B	05/15/2018	9045C
05/15/2018	Solid	J2680-16	Corrosivity	Cool 4 deg C	PSEG05		PL-01-051518-C	05/15/2018	9045C
05/15/2018	Solid	J2680-18	Corrosivity	Cool 4 deg C	PSEG05		PL-01-051518-D	05/15/2018	9045C
05/15/2018	Solid	J2680-20	Corrosivity	Cool 4 deg C	PSEG05		PL-01-051518-E	05/15/2018	9045C
05/15/2018	Solid	J2924-04	Corrosivity	Cool 4 deg C	PSEG01	I63	TP-12A	05/15/2018	9045C
05/15/2018	Solid	J2924-06	Corrosivity	Cool 4 deg C	PSEG01	I63	TP-12B	05/15/2018	9045C
05/14/2018	Solid	J2925-03	Corrosivity	Cool 4 deg C	LIRO01	I62	259-COMP1-SL	05/14/2018	9045C

Date/Time 05.15.18 5:20 PM
 Received by: 30
 Relinquished by: 30

Date/Time 05.15.18 5:40 PM
 Received by: 30
 Relinquished by: 30