



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: KETANKUMAR

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB95394

Ambient Barometric Pressure (mmHg): 760

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	27.0	80.6	80.6	05/16/2018	09:15
2	J2519-03		1	>100.0	>212.0	>212.0	05/16/2018	09:46
3	J2519-03DUP		1	>100.0	>212.0	>212.0	05/16/2018	10:20

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.33

LB 95394.

WORKLIST(Hardcopy Internal Chain)

WorkList Name : FP-051618 WorkList ID : 112143 Date : 5/16/2018 8:08:13 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
05/18/2018	Water	J2519-03 H	Flash Point	Cool 4 deg C	PSEG02	I62	OILY-WATER-TOTE	05/15/2018	1010A

Date/Time 05/16/18 8:30 AM
Received by: KS
Relinquished by: CP

Date/Time 05/16/18 10:25 AM
Received by: CP
Relinquished by: KS