



Analytical Summary Report

Analysis Method: SM5220 D

ANALYST: Indiana

Parameter: COD

SUPERVISOR REVIEW BY: apatel

Run Number: LB95398

Reagent/Standard	Lot/Log #
COD ICV-LCS std, 50ppm	WP64505
COD calibration std. 100 ppm	WP64503
COD calibration std. 10 ppm	WP64501
COD calibration std. 150 ppm	WP64504
COD calibration std. 50 ppm	WP64502
COD calibration std. 0 ppm	WP64500
COD Digestion Vials Low Level 0-150Mg/L	W2346
COD CCV std, 50ppm	WP64548

Temp In (C): <u>150</u>	Date In: <u>05/16/2018</u>	Time In: <u>11:00</u>
Temp Out (C): <u>150</u>	Date Out: <u>05/16/2018</u>	Time Out: <u>13:01</u>

Intercept: -0.8741

Slope: 1.0302

Regression: 1

Seq	Lab ID	TrueValue (mg/l)	DF	MATRIX	Reading	Anal Date	Anal Time
1	CAL1	0	1	Water	0.000	05/15/2018	12:20
2	CAL2	10	1	Water	9.000	05/15/2018	12:20
3	CAL3	50	1	Water	50.000	05/15/2018	12:21
4	CAL4	100	1	Water	102.000	05/15/2018	12:21
5	CAL5	150	1	Water	154.000	05/15/2018	12:22



Analytical Summary Report

Analysis Method: SM5220 D

ANALYST: Indiana

Parameter: COD

SUPERVISOR REVIEW BY: apatel

Run Number: LB95398

Seq	Lab ID	True Value (mg/l)	Initial Weight (g)	Final Vol (ml)	DF	MATRIX	Reading	Result	AnalDate	AnalTime
1	ICV	50			1	Water	51.000	50.353	05/15/2018	12:22
2	ICB				1	Water	0.000	0.848	05/15/2018	12:23
3	CCV1	50			1	Water	50.000	49.383	05/16/2018	14:03
4	CCB1				1	Water	0.000	0.848	05/16/2018	14:03
5	LB95398BL				1	Water	0.000	0.848	05/16/2018	14:04
6	LB95398BS	50			1	Water	51.000	50.353	05/16/2018	14:04
7	J2856-01				1	Water	0.000	0.848	05/16/2018	14:05
8	J2856-02				1	Water	0.000	0.848	05/16/2018	14:05
9	J2856-03				1	Water	0.000	0.848	05/16/2018	14:06
10	J2856-03DUP				1	Water	0.000	0.848	05/16/2018	14:06
11	J2856-04	50			1	Water	50.000	49.383	05/16/2018	14:07
12	J2856-05	50			1	Water	51.000	50.353	05/16/2018	14:07
13	J2856-06				1	Water	0.000	0.848	05/16/2018	14:08
14	J2858-01				1	Water	40.000	39.676	05/16/2018	14:08
15	CCV2	50			1	Water	51.000	50.353	05/16/2018	14:09
16	CCB2				1	Water	0.000	0.848	05/16/2018	14:09
17	J2858-02				1	Water	0.000	0.848	05/16/2018	14:10
18	J2858-03				1	Water	60.000	59.090	05/16/2018	14:10
19	J2858-04				1	Water	32.000	31.910	05/16/2018	14:11
20	J2860-01				1	Water	9.000	9.585	05/16/2018	14:11
21	J2860-02				1	Water	11.000	11.526	05/16/2018	14:12
22	J2860-03				1	Water	43.000	42.588	05/16/2018	14:12
23	J2860-04				1	Water	61.000	60.060	05/16/2018	14:13
24	J2860-05				1	Water	11.000	11.526	05/16/2018	14:13
25	CCV3	50			1	Water	51.000	50.353	05/16/2018	14:14
26	CCB3				1	Water	0.000	0.848	05/16/2018	14:14