



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB95404

Slope : 98.2

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

True Value of CCV4 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.3	4.01	05/16/2018	16:35
2	CAL2	1	Water			21.3	7.00	05/16/2018	16:36
3	CAL3	1	Water			21.3	10.03	05/16/2018	16:37
4	ICV	1	Water			21.3	6.99	05/16/2018	16:39
5	CCV1	1	Water			21.2	2.01	05/16/2018	16:40
6	J2669-08	1	Solid	20.02	20	21.8	9.79	05/16/2018	16:41
7	J2669-08DUP	1	Solid	20.03	20	21.9	9.80	05/16/2018	16:42
8	J2935-02	1	Solid	20.02	20	22.1	7.62	05/16/2018	16:44
9	J2936-02	1	Solid	20.04	20	21.4	7.51	05/16/2018	16:45
10	J2936-04	1	Solid	20.02	20	21.2	7.84	05/16/2018	16:46
11	J2936-06	1	Solid	20.01	20	21.4	7.54	05/16/2018	16:47
12	J2936-08	1	Solid	20.03	20	21.5	7.53	05/16/2018	16:48
13	J2944-02	1	Solid	20.05	20	21.4	8.34	05/16/2018	16:49
14	J2945-04	1	Solid	20.03	20	21.9	6.02	05/16/2018	16:50
15	J2963-13	1	Solid	20.02	20	22.0	9.37	05/16/2018	16:51
16	CCV2	1	Water			21.2	12.02	05/16/2018	17:00
17	J2963-14	1	Solid	20.03	20	22.1	9.22	05/16/2018	17:01
18	J2963-15	1	Solid	20.05	20	22.3	9.63	05/16/2018	17:04
19	J2963-16	1	Solid	20.02	20	21.6	10.60	05/16/2018	17:05
20	J2963-17	1	Solid	20.04	20	21.5	9.63	05/16/2018	17:06

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
21	J2963-18	1	Solid	20.02	20	21.8	8.76	05/16/2018	17:09
22	J2964-03	1	Solid	20.03	20	21.5	7.93	05/16/2018	17:10
23	J2965-05	1	Solid	20.04	20	21.8	10.93	05/16/2018	17:11
24	J2965-06	1	Solid	20.03	20	22.5	9.34	05/16/2018	17:12
25	J2966-04	1	Solid	20.04	20	21.9	8.06	05/16/2018	17:13
26	J2966-08	1	Solid	20.05	20	21.6	6.66	05/16/2018	17:15
27	CCV3	1	Water			21.2	2.01	05/16/2018	17:16
28	J2967-06	1	Solid	20.02	20	21.5	6.73	05/16/2018	17:18
29	J2967-06DUP	1	Solid	20.03	20	21.6	6.74	05/16/2018	17:20
30	CCV4	1	Water			21.2	12.01	05/16/2018	17:22

2905404

Date : 5/16/2018 2:39:06 PM

Fractal

Date/Time
Received
Relinquish