



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB95409

Slope : 97.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.1	4.00	05/17/2018	08:00
2	CAL2	1	Water			21.1	7.01	05/17/2018	08:01
3	CAL3	1	Water			21.1	10.02	05/17/2018	08:02
4	ICV	1	Water			21.1	6.99	05/17/2018	08:03
5	CCV1	1	Water			21.2	2.02	05/17/2018	08:05
6	J2966-09	1	Water			19.1	6.57	05/17/2018	08:06
7	J2966-09DUP	1	Water			19.2	6.58	05/17/2018	08:08
8	CCV2	1	Water			21.1	12.02	05/17/2018	08:10

12 05409

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH J2966

WorkList ID : 112189

Date : 5/17/2018 7:43:22 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
05/16/2018	Water	J2966-09	pH	Cool 4 deg C	PSEG01	I13	BUR-DRUM-1-TREN-DRUM-1	05/16/2018	9045C

Date/Time 05-17-18 7:43 AM
 Received by: JD
 Relinquished by: JD

Date/Time 05-17-18 8:32 AM
 Received by: JD
 Relinquished by: JD