



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB95410

Slope : 97.6

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.1	4.00	05/17/2018	08:00
2	CAL2	1	Water			21.1	7.01	05/17/2018	08:01
3	CAL3	1	Water			21.1	10.02	05/17/2018	08:02
4	ICV	1	Water			21.1	6.99	05/17/2018	08:03
5	CCV1	1	Water			21.2	2.01	05/17/2018	08:15
6	J2954-01	1	Solid	20.02	20	21.2	5.93	05/17/2018	08:20
7	J2954-01DUP	1	Solid	20.03	20	21.3	5.94	05/17/2018	08:21
8	J2957-01	1	Solid	20.05	20	21.0	5.34	05/17/2018	08:22
9	CCV2	1	Water			21.1	12.02	05/17/2018	08:24

1205410

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH J2954

WorkList ID : 112190

Date : 5/17/2018 7:44:01 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
05/16/2018	Solid	J2954-01-17	pH	Cool 4 deg C	TULL02	I53	FK-PS-01-026	05/16/2018	9045C
05/16/2018	Solid	J2957-01-17	pH	Cool 4 deg C	TULL02	I51	FK-PS-01.026	05/16/2018	9045C

Date/Time 05-17-18 7:46 AM
 Received by: JD
 Relinquished by: JD

Date/Time 05-17-18 9:00 AM
 Received by: JD
 Relinquished by: JD