



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: KETANKUMAR

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB95432

Ambient Barometric Pressure (mmHg): 755

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	27.0	80.6	82.3	05/17/2018	09:10
2	J2966-10		1	>100.0	>212.0	>212.0	05/17/2018	09:35
3	J2966-10DUP		1	>100.0	>212.0	>212.0	05/17/2018	10:02

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.33

LB95432

WORKLIST(Hardcopy Internal Chain)

WorkList Name : FP-051718

WorkList ID : 112209

Date : 5/17/2018 8:19:46 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
05/29/2018	Water	J2966-10 B	Flash Point	Cool 4 deg C	PSEG01	I13	BUR-DRUM-1-TREN-DRUM-1	05/16/2018	1010A

Date/Time 05/17/18 8:30 AM
Received by: KS
Relinquished by: CP

Date/Time 05/17/18 10:15 AM
Received by: CP
Relinquished by: KS