


7471B

LB 95439
INSTRUMENT ID: CV 2

Linear

 μ Abs.:

47090



A= 0.0000e+000

B= 2.1607e-004

C= -9.0173e-002

Rho= 0.9997435

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.051	-0.051	180	0.000	180				
0.2	0.200	0.224	0.024	1454	0.0 %	1454				
2.5	2.500	2.589	0.089	12398	0.0 %	12398				
5.0	5.000	5.008	0.008	23594	0.0 %	23594				
7.5	7.500	7.346	-0.154	34417	0.0 %	34417				
10.0	10.000	10.085	0.085	47090	0.0 %	47090				

LB95439 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 17 May 2018 17:12:21

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	180	-	0.0000	7471B	PPB	17 May 2018 17:21:28	S	Std
0.2 - 1	50.2	1454	-	0.2000	7471B	PPB	17 May 2018 17:23:46	S	Std
2.5 - 1	52.5	12398	-	2.5000	7471B	PPB	17 May 2018 17:26:04	S	Std
5.0 - 1	55	23594	-	5.0000	7471B	PPB	17 May 2018 17:28:21	S	Std
7.5 - 1	57.5	34417	-	7.5000	7471B	PPB	17 May 2018 17:30:38	S	Std
10.0 - 1	57.5	47090	-	10.0000	7471B	PPB	17 May 2018 17:38:11	S	Std
ICV53 - 1	ICV53	19968	4.2243	-7471B	PPB	PPB	17 May 2018 17:40:48	U	SMPL
ICB53 - 1	ICB53	145	-0.0588	-7471B	PPB	PPB	17 May 2018 17:43:05	U	SMPL
CCV74 - 1	CCV74	23300	4.9443	-7471B	PPB	PPB	17 May 2018 17:45:22	U	SMPL
CCB74 - 1	CCB74	545	0.0276	-7471B	PPB	PPB	17 May 2018 17:47:37	U	SMPL
CRA - 1	CRA	1419	0.2164	-7471B	PPB	PPB	17 May 2018 17:49:54	U	SMPL
HighStd - 1	HighStd	47107	10.0882	-7471B	PPB	PPB	17 May 2018 17:52:09	U	SMPL
ChkStd - 1	ChkStd	32390	6.9083	-7471B	PPB	PPB	17 May 2018 17:54:25	U	SMPL
PB109366BL - 1	PBS	420	0.0006	-7471B	PPB	PPB	17 May 2018 18:04:13	U	SMPL
PB109366BS - 1	LCSS	18128	3.8267	-7471B	PPB	PPB	17 May 2018 18:06:29	U	SMPL
J2669-07 - 1	SU-03-051618	963	0.1179	-7471B	PPB	PPB	17 May 2018 18:08:46	U	SMPL
J2935-01 - 1	NFS-OILY-STONE	-10	-0.0923	-7471B	PPB	PPB	17 May 2018 18:11:03	U	SMPL
J2936-01 - 1	TP-1-WC-1	959	0.1170	-7471B	PPB	PPB	17 May 2018 18:13:20	U	SMPL
J2936-03 - 1	TP-1-WC-2	395	-0.0048	-7471B	PPB	PPB	17 May 2018 18:15:38	U	SMPL
J2936-05 - 1	TP-1-WC-3	562	0.0313	-7471B	PPB	PPB	17 May 2018 18:17:54	U	SMPL
CCV75 - 1	CCV75	23467	4.9803	-7471B	PPB	PPB	17 May 2018 18:20:10	U	SMPL
CCB75 - 1	CCB75	329	-0.0191	-7471B	PPB	PPB	17 May 2018 18:22:25	U	SMPL
J2936-07 - 1	TP-1-WC-4	974	0.1203	-7471B	PPB	PPB	17 May 2018 18:24:42	U	SMPL
J2944-01 - 1	LSCY-STONE	113	-0.0658	-7471B	PPB	PPB	17 May 2018 18:26:58	U	SMPL
J2945-01 - 1	RO-1-4	1890	0.3182	-7471B	PPB	PPB	17 May 2018 18:29:14	U	SMPL
J2945-01DUP - 1	RO-1-4DUP	1954	0.3320	-7471B	PPB	PPB	17 May 2018 18:31:30	U	SMPL
J2945-01MS - 1	RO-1-4MS	19211	4.0607	-7471B	PPB	PPB	17 May 2018 18:33:47	U	SMPL
J2945-01MSD - 1	RO-1-4MSD	19241	4.0672	-7471B	PPB	PPB	17 May 2018 18:36:04	U	SMPL
J2950-01 - 1	EPE-MX-3(0-5)	892	0.1026	-7471B	PPB	PPB	17 May 2018 18:38:21	U	SMPL
J2950-02 - 1	EPE-MX-3(5-10)	1531	0.2406	-7471B	PPB	PPB	17 May 2018 18:40:39	U	SMPL
J2953-01 - 1	FK-SF-02-003-A	2357	0.4191	-7471B	PPB	PPB	17 May 2018 18:42:57	U	SMPL
J2953-03 - 1	FK-SF-02-003-B	1969	0.3353	-7471B	PPB	PPB	17 May 2018 18:45:13	U	SMPL
CCV76 - 1	CCV76	23340	4.9529	-7471B	PPB	PPB	17 May 2018 18:47:29	U	SMPL
CCB76 - 1	CCB76	358	-0.0128	-7471B	PPB	PPB	17 May 2018 18:49:45	U	SMPL
J2954-01 - 1	FK-PS-01-026	599	0.0393	-7471B	PPB	PPB	17 May 2018 18:52:02	U	SMPL
J2957-01 - 1	FK-PS-01.026	1353	0.2022	-7471B	PPB	PPB	17 May 2018 18:54:19	U	SMPL
J2964-01 - 1	BIN-7	2873	0.5306	-7471B	PPB	PPB	17 May 2018 18:56:35	U	SMPL
J2966-01 - 1	BUR-DRUM-1-2	594	0.0382	-7471B	PPB	PPB	17 May 2018 18:58:52	U	SMPL
J2966-05 - 1	TREN-DRUM-1	560	0.0308	-7471B	PPB	PPB	17 May 2018 19:01:08	U	SMPL
J2967-01 - 1	RS-SOIL	311	-0.0230	-7471B	PPB	PPB	17 May 2018 19:03:25	U	SMPL
J2945-01LX5 - 1		927	0.1101	-7471B	PPB	PPB	17 May 2018 19:05:43	U	SMPL
J2945-01A - 1		19127	4.0426	-7471B	PPB	PPB	17 May 2018 19:08:00	U	SMPL
CCV77 - 1	CCV77	23375	4.9605	-7471B	PPB	PPB	17 May 2018 19:10:18	U	SMPL
CCB77 - 1	CCB77	310	-0.0232	-7471B	PPB	PPB	17 May 2018 19:12:34	U	SMPL