

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

5/18/2018 9:16

Test: Phenolics-mg/L

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	1.002	0.0	0.109	
ICB1	0.003	0.0	0.004	
CCV1	1.017	0.0	0.110	
CCB1	0.008	0.0	0.004	
PB109408BL	0.000	0.0	0.003	
PB109408BS	1.013	0.0	0.110	
J2923-05	0.042	0.0	0.008	
J2923-06	0.044	0.0	0.008	
J2923-07	0.040	0.0	0.008	
J2923-08	0.039	0.0	0.008	
J2943-01	0.004	0.0	0.004	
J2943-01DUP	-0.002	0.0	0.003	
J2943-01MS	0.928	0.0	0.101	
J2943-01MSD	0.924	0.0	0.101	
CCV2	1.004	0.0	0.109	
CCB2	0.004	0.0	0.004	
J2943-02	-0.000	0.0	0.003	
J2943-03	0.003	0.0	0.004	
CCV3	0.997	0.0	0.108	
CCB3	0.001	0.0	0.003	

N 20
Mean 0.354
SD 0.4751
CV% 134.38

AS

Aquakem v. 7.2AQ1

Results from time period:

Fri May 18 08:05:22 2018

Fri May 18 08:50:48 2018

Sample Id	Sam/Ctr/c	Test short na	Test type	Result	Result unit	Result date and time	Stat
0.05PPM	A	Phenolics-	P	0.0494	mg/l	5/18/2018 8:05:22	
0.1PPM	A	Phenolics-	P	0.1015	mg/l	5/18/2018 8:05:23	
0.25PPM	A	Phenolics-	P	0.2512	mg/l	5/18/2018 8:05:24	
0.50PPM	A	Phenolics-	P	0.4988	mg/l	5/18/2018 8:05:25	
1.0PPM	A	Phenolics-	P	0.998	mg/l	5/18/2018 8:05:26	
2.0PPM	A	Phenolics-	P	2.0011	mg/l	5/18/2018 8:05:27	
ICV1	S	Phenolics-	P	1.0021	mg/l	5/18/2018 8:42:57	
ICB1	S	Phenolics-	P	0.0025	mg/l	5/18/2018 8:42:58	
CCV1	S	Phenolics-	P	1.0169	mg/l	5/18/2018 8:42:59	
CCB1	S	Phenolics-	P	0.0079	mg/l	5/18/2018 8:43:00	
PB109408BL	S	Phenolics-	P	0.0005	mg/l	5/18/2018 8:43:01	
PB109408BS	S	Phenolics-	P	1.0133	mg/l	5/18/2018 8:43:02	
J2923-05	S	Phenolics-	P	0.0422	mg/l	5/18/2018 8:43:03	
J2923-06	S	Phenolics-	P	0.0444	mg/l	5/18/2018 8:43:04	
J2923-07	S	Phenolics-	P	0.0396	mg/l	5/18/2018 8:43:05	
J2923-08	S	Phenolics-	P	0.0395	mg/l	5/18/2018 8:43:06	
J2943-01	S	Phenolics-	P	0.0037	mg/l	5/18/2018 8:43:07	
J2943-01DUP	S	Phenolics-	P	-0.0021	mg/l	5/18/2018 8:43:08	
J2943-01MS	S	Phenolics-	P	0.928	mg/l	5/18/2018 8:50:39	
J2943-01MSD	S	Phenolics-	P	0.9243	mg/l	5/18/2018 8:50:40	
CCV2	S	Phenolics-	P	1.0038	mg/l	5/18/2018 8:50:41	
CCB2	S	Phenolics-	P	0.004	mg/l	5/18/2018 8:50:42	
J2943-02	S	Phenolics-	P	-0.0002	mg/l	5/18/2018 8:50:43	
J2943-03	S	Phenolics-	P	0.003	mg/l	5/18/2018 8:50:44	
CCV3	S	Phenolics-	P	0.9969	mg/l	5/18/2018 8:50:45	
CCB3	S	Phenolics-	P	0.001	mg/l	5/18/2018 8:50:46	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

5/18/2018 8:13


Test Phenolics-

Accepted 5/18/2018 8:13

Factor 9.502
Bias 0.003

Coeff. of det. 0.999996

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	Phenol-2	0.009	0.0494	0.0500	
2	Phenol-2	0.014	0.1015	0.1000	
3	Phenol-2	0.030	0.2512	0.2500	
4	Phenol-2	0.056	0.4988	0.5000	
5	Phenol-2	0.108	0.9980	1.0000	
6	Phenol-2	0.214	2.0011	2.0000	