

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

5/18/2018 10:06

Test: Phenolics-mg/d

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	0.999	0.0	0.109	
ICB1	-0.000	0.0	0.003	
CCV1	1.008	0.0	0.109	
CCB1	-0.000	0.0	0.003	
PB109407BL	-0.001	0.0	0.003	
PB109407BS	1.012	0.0	0.110	
J2923-01	0.046	0.0	0.008	
J2923-02	0.043	0.0	0.008	
J2923-03	0.039	0.0	0.007	
J2923-04	0.040	0.0	0.008	
CCV2	1.021	0.0	0.111	
CCB2	-0.001	0.0	0.003	

N	12
Mean	0.350
SD	0.4873
CV%	139.06

AS

Aquakem v. 7.2AQ1

Results from time period:

Fri May 18 09:42:54 2018

Fri May 18 09:43:05 2018

Sample Id	Sam/Ctr/c/	Test short nan	Test type	Result	Result unit	Result date and time	Stat
0.05PPM	A	Phenolics-	P	0.0494	mg/l	5/18/2018 8:05:22	
0.1PPM	A	Phenolics-	P	0.1015	mg/l	5/18/2018 8:05:23	
0.25PPM	A	Phenolics-	P	0.2512	mg/l	5/18/2018 8:05:24	
0.50PPM	A	Phenolics-	P	0.4988	mg/l	5/18/2018 8:05:25	
1.0PPM	A	Phenolics-	P	0.998	mg/l	5/18/2018 8:05:26	
2.0PPM	A	Phenolics-	P	2.0011	mg/l	5/18/2018 8:05:27	
ICV1	S	Phenolics-	P	0.9991	mg/l	5/18/2018 9:42:54	
ICB1	S	Phenolics-	P	-0.0004	mg/l	5/18/2018 9:42:55	
CCV1	S	Phenolics-	P	1.0078	mg/l	5/18/2018 9:42:56	
CCB1	S	Phenolics-	P	-0.0004	mg/l	5/18/2018 9:42:57	
PB109407BL	S	Phenolics-	P	-0.0007	mg/l	5/18/2018 9:42:58	
PB109407BS	S	Phenolics-	P	1.0117	mg/l	5/18/2018 9:42:59	
J2923-01	S	Phenolics-	P	0.0461	mg/l	5/18/2018 9:43:00	
J2923-02	S	Phenolics-	P	0.0433	mg/l	5/18/2018 9:43:01	
J2923-03	S	Phenolics-	P	0.0392	mg/l	5/18/2018 9:43:02	
J2923-04	S	Phenolics-	P	0.0398	mg/l	5/18/2018 9:43:03	
CCV2	S	Phenolics-	P	1.0207	mg/l	5/18/2018 9:43:04	
CCB2	S	Phenolics-	P	-0.0008	mg/l	5/18/2018 9:43:05	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

5/18/2018 8:13


Test Phenolics-

Accepted 5/18/2018 8:13

Factor 9.502
Bias 0.003

Coeff. of det. 0.999996

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	Phenol-2	0.009	0.0494	0.0500	
2	Phenol-2	0.014	0.1015	0.1000	
3	Phenol-2	0.030	0.2512	0.2500	
4	Phenol-2	0.056	0.4988	0.5000	
5	Phenol-2	0.108	0.9980	1.0000	
6	Phenol-2	0.214	2.0011	2.0000	