

Daily Analysis Runlog For Sequence/QC Batch ID # LB95453

Review By		Review On				
SubDirectory		LB95453		Test : Phenolics		
STD. NAME		STD REF.#				
ICAL Standard		WP64598				
ICV Standard		WP64600				
CCV Standard		WP64599				
ICSA Standard		N/A				
CRI Standard		N/A				
LCS Standard		WP64432				
Chk Standard		WP64602,WP63060,WP63061				
Sr#	SampleID	ClientID	QcType	Date	Comment	Status
1	0.05PPM	0.05PPM	CAL1	05/18/18 08:05		OK
2	0.1PPM	0.1PPM	CAL2	05/18/18 08:05		OK
3	0.25PPM	0.25PPM	CAL3	05/18/18 08:05		OK
4	0.50PPM	0.50PPM	CAL4	05/18/18 08:05		OK
5	1.0PPM	1.0PPM	CAL5	05/18/18 08:05		OK
6	2.0PPM	2.0PPM	CAL6	05/18/18 08:05		OK
7	ICV1	ICV1	ICV	05/18/18 09:42		OK
8	ICB1	ICB1	ICB	05/18/18 09:42		OK
9	CCV1	CCV1	CCV	05/18/18 09:42		OK
10	CCB1	CCB1	CCB	05/18/18 09:42		OK
11	PB109407BL	PB109407BL	MB	05/18/18 09:42		OK
12	PB109407BS	PB109407BS	LCS	05/18/18 09:42		OK
13	J2923-01	LOQ-SCM-Precision-	LOQ	05/18/18 09:43	LOQ fail	Not Ok
14	J2923-02	LOQ-SCM-Precision-	LOQ	05/18/18 09:43	LOQ fail	Not Ok
15	J2923-03	LOQ-SCM-Precision-	LOQ	05/18/18 09:43	LOQ fail	Not Ok
16	J2923-04	LOQ-SCM-Precision-	LOQ	05/18/18 09:43	LOQ fail	Not Ok
17	CCV2	CCV2	CCV	05/18/18 09:43		OK
18	CCB2	CCB2	CCB	05/18/18 09:43		OK



GENCHEM

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: LB95453

SDG No.: LB95453

Contract: _____

Lab Code: CHEM

Case No.: LB95453

SAS No.: LB95453

Initial Calibration Source: _____

Continuing Calibration Source: _____

Sample ID	Analyte	Result mg/L	True Value	% Recovery	Qual	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV1	Phenolics	1	1	100		90 - 110		05/18/2018	09:42	lb95453



GENCHEM

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: LB95453

SDG No.: LB95453

Contract: _____ Lab Code: CHEM

Case No.: LB95453

SAS No.: LB95453

Initial Calibration Source: _____

Continuing Calibration Source: _____

Sample ID	Analyte	Result mg/L	True Value	% Recovery	Qual	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV1	Phenolics	1	1	100		90 - 110		05/18/2018	09:42	lb95453
CCV2	Phenolics	1	1	100		90 - 110		05/18/2018	09:43	lb95453



GENCHEM

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: LB95453

SDG No.: LB95453

Contract:

Lab Code: CHEM

Case No.: LB95453

SAS No.: LB95453

Sample ID	Analyte	Result mg/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB1	Phenolics	-0.0004	+/-0.05	U		0.05		05/18/2018	09:42	1b95453



GENCHEM

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: LB95453

SDG No.: LB95453

Contract:

Lab Code: CHEM

Case No.: LB95453

SAS No.: LB95453

Sample ID	Analyte	Result mg/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB1	Phenolics	-0.0004	+/-0.05	U		0.05		05/18/2018	09:42	1b95453
CCB2	Phenolics	-0.0008	+/-0.05	U		0.05		05/18/2018	09:43	1b95453



GENCHEM

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: LB95453

SDG No.: LB95453

Contract: Lab Code: CHEM

Case No.: LB95453

SAS No.: LB95453

Sample ID	Analyte	Result mg/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
-----------	---------	----------------	---------------------	--------------	-----	------	---	------------------	------------------	---------------



GENCHEM
- 3b -
PREPARATION BLANK SUMMARY

Client: LB95453

SDG No.: LB95453

Instrument: Konelab 20

Sample ID	Analyte	Result (mg/Kg)	Acceptance Limit	Conc Qual	LOD mg/Kg	CRQL mg/Kg	M	Analysis Date	Analysis Time	Run
PB109407BL		SOLID		Batch Number:		PB109407		Prep Date:	05/17/2018	
	Phenolics	-0.034	<2.5	U		2.5		05/18/2018	09:42	lb95453



GENCHEM

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: LB95453

SDG No.: LB95453

Contract:

Lab Code: CHEM

Case No.: LB95453

SAS No.: LB95453

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
J2923-01 Phenolics	mg/Kg	4.8	2.2	J	45.8	50 - 150	



GENCHEM

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: LB95453

SDG No.: LB95453

Contract:

Lab Code: CHEM

Case No.: LB95453

SAS No.: LB95453

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
J2923-02 Phenolics	mg/Kg	4.8	2.1	J	43.7	50 - 150	



GENCHEM

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: LB95453

SDG No.: LB95453

Contract:

Lab Code: CHEM

Case No.: LB95453

SAS No.: LB95453

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
J2923-03 Phenolics	mg/Kg	4.8	1.9	J	39.6	50 - 150	



GENCHEM

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: LB95453

SDG No.: LB95453

Contract:

Lab Code: CHEM

Case No.: LB95453

SAS No.: LB95453

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
J2923-04 Phenolics	mg/Kg	4.7	1.9	J	40.4	50 - 150	



GENCHEM

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: LB95453

SDG No.: LB95453

Contract:

Lab Code: CHEM

Case No.: LB95453

SAS No.: LB95453

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB109407BS Phenolics	mg/Kg	48.5	49.1		101.2	80 - 120	