

LB95459


7471B

INSTRUMENT 10: CV4

Linear

 μ Abs.:

36707



A= 0.0000e+000

B= 2.7179e-004

C= -6.9248e-002

Rho= 0.9998660

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.057	-0.057	46	0.000	46				
0.2	0.200	0.183	-0.017	928	0.0 %	928				
2.5	2.500	2.535	0.035	9583	0.0 %	9583				
5.0	5.000	5.065	0.065	18892	0.0 %	18892				
7.5	7.500	7.565	0.065	28090	0.0 %	28090				
10.0	10.000	9.907	-0.093	36707	0.0 %	36707				

LB95459 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 18 May 2018 13:50:57

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	46	-	0.00007471B	PPB		18 May 2018 13:51:06	S	Std
0.2 - 1	50.2	928	-	0.20007471B	PPB		18 May 2018 13:53:23	S	Std
2.5 - 1	52.5	9583	-	2.50007471B	PPB		18 May 2018 13:55:41	S	Std
5.0 - 1	55	18892	-	5.00007471B	PPB		18 May 2018 13:57:58	S	Std
7.5 - 1	57.5	28090	-	7.50007471B	PPB		18 May 2018 14:00:15	S	Std
10.0 - 1	510	36707	-	10.00007471B	PPB		18 May 2018 14:05:15	S	Std
ICV55 - 1	ICV55	14470	3.8636	-7471B	PPB		18 May 2018 14:08:00	U	SMPL
ICB55 - 1	ICB55	-107	-0.0983	-7471B	PPB		18 May 2018 14:10:16	U	SMPL
CCV85 - 1	CCV85	18949	5.0810	-7471B	PPB		18 May 2018 14:12:32	U	SMPL
CCB85 - 1	CCB85	-74	-0.0894	-7471B	PPB		18 May 2018 14:14:48	U	SMPL
CRA - 1	CRA	915	0.1794	-7471B	PPB		18 May 2018 14:17:04	U	SMPL
HighStd - 1	HighStd	37161	10.0309	-7471B	PPB		18 May 2018 14:19:19	U	SMPL
ChkStd - 1	ChkStd	25922	6.9762	-7471B	PPB		18 May 2018 14:21:36	U	SMPL
PB109414BL - 1	PBS	-17	-0.0739	-7471B	PPB		18 May 2018 14:27:47	U	SMPL
PB109414BS - 1	LCSS	14084	3.7587	-7471B	PPB		18 May 2018 14:30:03	U	SMPL
J2673-01 - 1	JC-03-051718-A	9989	2.6457	-7471B	PPB		18 May 2018 14:32:20	U	SMPL
J2673-03 - 1	JC-03-051718-B	3321	0.8334	-7471B	PPB		18 May 2018 14:34:37	U	SMPL
J2673-05 - 1	JC-03-051718-C	6608	1.7268	-7471B	PPB		18 May 2018 14:36:55	U	SMPL
J2673-07 - 1	JC-03-051718-D	8128	2.1399	-7471B	PPB		18 May 2018 14:39:13	U	SMPL
J2673-09 - 1	JC-03-051718-E	2940	0.7298	-7471B	PPB		18 May 2018 14:41:29	U	SMPL
CCV86 - 1	CCV86	18675	5.0065	-7471B	PPB		18 May 2018 14:43:44	U	SMPL
CCB86 - 1	CCB86	-104	-0.0975	-7471B	PPB		18 May 2018 14:46:00	U	SMPL
J2673-11 - 1	JC-03-051718-F	4609	1.1835	-7471B	PPB		18 May 2018 14:48:16	U	SMPL
J2673-13 - 1	JC-03-051718-G	6451	1.6841	-7471B	PPB		18 May 2018 14:50:31	U	SMPL
J2673-15 - 1	JC-03-051718-H	3012	0.7494	-7471B	PPB		18 May 2018 14:52:48	U	SMPL
J2677-05 - 1	OK-02-051718	1683	0.3882	-7471B	PPB		18 May 2018 14:55:04	U	SMPL
J2971-01 - 1	TP6-WC-1	585	0.0898	-7471B	PPB		18 May 2018 14:57:20	U	SMPL
J2971-03 - 1	TP6-WC-2	580	0.0884	-7471B	PPB		18 May 2018 14:59:37	U	SMPL
J2971-05 - 1	TP6-WC-3	602	0.0944	-7471B	PPB		18 May 2018 15:01:55	U	SMPL
J2971-07 - 1	TP6-WC-4	596	0.0927	-7471B	PPB		18 May 2018 15:04:12	U	SMPL
J2971-09 - 1	TP6-WC-5	875	0.1686	-7471B	PPB		18 May 2018 15:06:30	U	SMPL
J2999-01 - 1	FK-BPM-01-F4-G4-F4A-G4A	979	0.1968	-7471B	PPB		18 May 2018 15:08:46	U	SMPL
CCV87 - 1	CCV87	18640	4.9970	-7471B	PPB		18 May 2018 15:11:02	U	SMPL
CCB87 - 1	CCB87	-82	-0.0915	-7471B	PPB		18 May 2018 15:13:18	U	SMPL
J3000-01 - 1	FK-BPM-01-G3-H3-G3A-H3A	1064	0.2199	-7471B	PPB		18 May 2018 15:15:35	U	SMPL
J3000-01DUP - 1	FK-BPM-01-G3-H3-G3A-H3ADUP	1108	0.2319	-7471B	PPB		18 May 2018 15:17:51	U	SMPL
J3000-01MS - 1	FK-BPM-01-G3-H3-G3A-H3AMS	15525	4.1504	-7471B	PPB		18 May 2018 15:20:08	U	SMPL
J3000-01MSD - 1	FK-BPM-01-G3-H3-G3A-H3AMSD	15663	4.1879	-7471B	PPB		18 May 2018 15:22:24	U	SMPL
PB109415BL - 1	PBS	-86	-0.0926	-7471B	PPB		18 May 2018 15:24:41	U	SMPL
PB109415BS - 1	LCSS	14325	3.8242	-7471B	PPB		18 May 2018 15:26:58	U	SMPL
J3001-01 - 1	TP-1-STERN-AVE	25918	6.9751	-7471B	PPB		18 May 2018 15:29:15	U	SMPL
J3001-05 - 1	TP-2-STERN-AVE	49977	13.5142	-7471B	PPB		18 May 2018 15:31:33	U	SMPL
J3001-09 - 1	TP-3-STERN-AVE	36444	9.8360	-7471B	PPB		18 May 2018 15:33:51	U	SMPL
J3001-13 - 1	TP-4-STERN-AVE	74938	20.2984	-7471B	PPB		18 May 2018 15:36:10	U	SMPL
CCV88 - 1	CCV88	18685	5.0092	-7471B	PPB		18 May 2018 15:38:28	U	SMPL
CCB88 - 1	CCB88	96	-0.0432	-7471B	PPB		18 May 2018 15:43:30	U	SMPL
J3001-17 - 1	TP-5-STERN-AVE	29287	7.8908	-7471B	PPB		18 May 2018 15:45:46	U	SMPL
J3001-21 - 1	TP-6-STERN-AVE	40490	10.9357	-7471B	PPB		18 May 2018 15:48:02	U	SMPL
J3004-01 - 1	TP-10-WC-1	471	0.0588	-7471B	PPB		18 May 2018 15:50:21	U	SMPL
J3004-03 - 1	TP-10-WC-2	865	0.1659	-7471B	PPB		18 May 2018 15:52:40	U	SMPL
J3004-05 - 1	TP-10-WC-3	753	0.1354	-7471B	PPB		18 May 2018 15:54:56	U	SMPL
J3005-01 - 1	TPA-S	2940	0.7298	-7471B	PPB		18 May 2018 15:57:13	U	SMPL
J3005-03 - 1	TPA-D	830	0.1563	-7471B	PPB		18 May 2018 15:59:30	U	SMPL
J3006-01 - 1	BYR-1-E(0-1)	5812	1.5104	-7471B	PPB		18 May 2018 16:01:47	U	SMPL
J3006-02 - 1	BYR-1-E(4-6)	5484	1.4213	-7471B	PPB		18 May 2018 16:04:05	U	SMPL
J3006-03 - 1	BYR-1-E(7-8)	5565	1.4433	-7471B	PPB		18 May 2018 16:06:21	U	SMPL
CCV89 - 1	CCV89	19660	5.2742	-7471B	PPB		18 May 2018 16:08:37	U	SMPL
CCB89 - 1	CCB89	-68	-0.0877	-7471B	PPB		18 May 2018 16:10:54	U	SMPL
J3006-03DUP - 1	BYR-1-E(7-8)DUP	5825	1.5140	-7471B	PPB		18 May 2018 16:13:10	U	SMPL
J3006-03MS - 1	BYR-1-E(7-8)MS	20787	5.5805	-7471B	PPB		18 May 2018 16:15:27	U	SMPL
J3006-03MSD - 1	BYR-1-E(7-8)MSD	22397	6.0181	-7471B	PPB		18 May 2018 16:17:43	U	SMPL
J3000-01LX5 - 1		36	-0.0595	-7471B	PPB		18 May 2018 16:20:00	U	SMPL
J3000-01A - 1		16849	4.5102	-7471B	PPB		18 May 2018 16:22:18	U	SMPL
J3006-03LX5 - 1		1146	0.2422	-7471B	PPB		18 May 2018 16:31:22	U	SMPL
J3006-03A - 1		21310	5.7227	-7471B	PPB		18 May 2018 16:33:39	U	SMPL
J3001-05DLX2 - 1		24285	6.5313	-7471B	PPB		18 May 2018 16:35:56	U	SMPL
J3001-13DLX5 - 1		13998	3.7353	-7471B	PPB		18 May 2018 16:38:15	U	SMPL
J3001-21DLX2 - 1		19735	5.2946	-7471B	PPB		18 May 2018 16:40:32	U	SMPL
CCV90 - 1	CCV90	20276	5.4416	-7471B	PPB		18 May 2018 16:42:49	U	SMPL
CCB90 - 1	CCB90	-21	-0.0750	-7471B	PPB		18 May 2018 16:45:06	U	SMPL