


ISM02.4

LB 95467
INSTRUMENT ID: CV1

Linear

 μ Abs.:

41242



A= 0.0000e+000

B= 2.4160e-004

C= 1.3205e-002

Rho= 0.9996772

Accept=Accepted

slope
 $y = 1.3205e-002 x$

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	-0.002	-0.002	-65	0.000	-65					-
0.2	0.200	0.235	0.035	917	0.0 %	917					18
2.5	2.500	2.543	0.043	10469	0.0 %	10469					2
5.0	5.000	4.819	-0.181	19893	0.0 %	19893					-4
7.5	7.500	7.629	0.129	31521	0.0 %	31521					2
10.0	10.000	9.977	-0.023	41242	0.0 %	41242					0

ns
5/18/18

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.0	-65	-		0 ISM02.4	PPB	5/18/2018 17:03	Std
	0.2 S0.2	917	-		0.2 ISM02.4	PPB	5/18/2018 17:05	Std
	2.5 S2.5	10469	-		2.5 ISM02.4	PPB	5/18/2018 17:07	Std
	5 S5	19893	-		5 ISM02.4	PPB	5/18/2018 17:09	Std
	7.5 S7.5	31521	-		7.5 ISM02.4	PPB	5/18/2018 17:12	Std
	10 S10.0	41242	-		10 ISM02.4	PPB	5/18/2018 17:14	Std
ICV	ICV	15734	3.8145 -		ISM02.4	PPB	5/18/2018 17:17	SMPL
ICB	ICB	-207	-0.0368 -		ISM02.4	PPB	5/18/2018 17:19	SMPL
CCV091	CCV091	19782	4.7925 -		ISM02.4	PPB	5/18/2018 17:22	SMPL
CCB091	CCB091	-205	-0.0363 -		ISM02.4	PPB	5/18/2018 17:24	SMPL
PB109434BL	PBW003	0	0.0132 -		ISM02.4	PPB	5/18/2018 17:26	SMPL
J2989-09	MC0AC3	-83	-0.0068 -		ISM02.4	PPB	5/18/2018 17:28	SMPL
CCV092	CCV092	19989	4.8425 -		ISM02.4	PPB	5/18/2018 17:31	SMPL
CCB092	CCB092	-217	-0.0392 -		ISM02.4	PPB	5/18/2018 17:33	SMPL