



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB95472

Slope : 97.8

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.1	4.01	05/18/2018	17:00
2	CAL2	1	Water			21.1	6.99	05/18/2018	17:01
3	CAL3	1	Water			21.1	10.02	05/18/2018	17:02
4	ICV	1	Water			21.1	7.00	05/18/2018	17:03
5	CCV1	1	Water			21.2	2.01	05/18/2018	17:04
6	J3010-24	1	Water			19.1	6.18	05/18/2018	17:05
7	J3013-04	1	Water			18.5	7.45	05/18/2018	17:06
8	J3013-04DUP	1	Water			18.6	7.46	05/18/2018	17:09
9	CCV2	1	Water			21.2	12.02	05/18/2018	17:10

LB95472

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosivity j3010

WorkList ID : 112313

Date : 5/18/2018 4:37:33 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
05/18/2018	Water	J3010-24~A	Corrosivity	Cool 4 deg C	PSEG01	I31	OILY-WATER	05/18/2018	9040C
05/18/2018	Water	J3013-04~B	Corrosivity	Cool 4 deg C	PSEG01	I23	FP-WATER	05/18/2018	9040C

Date/Time 05-18-18 4:40 PM
Received by: JN
Relinquished by:

Date/Time 05-18-18 6:00 PM
Received by: CH
Relinquished by: JN