


7470A

LB 95498
INSTRUMENT ID: CV4

Linear

 μ Abs.:

40040



A= 0.0000e+000

B= 2.5121e-004

C= 1.0750e-002

Rho= 0.9999237

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.016	0.016	20	0.000	20				
0.2	0.200	0.213	0.013	804	0.0 %	804				
2.5	2.500	2.518	0.018	9979	0.0 %	9979				
5.0	5.000	4.955	-0.045	19680	0.0 %	19680				
7.5	7.500	7.430	-0.070	29534	0.0 %	29534				
10.0	10.000	10.069	0.069	40040	0.0 %	40040				

LB95498 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 21 May 2018 09:10:02

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	S'0	20	-	0.0000	7470A	PPB	21 May 2018 09:16:34	S	Std
0.2 - 1	S0.2	804	-	0.2000	7470A	PPB	21 May 2018 09:18:42	S	Std
2.5 - 1	S2.5	9979	-	2.5000	7470A	PPB	21 May 2018 09:20:50	S	Std
5.0 - 1	S5	19680	-	5.0000	7470A	PPB	21 May 2018 09:22:58	S	Std
7.5 - 1	S7.5	29534	-	7.5000	7470A	PPB	21 May 2018 09:25:06	S	Std
10.0 - 1	S10	40040	-	10.0000	7470A	PPB	21 May 2018 09:27:14	S	Std
ICV56 - 1	ICV56	16285	4.1017	-	7470A	PPB	21 May 2018 09:29:46	U	SMPL
ICB56 - 1	ICB56	-163	-0.0302	-	7470A	PPB	21 May 2018 09:31:52	U	SMPL
CCV97 - 1	CCV97	19874	5.0033	-	7470A	PPB	21 May 2018 09:33:59	U	SMPL
CCB97 - 1	CCB97	-122	-0.0199	-	7470A	PPB	21 May 2018 09:36:06	U	SMPL
CRA - 1	CRA	852	0.2248	-	7470A	PPB	21 May 2018 09:38:13	U	SMPL
HighStd - 1	HighStd	40484	10.1808	-	7470A	PPB	21 May 2018 09:40:19	U	SMPL
ChkStd - 1	ChkStd	26997	6.7927	-	7470A	PPB	21 May 2018 09:42:25	U	SMPL
PB109457BL - 1	PBW	-207	-0.0413	-	7470A	PPB	21 May 2018 09:44:34	U	SMPL
PB109457BS - 1	LCSW	16079	4.0500	-	7470A	PPB	21 May 2018 09:46:43	U	SMPL
PB109457TB - 1	PB109412TB	-135	-0.0232	-	7470A	PPB	21 May 2018 09:48:50	U	SMPL
J2673-02 - 1	JC-03-051718-A	-17	0.0065	-	7470A	PPB	21 May 2018 09:50:59	U	SMPL
J2673-04 - 1	JC-03-051718-B	-20	0.0057	-	7470A	PPB	21 May 2018 09:53:07	U	SMPL
J2673-06 - 1	JC-03-051718-C	-3	0.0100	-	7470A	PPB	21 May 2018 09:55:15	U	SMPL
J2673-08 - 1	JC-03-051718-D	10	0.0133	-	7470A	PPB	21 May 2018 09:57:22	U	SMPL
CCV98 - 1	CCV98	19987	5.0317	-	7470A	PPB	21 May 2018 09:59:29	U	SMPL
CCB98 - 1	CCB98	-107	-0.0161	-	7470A	PPB	21 May 2018 10:01:36	U	SMPL
J2673-10 - 1	JC-03-051718-E	41	0.0210	-	7470A	PPB	21 May 2018 10:03:43	U	SMPL
J2673-12 - 1	JC-03-051718-F	58	0.0253	-	7470A	PPB	21 May 2018 10:05:50	U	SMPL
J2673-14 - 1	JC-03-051718-G	40	0.0208	-	7470A	PPB	21 May 2018 10:07:56	U	SMPL
J2673-16 - 1	JC-03-051718-H	-10	0.0082	-	7470A	PPB	21 May 2018 10:10:03	U	SMPL
J2677-06 - 1	OK-02-051718	19	0.0155	-	7470A	PPB	21 May 2018 10:12:11	U	SMPL
J2677-06DUP - 1	OK-02-051718DUP	16	0.0148	-	7470A	PPB	21 May 2018 10:14:18	U	SMPL
J2677-06MS - 1	OK-02-051718MS	14877	3.7480	-	7470A	PPB	21 May 2018 10:16:26	U	SMPL
J2677-06MSD - 1	OK-02-051718MSD	14419	3.6330	-	7470A	PPB	21 May 2018 10:18:34	U	SMPL
J2971-02 - 1	TP6-WC-1	-103	-0.0151	-	7470A	PPB	21 May 2018 10:20:43	U	SMPL
J2971-04 - 1	TP6-WC-2	-7	0.0090	-	7470A	PPB	21 May 2018 10:22:50	U	SMPL
CCV99 - 1	CCV99	19879	5.0046	-	7470A	PPB	21 May 2018 10:24:56	U	SMPL
CCB99 - 1	CCB99	-104	-0.0154	-	7470A	PPB	21 May 2018 10:27:03	U	SMPL
J2971-06 - 1	TP6-WC-3	1	0.0110	-	7470A	PPB	21 May 2018 10:29:10	U	SMPL
J2971-08 - 1	TP6-WC-4	-19	0.0060	-	7470A	PPB	21 May 2018 10:31:17	U	SMPL
J2971-10 - 1	TP6-WC-5	17	0.0150	-	7470A	PPB	21 May 2018 10:33:24	U	SMPL
J3001-04 - 1	TP-1-STERN-AVE	-37	0.0015	-	7470A	PPB	21 May 2018 10:35:31	U	SMPL
J3001-08 - 1	TP-2-STERN-AVE	1	0.0110	-	7470A	PPB	21 May 2018 10:37:38	U	SMPL
J3001-12 - 1	TP-3-STERN-AVE	4	0.0118	-	7470A	PPB	21 May 2018 10:39:46	U	SMPL
J3001-16 - 1	TP-4-STERN-AVE	-6	0.0092	-	7470A	PPB	21 May 2018 10:41:53	U	SMPL
J3001-20 - 1	TP-5-STERN-AVE	55	0.0246	-	7470A	PPB	21 May 2018 10:44:01	U	SMPL
J3001-24 - 1	TP-6-STERN-AVE	-56	-0.0033	-	7470A	PPB	21 May 2018 10:46:10	U	SMPL
CCV01 - 1	CCV01	19756	4.9737	-	7470A	PPB	21 May 2018 10:48:17	U	SMPL
CCB01 - 1	CCB01	-101	-0.0146	-	7470A	PPB	21 May 2018 10:50:24	U	SMPL
J2677-06LX5 - 1		19	0.0155	-	7470A	PPB	21 May 2018 10:52:32	U	SMPL
J2677-06A - 1		15447	3.8912	-	7470A	PPB	21 May 2018 10:54:39	U	SMPL
CCV02 - 1	CCV02	20057	5.0493	-	7470A	PPB	21 May 2018 10:56:46	U	SMPL
CCB02 - 1	CCB02	-132	-0.0224	-	7470A	PPB	21 May 2018 10:58:53	U	SMPL