



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: KETANKUMAR

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB95503

Ambient Barometric Pressure (mmHg): 760

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	27.0	80.6	80.6	05/21/2018	09:10
2	J3010-24		1	>100.0	>212.0	>212.0	05/21/2018	09:40
3	J3010-24DUP		1	>100.0	>212.0	>212.0	05/21/2018	10:10
4	J3013-04		1	>100.0	>212.0	>212.0	05/21/2018	10:42

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.33

LB95503

WORKLIST(Hardcopy Internal Chain)

WorkList Name : FP-052118

WorkList ID : 112373

Date : 5/21/2018 8:17:54 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
05/23/2018	Water	J3010-24	Flash Point	Cool 4 deg C	PSEG01	I31	OILY-WATER	05/18/2018	1010A
05/23/2018	Water	J3013-04	Flash Point	Cool 4 deg C	PSEG01	I23	FP-WATER	05/18/2018	1010A

Date/Time 05/21/18 8:30 AM
 Received by: KS
 Relinquished by: CP

Date/Time 05/21/18 10:45 AM
 Received by: CP
 Relinquished by: KS