


7471B

LB95513
INSTRUMENT ID: CVL

Linear

 μ Abs.:

47864



A= 0.0000e+000

B= 2.1061e-004

C= -7.5222e-003

Rho= 0.9997527

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.010	-0.010	-10	0.000	-10				
0.2	0.200	0.228	0.028	1118	0.0 %	1118				
2.5	2.500	2.471	-0.029	11769	0.0 %	11769				
5.0	5.000	5.094	0.094	24224	0.0 %	24224				
7.5	7.500	7.343	-0.157	34904	0.0 %	34904				
10.0	10.000	10.073	0.073	47864	0.0 %	47864				

LB95513 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 21 May 2018 14:57:19

Sample ID	Extended ID	μ Abs.	Conc.	Std ConcMethod	Units	Date	Type	Type
0.0 - 1	50	-10	-	0.00007471B	PPB	21 May 2018 14:59:41	S	Std
0.2 - 1	50.2	1118	-	0.20007471B	PPB	21 May 2018 15:01:59	S	Std
2.5 - 1	52.5	11769	-	2.50007471B	PPB	21 May 2018 15:04:17	S	Std
5.0 - 1	55	24224	-	5.00007471B	PPB	21 May 2018 15:06:35	S	Std
7.5 - 1	57.5	34904	-	7.50007471B	PPB	21 May 2018 15:08:52	S	Std
10.0 - 1	59	47864	-	10.00007471B	PPB	21 May 2018 15:11:08	S	Std
ICV57 - 1	ICV57	19073	4.0094	-7471B	PPB	21 May 2018 15:13:56	U	SMPL
ICB57 - 1	ICB57	-158	-0.0408	-7471B	PPB	21 May 2018 15:16:11	U	SMPL
CCV03 - 1	CCV03	22600	4.7522	-7471B	PPB	21 May 2018 15:18:26	U	SMPL
CCB03 - 1	CCB03	-183	-0.0461	-7471B	PPB	21 May 2018 15:20:42	U	SMPL
CRA - 1	CRA	1097	0.2235	-7471B	PPB	21 May 2018 15:22:58	U	SMPL
HighStd - 1	HighStd	48773	10.2644	-7471B	PPB	21 May 2018 15:25:14	U	SMPL
ChkStd - 1	ChkStd	31253	6.5745	-7471B	PPB	21 May 2018 15:27:30	U	SMPL
PB109461BL - 1	PBS	-114	-0.0315	-7471B	PPB	21 May 2018 15:29:48	U	SMPL
PB109461BS - 1	LCSS	17084	3.5905	-7471B	PPB	21 May 2018 15:32:06	U	SMPL
J2676-21 - 1	CL-01-051818-A	1591	0.3276	-7471B	PPB	21 May 2018 15:34:23	U	SMPL
J2676-23 - 1	CL-01-051818-B	2163	0.4480	-7471B	PPB	21 May 2018 15:36:40	U	SMPL
J2676-25 - 1	CL-01-051818-C	2186	0.4529	-7471B	PPB	21 May 2018 15:38:57	U	SMPL
J2676-27 - 1	CL-01-051818-D	1783	0.3680	-7471B	PPB	21 May 2018 15:41:15	U	SMPL
J2676-29 - 1	CL-01-051818-E	1544	0.3177	-7471B	PPB	21 May 2018 15:43:31	U	SMPL
CCV04 - 1	CCV04	23085	4.8543	-7471B	PPB	21 May 2018 15:45:47	U	SMPL
CCB04 - 1	CCB04	-199	-0.0494	-7471B	PPB	21 May 2018 15:48:03	U	SMPL
J3010-01 - 1	132-2-PHASE-1	419	0.0807	-7471B	PPB	21 May 2018 15:50:19	U	SMPL
J3010-04 - 1	132-2-PHASE-2	483	0.0942	-7471B	PPB	21 May 2018 15:52:35	U	SMPL
J3010-07 - 1	132-2-PHASE-3	211	0.0369	-7471B	PPB	21 May 2018 15:54:51	U	SMPL
J3010-10 - 1	SOIL-PHASE-1	2285	0.4737	-7471B	PPB	21 May 2018 15:57:07	U	SMPL
J3010-14 - 1	SOIL-PHASE-2	2005	0.4147	-7471B	PPB	21 May 2018 15:59:24	U	SMPL
J3010-14DUP - 1	SOIL-PHASE-2DUP	2058	0.4259	-7471B	PPB	21 May 2018 16:01:41	U	SMPL
J3010-14MS - 1	SOIL-PHASE-2MS	19594	4.1191	-7471B	PPB	21 May 2018 16:03:58	U	SMPL
J3010-14MSD - 1	SOIL-PHASE-2MSD	19577	4.1155	-7471B	PPB	21 May 2018 16:06:15	U	SMPL
J3010-18 - 1	SOIL-PHASE-3	3717	0.7753	-7471B	PPB	21 May 2018 16:08:33	U	SMPL
J3011-01 - 1	MS-BLDG-FOUNDATION	281	0.0517	-7471B	PPB	21 May 2018 16:10:49	U	SMPL
CCV05 - 1	CCV05	23074	4.8520	-7471B	PPB	21 May 2018 16:13:06	U	SMPL
CCB05 - 1	CCB05	-176	-0.0446	-7471B	PPB	21 May 2018 16:15:22	U	SMPL
J3013-01 - 1	FP-SOIL	5761	1.2058	-7471B	PPB	21 May 2018 16:17:38	U	SMPL
J3019-01 - 1	WJTS-051818-01	8890	1.8648	-7471B	PPB	21 May 2018 16:19:54	U	SMPL
J3019-02 - 1	WJTS-051818-02	230	0.0409	-7471B	PPB	21 May 2018 16:22:11	U	SMPL
J3019-03 - 1	WJTS-051818-03	853	0.1721	-7471B	PPB	21 May 2018 16:24:27	U	SMPL
PB109462BL - 1	PBS	110	0.0156	-7471B	PPB	21 May 2018 16:26:43	U	SMPL
PB109462BS - 1	LCSS	17246	3.6246	-7471B	PPB	21 May 2018 16:29:00	U	SMPL
J3023-01 - 1	FOX-1-E(0-1)	17261	3.6277	-7471B	PPB	21 May 2018 16:31:18	U	SMPL
J3023-02 - 1	FOX-1-E(4-5)	1868	0.3859	-7471B	PPB	21 May 2018 16:33:35	U	SMPL
J3023-03 - 1	FOX-1-E(9-10)	3242	0.6753	-7471B	PPB	21 May 2018 16:35:53	U	SMPL
J3023-03DUP - 1	FOX-1-E(9-10)DUP	3436	0.7161	-7471B	PPB	21 May 2018 16:38:10	U	SMPL
CCV06 - 1	CCV06	22954	4.8267	-7471B	PPB	21 May 2018 16:40:37	U	SMPL
CCB06 - 1	CCB06	-200	-0.0496	-7471B	PPB	21 May 2018 16:42:52	U	SMPL
J3023-03MS - 1	FOX-1-E(9-10)MS	20609	4.3329	-7471B	PPB	21 May 2018 16:45:09	U	SMPL
J3023-03MSD - 1	FOX-1-E(9-10)MSD	20488	4.3074	-7471B	PPB	21 May 2018 16:47:26	U	SMPL
J3024-01 - 1	EPE-MX-3(75-77)	1026	0.2086	-7471B	PPB	21 May 2018 16:49:43	U	SMPL
J3024-02 - 1	EPE-MX-3(85-87)	828	0.1669	-7471B	PPB	21 May 2018 16:52:00	U	SMPL
J3024-03 - 1	EPE-F-3A(0-5)	7409	1.5529	-7471B	PPB	21 May 2018 16:54:17	U	SMPL
J3024-04 - 1	EPE-F-3A(5-10)	1857	0.3836	-7471B	PPB	21 May 2018 16:56:34	U	SMPL
J3024-05 - 1	EPE-MX-3(125-127)	1336	0.2738	-7471B	PPB	21 May 2018 16:58:51	U	SMPL
J3024-06 - 1	EPE-MX-3(135-137)	905	0.1831	-7471B	PPB	21 May 2018 17:01:08	U	SMPL
J3024-07 - 1	EPE-MX-3(145-147)	1398	0.2869	-7471B	PPB	21 May 2018 17:03:25	U	SMPL
CCV07 - 1	CCV07	23342	4.9084	-7471B	PPB	21 May 2018 17:05:42	U	SMPL
CCB07 - 1	CCB07	-172	-0.0437	-7471B	PPB	21 May 2018 17:07:59	U	SMPL
J3010-14LX5 - 1		431	0.0832	-7471B	PPB	21 May 2018 17:10:16	U	SMPL
J3010-14A - 1		20683	4.3484	-7471B	PPB	21 May 2018 17:12:33	U	SMPL
J3023-03LX5 - 1		542	0.1066	-7471B	PPB	21 May 2018 17:14:50	U	SMPL
J3023-03A - 1		22019	4.6298	-7471B	PPB	21 May 2018 17:17:07	U	SMPL
CCV08 - 1	CCV08	22904	4.8162	-7471B	PPB	21 May 2018 17:19:24	U	SMPL
CCB08 - 1	CCB08	-130	-0.0349	-7471B	PPB	21 May 2018 17:21:41	U	SMPL
						21 May 2018 17:23:58	U	SMPL
						21 May 2018 17:26:36	U	SMPL