



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Sweta

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB95520

Slope : 99.3

BalanceID: WC SC-4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.5	4.01	05/21/2018	16:45
2	CAL2	1	Water			21.2	7.00	05/21/2018	16:47
3	CAL3	1	Water			21.3	10.03	05/21/2018	16:49
4	ICV	1	Water			21.4	7.00	05/21/2018	16:51
5	CCV1	1	Water			21.5	2.00	05/21/2018	16:53
6	J2675-10	1	Solid	20.02	20	21.2	6.88	05/21/2018	16:55
7	J2675-10DUP	1	Solid	20.03	20	21.3	6.89	05/21/2018	16:57
8	J2675-12	1	Solid	20.02	20	21.5	9.09	05/21/2018	16:59
9	J2679-20	1	Solid	20.04	20	22.5	7.47	05/21/2018	17:01
10	J2679-22	1	Solid	20.02	20	22.6	7.32	05/21/2018	17:03
11	J2679-24	1	Solid	20.05	20	21.9	8.90	05/21/2018	17:05
12	J3055-04	1	Solid	20.03	20	22.2	5.52	05/21/2018	17:07
13	J3055-08	1	Solid	20.02	20	21.4	6.11	05/21/2018	17:09
14	J3055-12	1	Solid	20.03	20	22.1	5.74	05/21/2018	17:11
15	J3056-04	1	Solid	20.05	20	22.1	7.58	05/21/2018	17:13
16	CCV2	1	Water			21.5	12.00	05/21/2018	17:15
17	J3056-08	1	Solid	20.02	20	21.6	8.98	05/21/2018	17:17
18	J3058-03	1	Solid	20.03	20	21.8	8.27	05/21/2018	17:19
19	CCV3	1	Water			21.3	2.00	05/21/2018	17:21

WORKLIST(Hardcopy Internal Chain)

LB 95520

WorkList Name : corrosivity052118

WorkList ID : 112403

Date : 5/21/2018 3:23:29 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
05/21/2018	Solid	J2675-10	Corrosivity	Cool 4 deg C	PSEG05		HR-05-052118-A	05/21/2018	9045C
05/21/2018	Solid	J2675-12	Corrosivity	Cool 4 deg C	PSEG05		HR-05-052118-B	05/21/2018	9045C
05/21/2018	Solid	J2679-20	Corrosivity	Cool 4 deg C	PSEG05		EO-01-052118-A	05/21/2018	9045C
05/21/2018	Solid	J2679-22	Corrosivity	Cool 4 deg C	PSEG05		EO-01-052118-B	05/21/2018	9045C
05/21/2018	Solid	J2679-24	Corrosivity	Cool 4 deg C	PSEG05		EO-01-052118-C	05/21/2018	9045C
05/21/2018	Solid	J3055-04	Corrosivity	Cool 4 deg C	PSEG01	I23	PHASE-1	05/21/2018	9045C
05/21/2018	Solid	J3055-08	Corrosivity	Cool 4 deg C	PSEG01	I23	PHASE-2	05/21/2018	9045C
05/21/2018	Solid	J3055-12	Corrosivity	Cool 4 deg C	PSEG01	I23	PHASE-3	05/21/2018	9045C
05/21/2018	Solid	J3056-04	Corrosivity	Cool 4 deg C	PSEG01	I23	TP-1	05/21/2018	9045C
05/21/2018	Solid	J3056-08	Corrosivity	Cool 4 deg C	PSEG01	I23	TP-2	05/21/2018	9045C
05/21/2018	Solid	J3058-03	Corrosivity	Cool 4 deg C	PSEG01	I62	VNJ-229	05/21/2018	9045C

Date/Time 05/21/18 4:30 PM

Received by: SP

Relinquished by: CP

Date/Time 05/21/18 5:30 PM

Received by: CP

Relinquished by: SP