


7470A

LB95526
INSTRUMENT ID: CV1

Linear

 μ Abs.:

39039



A= 0.0000e+000

B= 2.5750e-004

C= -2.6403e-002

Rho= 0.9998991

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.018	-0.018	31	0.000	31				
0.2	0.200	0.229	0.029	992	0.0 %	992				
2.5	2.500	2.461	-0.039	9658	0.0 %	9658				
5.0	5.000	5.082	0.082	19839	0.0 %	19839				
7.5	7.500	7.421	-0.079	28920	0.0 %	28920				
10.0	10.000	10.026	0.026	39039	0.0 %	39039				

LB95526 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 22 May 2018 09:10:37

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	31	-	0.0000	7470A	PPB	22 May 2018 09:10:48	S	Std
0.2 - 1	50.2	992	-	0.2000	7470A	PPB	22 May 2018 09:12:55	S	Std
2.5 - 1	52.5	9658	-	2.5000	7470A	PPB	22 May 2018 09:15:04	S	Std
5.0 - 1	55	19839	-	5.0000	7470A	PPB	22 May 2018 09:17:12	S	Std
7.5 - 1	57.5	28920	-	7.5000	7470A	PPB	22 May 2018 09:19:20	S	Std
10.0 - 1	510	39039	-	10.0000	7470A	PPB	22 May 2018 09:21:28	S	Std
ICV59 - 1	ICV59	15610	3.9932	-	7470A	PPB	22 May 2018 09:26:35	U	SMPL
ICB59 - 1	ICB59	92	-0.0027	-	7470A	PPB	22 May 2018 09:28:41	U	SMPL
CCV16 - 1	CCV16	19182	4.9130	-	7470A	PPB	22 May 2018 09:30:48	U	SMPL
CCB16 - 1	CCB16	97	-0.0014	-	7470A	PPB	22 May 2018 09:32:55	U	SMPL
CRA - 1	CRA	951	0.2185	-	7470A	PPB	22 May 2018 09:35:02	U	SMPL
HighStd - 1	HighStd	39430	10.1269	-	7470A	PPB	22 May 2018 09:37:09	U	SMPL
ChkStd - 1	ChkStd	26536	6.8066	-	7470A	PPB	22 May 2018 09:39:16	U	SMPL
PB109518BL - 1	PBW	-179	-0.0725	-	7470A	PPB	22 May 2018 09:41:25	U	SMPL
PB109518BS - 1	LCSW	15295	3.9121	-	7470A	PPB	22 May 2018 09:43:33	U	SMPL
PB109518TB - 1	PB109455TB	-53	-0.0401	-	7470A	PPB	22 May 2018 09:45:41	U	SMPL
J2676-22 - 1	CL-01-051818-A	40	-0.0161	-	7470A	PPB	22 May 2018 09:47:49	U	SMPL
J2676-24 - 1	CL-01-051818-B	63	-0.0102	-	7470A	PPB	22 May 2018 09:49:57	U	SMPL
J2676-26 - 1	CL-01-051818-C	51	-0.0133	-	7470A	PPB	22 May 2018 09:52:05	U	SMPL
J2676-28 - 1	CL-01-051818-D	30	-0.0187	-	7470A	PPB	22 May 2018 09:54:11	U	SMPL
CCV17 - 1	CCV17	19407	4.9709	-	7470A	PPB	22 May 2018 09:56:18	U	SMPL
CCB17 - 1	CCB17	166	0.0163	-	7470A	PPB	22 May 2018 09:58:25	U	SMPL
J2676-30 - 1	CL-01-051818-E	29	-0.0189	-	7470A	PPB	22 May 2018 10:00:32	U	SMPL
J3004-02 - 1	TP-10-WC-1	22	-0.0207	-	7470A	PPB	22 May 2018 10:02:39	U	SMPL
J3004-04 - 1	TP-10-WC-2	15	-0.0225	-	7470A	PPB	22 May 2018 10:04:45	U	SMPL
J3004-06 - 1	TP-10-WC-3	12	-0.0233	-	7470A	PPB	22 May 2018 10:06:52	U	SMPL
J3005-02 - 1	TPA-S	36	-0.0171	-	7470A	PPB	22 May 2018 10:09:00	U	SMPL
J3005-04 - 1	TPA-D	30	-0.0187	-	7470A	PPB	22 May 2018 10:11:08	U	SMPL
J3010-03 - 1	132-2-PHASE-1	38	-0.0166	-	7470A	PPB	22 May 2018 10:13:15	U	SMPL
J3010-06 - 1	132-2-PHASE-2	39	-0.0164	-	7470A	PPB	22 May 2018 10:15:23	U	SMPL
J3010-09 - 1	132-2-PHASE-3	20	-0.0213	-	7470A	PPB	22 May 2018 10:17:31	U	SMPL
J3010-13 - 1	SOIL-PHASE-1	4	-0.0254	-	7470A	PPB	22 May 2018 10:19:38	U	SMPL
CCV18 - 1	CCV18	19463	4.9853	-	7470A	PPB	22 May 2018 10:21:45	U	SMPL
CCB18 - 1	CCB18	-6	-0.0279	-	7470A	PPB	22 May 2018 10:23:51	U	SMPL
J3010-17 - 1	SOIL-PHASE-2	-25	-0.0328	-	7470A	PPB	22 May 2018 10:25:59	U	SMPL
J3010-21 - 1	SOIL-PHASE-3	11	-0.0236	-	7470A	PPB	22 May 2018 10:28:05	U	SMPL
J3013-02 - 1	FP-SOIL	12	-0.0233	-	7470A	PPB	22 May 2018 10:30:13	U	SMPL
J3013-02DUP - 1	FP-SOILDUP	24	-0.0202	-	7470A	PPB	22 May 2018 10:32:20	U	SMPL
J3013-02MS - 1	FP-SOILMS	14827	3.7916	-	7470A	PPB	22 May 2018 10:34:27	U	SMPL
J3013-02MSD - 1	FP-SOILMSD	14461	3.6973	-	7470A	PPB	22 May 2018 10:36:35	U	SMPL
PB109519BL - 1	PBW	-45	-0.0380	-	7470A	PPB	22 May 2018 10:38:43	U	SMPL
PB109519BS - 1	LCSW	15251	3.9007	-	7470A	PPB	22 May 2018 10:40:51	U	SMPL
PB109519TB - 1	PB109460TB	-137	-0.0617	-	7470A	PPB	22 May 2018 10:43:00	U	SMPL
J3010-24 - 1	OILY-WATER	100	-0.0007	-	7470A	PPB	22 May 2018 10:45:08	U	SMPL
CCV19 - 1	CCV19	19470	4.9871	-	7470A	PPB	22 May 2018 10:47:15	U	SMPL
CCB19 - 1	CCB19	-100	-0.0522	-	7470A	PPB	22 May 2018 10:49:22	U	SMPL
J3013-04 - 1	FP-WATER	94	-0.0022	-	7470A	PPB	22 May 2018 10:51:30	U	SMPL
J3013-04DUP - 1	FP-WATERDUP	42	-0.0156	-	7470A	PPB	22 May 2018 10:53:37	U	SMPL
J3013-04MS - 1	FP-WATERMS	14271	3.6484	-	7470A	PPB	22 May 2018 10:55:44	U	SMPL
J3013-04MSD - 1	FP-WATERMSD	14604	3.7341	-	7470A	PPB	22 May 2018 10:57:52	U	SMPL
PB109520BL - 1	PBW	-40	-0.0367	-	7470A	PPB	22 May 2018 10:59:59	U	SMPL
PB109520BS - 1	LCSW	14406	3.6832	-	7470A	PPB	22 May 2018 11:02:07	U	SMPL
J3010-23 - 1	OILY-WATER	154	0.0133	-	7470A	PPB	22 May 2018 11:04:15	U	SMPL
J3013-03 - 1	FP-WATER	20463	5.2428	-	7470A	PPB	22 May 2018 11:06:23	U	SMPL
J3013-03DUP - 1	FP-WATERDUP	18105	4.6356	-	7470A	PPB	22 May 2018 11:08:32	U	SMPL
J3013-03MS - 1	FP-WATERMS	32780	8.4145	-	7470A	PPB	22 May 2018 11:10:41	U	SMPL
CCV20 - 1	CCV20	19727	5.0533	-	7470A	PPB	22 May 2018 11:12:49	U	SMPL
CCB20 - 1	CCB20	-79	-0.0467	-	7470A	PPB	22 May 2018 11:14:58	U	SMPL
J3013-03MSD - 1	FP-WATERMSD	32785	8.4158	-	7470A	PPB	22 May 2018 11:17:06	U	SMPL
J3013-02LX5 - 1		-251	-0.0910	-	7470A	PPB	22 May 2018 11:19:13	U	SMPL
J3013-02A - 1		14776	3.7784	-	7470A	PPB	22 May 2018 11:21:22	U	SMPL
J3013-04LX5 - 1		-118	-0.0568	-	7470A	PPB	22 May 2018 11:23:29	U	SMPL
J3013-04A - 1		14750	3.7717	-	7470A	PPB	22 May 2018 11:25:37	U	SMPL
J3013-03LX5 - 1		3640	0.9109	-	7470A	PPB	22 May 2018 11:27:44	U	SMPL
J3013-03A - 1		35679	9.1610	-	7470A	PPB	22 May 2018 11:29:52	U	SMPL
CCV21 - 1	CCV21	19145	4.9034	-	7470A	PPB	22 May 2018 11:32:00	U	SMPL
CCB21 - 1	CCB21	142	0.0102	-	7470A	PPB	22 May 2018 11:34:10	U	SMPL