



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Sweta

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB95539

Slope : 98.7

BalanceID: WC SC-4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.6	4.01	05/22/2018	15:55
2	CAL2	1	Water			21.3	7.00	05/22/2018	15:57
3	CAL3	1	Water			21.6	10.03	05/22/2018	15:59
4	ICV	1	Water			21.4	7.00	05/22/2018	16:01
5	CCV1	1	Water			21.5	2.00	05/22/2018	16:03
6	J2669-10	1	Solid	20.02	20	22.3	8.43	05/22/2018	16:05
7	J2669-10DUP	1	Solid	20.04	20	22.4	8.44	05/22/2018	16:07
8	J3060-02	1	Solid	20.03	20	21.9	7.57	05/22/2018	16:09
9	J3060-04	1	Solid	20.02	20	21.7	6.60	05/22/2018	16:11
10	J3060-06	1	Solid	20.05	20	21.8	7.25	05/22/2018	16:13
11	J3071-01	1	Solid	20.02	20	22.3	6.64	05/22/2018	16:15
12	J3075-04	1	Solid	20.03	20	21.7	5.93	05/22/2018	16:17
13	CCV2	1	Water			21.5	12.00	05/22/2018	16:19

LB95539

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosivity052218

WorkList ID : 112453

Date : 5/22/2018 3:28:31 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
05/22/2018	Solid	J2669-10	Corrosivity	Cool 4 deg C	PSEG05		SU-02-052218	05/22/2018	9045C
05/21/2018	Solid	J3060-02	Corrosivity	Cool 4 deg C	PSEG01	I32	TP-3-WC-1	05/21/2018	9045C
05/21/2018	Solid	J3060-04	Corrosivity	Cool 4 deg C	PSEG01	I32	TP-3-WC-2	05/21/2018	9045C
05/21/2018	Solid	J3060-06	Corrosivity	Cool 4 deg C	PSEG01	I32	TP-3-WC-3	05/21/2018	9045C
05/22/2018	Solid	J3071-01	Corrosivity	Cool 4 deg C	SEAC01	L52	SOIL-CUTTINGS	05/22/2018	9045C
05/22/2018	Solid	J3075-04	Corrosivity	Cool 4 deg C	PSEG01	H32	RO-4-1	05/22/2018	9045C

Date/Time 05/22/18 3:45 PM

Received by: sp

Relinquished by: CP

Date/Time 05/22/18 4:30 PM

Received by: CP

Relinquished by: SP