


7471B

LB 95540
INSTRUMENT ID: CV1

Linear

 μ Abs.:

46551



A= 0.0000e+000

B= 2.1358e-004

C= -3.2235e-002

Rho= 0.9995224

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.008	-0.008	114	0.000	114				
0.2	0.200	0.222	0.022	1191	0.0 %	1191				
2.5	2.500	2.531	0.031	12001	0.0 %	12001				
5.0	5.000	4.839	-0.161	22805	0.0 %	22805				
7.5	7.500	7.706	0.206	36230	0.0 %	36230				
10.0	10.000	9.910	-0.090	46551	0.0 %	46551				

LB95540 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 22 May 2018 16:08:21

Sample ID	Extended ID	μ Abs.	Conc.	Std ConcMethod	Units	Date	Type	Type
0.0 - 1	50	114	-	0.00007471B	PPB	22 May 2018 16:13:24	S	Std
0.2 - 1	50.2	1191	-	0.20007471B	PPB	22 May 2018 16:15:41	S	Std
2.5 - 1	52.5	12001	-	2.50007471B	PPB	22 May 2018 16:17:59	S	Std
5.0 - 1	55	22805	-	5.00007471B	PPB	22 May 2018 16:20:16	S	Std
7.5 - 1	57.5	36230	-	7.50007471B	PPB	22 May 2018 16:22:34	S	Std
10.0 - 1	50	46551	-	10.00007471B	PPB	22 May 2018 16:27:27	S	Std
ICV60 - 1	ICV60	17229	3.6476	-7471B	PPB	22 May 2018 16:50:52	U	SMPL
ICB60 - 1	ICB60	0	-0.0322	-7471B	PPB	22 May 2018 16:53:08	U	SMPL
CCV24 - 1	CCV24	23156	4.9135	-7471B	PPB	22 May 2018 16:55:24	U	SMPL
CCB24 - 1	CCB24	3	-0.0316	-7471B	PPB	22 May 2018 16:57:39	U	SMPL
CRA - 1	CRA	1192	0.2224	-7471B	PPB	22 May 2018 16:59:55	U	SMPL
HighStd - 1	HighStd	46366	9.8708	-7471B	PPB	22 May 2018 17:02:11	U	SMPL
ChkStd - 1	ChkStd	33282	7.0763	-7471B	PPB	22 May 2018 17:04:27	U	SMPL
PB109531BL - 1	PBS	-128	-0.0596	-7471B	PPB	22 May 2018 17:06:45	U	SMPL
PB109531BS - 1	LCSS	17008	3.6004	-7471B	PPB	22 May 2018 17:09:03	U	SMPL
J2675-09 - 1	HR-05-052118-A	558	0.0869	-7471B	PPB	22 May 2018 17:11:20	U	SMPL
J2675-11 - 1	HR-05-052118-B	312694	HIGH	-7471B	PPB	22 May 2018 17:13:38	X	
J2679-19 - 1	EO-01-052118-A	2434	0.4876	-7471B	PPB	22 May 2018 17:15:30	U	SMPL
J2679-21 - 1	EO-01-052118-B	505	0.0756	-7471B	PPB	22 May 2018 17:17:53	U	SMPL
J2679-23 - 1	EO-01-052118-C	852	0.1497	-7471B	PPB	22 May 2018 17:20:08	U	SMPL
CCV25 - 1	CCV25	22649	4.8052	-7471B	PPB	22 May 2018 17:22:25	U	SMPL
CCB25 - 1	CCB25	-61	-0.0453	-7471B	PPB	22 May 2018 17:24:41	U	SMPL
J3053-01 - 1	FK-GF-02-019-A	2460	0.4932	-7471B	PPB	22 May 2018 17:26:58	U	SMPL
J3053-01DUP - 1	FK-GF-02-019-ADUP	2663	0.5365	-7471B	PPB	22 May 2018 17:29:13	U	SMPL
J3053-01MS - 1	FK-GF-02-019-AMS	21694	4.6013	-7471B	PPB	22 May 2018 17:31:29	U	SMPL
J3053-01MSD - 1	FK-GF-02-019-AMSD	19359	4.1025	-7471B	PPB	22 May 2018 17:33:46	U	SMPL
J3053-03 - 1	FK-GF-02-019-B	1751	0.3418	-7471B	PPB	22 May 2018 17:36:02	U	SMPL
J3054-01 - 1	FK-GF-01-006-A	2096	0.4154	-7471B	PPB	22 May 2018 17:38:19	U	SMPL
J3054-03 - 1	FK-GF-01-006-B	1635	0.3170	-7471B	PPB	22 May 2018 17:40:36	U	SMPL
J3055-01 - 1	PHASE-1	1837	0.3601	-7471B	PPB	22 May 2018 17:42:54	U	SMPL
J3055-05 - 1	PHASE-2	2143	0.4255	-7471B	PPB	22 May 2018 17:45:12	U	SMPL
J3055-09 - 1	PHASE-3	2165	0.4302	-7471B	PPB	22 May 2018 17:47:28	U	SMPL
CCV26 - 1	CCV26	22786	4.8345	-7471B	PPB	22 May 2018 17:49:44	U	SMPL
CCB26 - 1	CCB26	-58	-0.0446	-7471B	PPB	22 May 2018 17:52:00	U	SMPL
J3056-01 - 1	TP-1	37522	7.9819	-7471B	PPB	22 May 2018 17:54:17	U	SMPL
J3056-05 - 1	TP-2	98973	21.1068	-7471B	PPB	22 May 2018 17:56:33	U	SMPL
J3058-01 - 1	VNJ-229	23431	4.9722	-7471B	PPB	22 May 2018 17:58:51	U	SMPL
J2675-11DLX10 - 1		41991	8.9364	-7471B	PPB	22 May 2018 18:01:10	U	SMPL
J3053-01LX5 - 1		209	0.0124	-7471B	PPB	22 May 2018 18:03:28	U	SMPL
J3053-01A - 1		21167	4.4887	-7471B	PPB	22 May 2018 18:05:47	U	SMPL
J3056-05DLX5 - 1		18591	3.9385	-7471B	PPB	22 May 2018 18:08:04	U	SMPL
CCV27 - 1	CCV27	22170	4.7029	-7471B	PPB	22 May 2018 18:10:22	U	SMPL
CCB27 - 1	CCB27	-51	-0.0431	-7471B	PPB	22 May 2018 18:12:40	U	SMPL